End-to-End 4D Heart Mesh Recovery Across Full-Stack and Sparse Cardiac MRI

Abstract

Reconstructing cardiac motion from CMR sequences is critical for diagnosis, prognosis, and intervention. Existing methods rely on complete CMR stacks to infer full heart motion, limiting their applicability during intervention when only sparse observations are available. We present *TetHeart*, the first end-to-end framework for unified 4D heart mesh recovery from both offline full-stack and intra-procedural sparse-slice observations. Our method leverages deformable tetrahedra to capture shape and motion in a coherent space shared across cardiac structures. Before a procedure, it initializes detailed, patient-specific heart meshes from high-quality full stacks, which can then be updated using whatever slices can be obtained in real-time, down to a single one during the procedure. *TetHeart* incorporates several key innovations: (i) an attentive slice-adaptive 2D–3D feature assembly mechanism that integrates information from arbitrary numbers of slices at any position; (ii) a distillation strategy to ensure accurate reconstruction under extreme sparsity; and (iii) a weakly supervised motion learning scheme requiring annotations only at keyframes, such as the end-diastolic and end-systolic phases. Trained and validated on three large public datasets and evaluated on additional private interventional and public datasets without retraining, *TetHeart* achieves state-of-the-art accuracy in both pre- and intra-procedural settings.

1 Introduction

Modeling cardiac shape and motion is a vital 4D reconstruction problem with direct implications for medical imaging, motion analysis, and image-guided intervention Sanz & Fayad (2008); Li et al. (2024); Qiao et al. (2025); Qian et al. (2025). Cardiac magnetic resonance (CMR) provides high-quality temporal anatomy and has been widely used for both shape reconstruction Bai et al. (2015); Ye et al. (2023); Dou et al. (2024); Yang et al. (2024); Xiao et al. (2024) and motion estimation Qin et al. (2018); Meng et al. (2022a;b; 2023); Yuan et al. (2023). However, as shown in Fig. 1 (a), existing methods rely on full volumetric CMR stacks and are thus restricted to offline analysis and pre-operative planning. In interventional settings, as depicted by Fig. 1 (b), only a handful of low-resolution slices can be acquired in real time Gao et al. (2012); Nayak et al. (2022); Rogers et al. (2023), a limitation we observed in practice at our collaborating university hospital even though state-of-the-art MRI equipment is available, as shown in Fig. 1(c).

This limitation makes current 4D modeling pipelines unusable *during interventions*. Currently, interventional MRI technology to guide electrophysiological ablations is gaining increasing attention among electrophysiologists and this technique is used in more than ten countries. The most frequently performed iCMR-guided ablation nowadays is right atrial flutter ablation Paetsch et al. (2019) with an estimated number of over 400 cases performed so far. In addition, left ventricular arrhythmia ablations have also been performed recently Götte et al. (2025). Thus, making 3D modeling fast enough for use in such an interventional MRI setting is what motivates us to go beyond the current state of the art and to recover full 3D shape and motion from sparse and partial observations. Our goal is to enable real-time tracking of patient-specific heart dynamics, with a view to delivering guidance to interventional teams.

To this end, we introduce *TetHeart*, a unified framework for 4D cardiac motion reconstruction that operates consistently across both full-stack and sparse-slice CMR inputs, the latter being what can be acquired in

Figure 1: *Cardiac motion reconstruction.* (a) Standard offline scenario. (b) Intervention room setting. (c) Interventional MRI setup at our collaborating hospital: A. Preparation room located next to the iCMR room. B. iCMR room equipped with screens to visualize the heart and catheters during the intervention. C. Preparation of the intubated and ventilated patient inside the iCMR. D. Control room with visualizations.

real-time during an intervention. As shown in Fig. 2, *TetHeart* first constructs a patient-specific mesh representation from pre-operative data and then dynamically updates it during procedures using only the slices that can be acquired in real-time, down to a single one. *TetHeart* builds on deep marching tetrahedra Shen et al. (2021), which combines the optimization efficiency of signed distance fields with the geometric flexibility of tetrahedral meshes, providing a coherent space for joint shape–motion learning. Given this, our main contributions are as follows:

- We propose a slice-adaptive 2D–3D feature fusion mechanism that allows us to dynamically aggregate features from arbitrarily located slices, enabling robust reconstruction across varying input sparsity. We further introduce a full-to-sparse distillation strategy that transfers knowledge from dense to sparse regimes to ensure accurate reconstruction under extreme input sparsity.

- We develop a weakly supervised motion learning scheme that requires annotations only at key phases (end-diastolic and end-systolic) and learns motion through temporal consistency. This is essential in practice because it makes the training feasible and affordable across diverse clinical environments.

- We demonstrate that a single model trained once perform well across offline and online scenarios, achieving state-of-the-art results on ACDC Bernard et al. (2018), M&Ms Campello et al. (2021), and M&Ms-2 Martín-Isla et al. (2023). It also shows strong performance by further validating on a privately collected interventional CMR dataset and the public 4DM Yuan et al. (2023) dataset without retraining, demonstrating its potential for deployment in real-world clinical environments.

In short, the combination of our feature fusion and training scheme yield a novel unified framework that enables consistent reconstruction and motion inference across a wide spectrum of observational sparsity. As a result, *TetHeart* can operate both offline and online, enabling comprehensive, patient-specific motion tracking across the entire clinical workflow. We achieve state-of-the-art accuracy for offline use with full-stack CMR, and perhaps more importantly, we unlock online use in intra-procedural scenarios where only sparse observations are available, which cannot be handled using current methods. Code will be public.

2 Related Work

We first review current methods to recovering static/dynamic heart, then the imaging modality we focus on.

2.1 Static Cardiac Shape Modeling

Reconstructing the heart from 3D medical images holds significant clinical value. However, CMR images often exhibit low through-plane resolution, making accurate reconstruction challenging. Current solutions Van Assen et al. (2006); Bai et al. (2015); Villard et al. (2018); Duan et al. (2019); Attar et al. (2019); Chen et al.

(2021); Xia et al. (2022); Beetz et al. (2022); Hu et al. (2023); Xiao et al. (2024) use either traditional optimization-based techniques or deep learning-based methods.

Optimization-Based Approaches usually rely on a two-stage pipeline, often starting from contours manually extracted from the CMR images. For example, Villard et al. (2018) starts from a predefined tubular mesh that is deformed to minimize a contour-matching loss and, in Hu et al. (2023), a 3D active shape model and an intensity model are used to align initial meshes with GT contours. Even though this works, relying on test-time optimization makes the reconstruction time-consuming, taking from tens of seconds to several minutes per frame. When applied to a full image sequence, this may extend to over an hour, making it impractical for real-time applications. Moreover, it has only been evaluated using GT contours, requiring labor-intensive manual annotation.

Deep Learning-Based Approaches leverage prior knowledge learned from large datasets to accelerate mesh reconstruction. For instance, in Beetz et al. (2022), contours are treated as sparse point cloud and decoded into 3D meshes using point- and graph-convolutions. Similarly, in Chen et al. (2021), contours are embedded into a 3D volume and a model made of 3D CNNs and GCNs is used to progressively deform a template mesh. These methods are much faster but also have drawbacks. Reconstructing meshes solely from extracted contours or points fails to exploit the rich appearance information contained in cine images Xiao et al. (2024). Furthermore, as contour extraction is not necessarily perfect, such two-step approaches are subject to error accumulation Chen et al. (2021).

2.2 Dynamic Cardiac Motion Modeling

Arguably, when reconstructing deforming hearts from sequences of CMR scans, one could simply run a static reconstruction algorithm on each individual frame. However, this would fail to exploit temporal consistency and make it difficult to establish inter-frame correspondences. Thus, several motion-based approaches that compute the deformation from frame to frame have been proposed.

When full CMR stacks are available, recent deep learning-based approaches Meng et al. (2022a;b; 2023); Yuan et al. (2023); Ye et al. (2023) can be used. MulViMotion Meng et al. (2022b) trains 3D CNNs to estimate left ventricular myocardial motion by predicting voxel-based deformation fields from CMR data, and then building 3D meshes by warping segmentation masks. DeepMesh Meng et al. (2023) decouples mesh reconstruction from motion estimation. It uses 3D CNNs and interpolation to predict per-vertex deformation. 4DMR Yuan et al. (2023) models the myocardium using implicit representation. Decoupled shape and motion latent codes are optimized at test time to predict a source shape and motion fields. We use MulViMotion, DeepMesh and 4DMR as baselines because they are among the best current methods.

Unfortunately, they also suffer from several limitations. First, some of them focus solely on myocardium, neglecting other structures, preventing comprehensive heart assessment. Second, optimization-based methods like 4DMR are time-consuming at inference time. Finally and most damagingly, they require full scans that cannot be acquired at sufficient frame rate, which precludes intra-procedural use as discussed below.

2.3 Real-Time MR Imaging

CMR imaging has a reputation for being a "slow" modality. Acquiring a full MRI stack involves sequential slice acquisition over a 4–6 heartbeat breath-hold, which can take several minutes Nayak et al. (2022); Rogers et al. (2023). This is feasible for pre-interventional planning but impractical for intra-procedural work.

Recently, real-time MRI (RT-MRI) has gained favor due to it can capture dynamic processes without gating, synchronization, or repetition Nayak et al. (2022). Its applications include complex electrophysiology procedures Paetsch et al. (2019), congenital heart disease Hauptmann et al. (2019), and cardiac interventions Gao et al. (2012); Xu et al. (2014). RT-MRI provides excellent soft-tissue contrast for assessing interventions like ablations Bhagirath et al. (2015); Rogers et al. (2023) and allows continuous visualization and tracking of devices, and it is free of ionizing radiation Paetsch et al. (2019); Campbell-Washburn et al. (2017).

Although RT-MRI enables rapid continuous image acquisition, it still involves a fundamental tradeoff between spatial resolution, temporal resolution, artifacts and reconstruction latency Nayak et al. (2022). Therefore,

Figure 2: *TetHeart* reconstructs cardiac motion from observed 2D slices. At $t = 0$, a 3D CMR stack $\mathbf{I}^0$ is processed by the AFA module with self-attention to extract a volumetric static feature V^0_{static}, which is used to generate the initial tetrahedra $\mathcal{G}^0$. For each subsequent time step t, the observed 2D slices $\mathbf{O}^t$ are encoded via the AFA module with cross-attention to obtain a volumetric motion feature V^t_{motion}, which is used to recover motion by predicting offsets to deform $\mathcal{G}^0$. Weights are shared between static and dynamic branch.

achieving a sufficient temporal resolution (e.g., 20–25 Hz) for use during invasive procedures, can only be done at the cost of acquiring a limited number of 2D slices at lower spatial resolution Gao et al. (2012); Xu et al. (2014); Campbell-Washburn et al. (2017); Nayak et al. (2022); Rogers et al. (2023). We are not aware of any current method that can reconstruct 3D cardiac motion using this kind of data. Yet, delivering a high-quality, dynamic 3D heart model to the intervention room based on real-time observations is crucial for guiding high-precision minimally invasive cardiac procedures Linte et al. (2009); Huang et al. (2009); Gao et al. (2012) and for real-time localization of devices within the heart Paetsch et al. (2019); Tampakis et al. (2023). This is the application that motivated us to begin this study.

3 Method

We now present *TetHeart*, a framework for 4D cardiac shape and motion reconstruction from either full CMR volumes or sparse 2D slice sets, as shown in Fig. 2. We first describe our heart representation based on deep marching tetrahedra Shen et al. (2021), an explicit–implicit hybrid model combining the optimization efficiency of signed distance fields with the geometric flexibility of tetrahedral meshes. Next, we introduce the Attentive 2D–3D Feature Assembler (AFA) we designed to adaptively aggregate features from arbitrary numbers of 2D slices at varying spatial locations. Finally, we detail our weakly supervised training scheme, enabling effective motion learning from limited annotations available only at key cardiac phases.

3.1 Formalization

To model deforming hearts, we use the deep marching tetrahedra formalism. We discretize the 3D space using a deformable tetrahedral grid, in which each vertex possesses an SDF, denoted as $\mathcal{G} = (\{\mathbf{v}_i\}, \{s_i\}, \mathcal{T})$, where $\mathbf{v}_i$ and s_i are the vertices and their corresponding SDFs in $\mathcal{T}$, the set of all tetrahedrons. The explicit surface can then be extracted using the differentiable marching tetrahedra algorithm Doi & Koide (1991).

Given a 3D heart model $\mathcal{G}^0$ reconstructed from the initial frame $\mathbf{I}^0 \in \mathbb{R}^{D \times H \times W}$, our goal is to estimate a deformed model $\mathcal{G}^t$ at each time instant t, based on subsequent observations $\mathbf{O}^t$ that reflect the ongoing cardiac motion. We do this by updating the vertices of $\mathcal{G}^0$:

$$\mathcal{G}^t = (\{\mathbf{v}^t_i\}, \{s_i\}, \mathcal{T}) , \ \mathbf{v}^t_i = \mathcal{D}(\mathbf{v}^0_i, \mathbf{O}^t) , \tag{1}$$

where $\mathbf{v}^0_i$ is a grid vertex from $\mathcal{G}^0$, $\mathbf{v}^t_i$ its counterpart after deformation. $\mathcal{D}$ is the deformation model which we instantiate in the following sections and makes it possible to recover deformations across the whole sequence.

The observations $\mathbf{O}^t$ can be obtained in two manners.

1. **Online.** For intra-procedural use, $\mathbf{I}^0 \in \mathbb{R}^{D \times H \times W}$ is a full-slice image acquired before the intervention, taking as much time as needed. However, it has to be possible to acquire $\mathbf{O}^t$ at sufficient frame rate and, consequently, it can only comprise a small set—from 1 to 3—of 2D slices at lower spatial resolution and at arbitrary locations Gao et al. (2012).

2. **Offline.** Full-slice CMR images are available both initially and at time t, which is acquired from a slow slice-by-slice scanning process. In this case, $\mathbf{O}^t$ and $\mathbf{I}^0$ are of the same dimension $\mathbb{R}^{D \times H \times W}$.

In the following sections, we first introduce the AFA module and then describe how it is incorporated in our approach to generate the initial shape model from $\mathbf{I}^0$ and then deforming it given $\mathbf{O}^t$. Whether it is a full scan or a sparse set of slices, we use the same approach in both cases.

3.2   Attentive 2D-3D Feature Assembler

Because the observations can be either a full stack, a sparse set of slices, or anything in between, we need a mechanism to extract useful features in all cases. To this end, we developed the Attentive 2D-3D Feature Assembler (AFA) module that takes $\mathbf{I}^0$ and $\mathbf{O}^t$ as input and produces features.

Let us write $\mathbf{O}^t$ as $\{\mathbf{I}_s^t \in \mathbb{R}^{H \times W} | s = 1, \ldots, S\}$, a collection of 2D slices where S is the number of slices. With this, $\mathbf{I}^0$ can be written as $\{\mathbf{I}_d^0 \in \mathbb{R}^{H \times W} | d = 1, \ldots, D\}$. A full-slice observation is simply one where $S = D$.

In practice, we first perform 2D convolutions on each individual slice to create a collection of 2D feature maps F_{2d}^0 and F_{2d}^t from $\mathbf{I}^0$ and $\mathbf{O}^t$. Given that $\mathbf{O}^t$ is not necessarily in the same coordinate system as $\mathbf{I}^0$, we use the metadata associated to these slices in the DICOM header to position them spatially in the $\mathbf{I}^0$ volume. This links each 2D location in the feature maps to a precise 3D location. In the remainder of this section, we first discuss how we use these 2D feature maps to compute volumetric *motion features* from F_{2d}^0 and F_{2d}^t, which are obtained from slices acquired at different times and can be used to estimate deformations. We then describe how we derive *static features* from a dense set of slices all acquired at the same time, which can be used to reconstruct the initial static model.

Motion Features. A naive way to use attention to construct motion features would be to treat the F_{2d}^0 as the query and the F_{2d}^t as both keys and values. However, this would be inefficient because global attention aims at capturing long-range dependencies, which is not needed to model the inherently local nature of anatomical motion. Also, this requires computing full pairwise relationships between all query and key positions, which is computationally expensive when using many slices and high-resolution feature maps.

To avoid this, we take our inspiration from the Swin Transformer Liu et al. (2021) and use instead a variant of attention that preserves locality and significantly reduces computational complexity. For each voxel in F_{2d}^0, we first identify k_1 spatially nearest 2D slices, or all slices if there are only k_1 or less. Within each selected slice, we then retrieve the features at the k_2 spatially closest positions to serve as keys and values. Attention is subsequently computed only between the query and its corresponding set of localized keys and values. This can be written as

$$V_{\text{motion}}^t = AFA(F_{2d}^0, F_{2d}^t) = \text{MHAttention}(Q, K, V), \text{ where } Q = F_{2d}^0, \ K = V = \text{NN-Select}_{F_{2d}^0, k_1, k_2}(F_{2d}^t),$$

where *NN-Select* denotes the nearest slice and position selecting operation. *MHAttention* denotes a Multi-Head Attention Vaswani (2017) operation. A learnable position embedding is used to encode the spatial information. In practice, we set $k_1 = 3$ and $k_2 = 9$. This ensures that the computational complexity of our attention mechanism is approximately equivalent to that of a 3D convolution with a kernel size of 3. As a result, the AFA module achieves high computational efficiency while retaining the flexibility and adaptability of attention-based feature aggregation. The complete algorithm pipeline is described in Sec. B.

While our AFA module effectively encodes features from varying 2D slice configurations, a persistent challenge remains in the online scenario, where the limited number of slices can hinder accurate 3D motion reconstruction. To alleviate this issue, we introduce a distillation loss that encourages the model to transfer knowledge from full-slice inputs to partial-slice inputs during training. In short, we predict two versions of motion features from full-slice observation and its randomly sampled slice subset. The distillation loss encourages the encoded feature from partial slices to be close to the feature from full-slice input. We formulate the loss we minimize to achieve this in Section 3.5.

Static Features. The motion features described above relate $\mathbf{I}^0$ to $\mathbf{O}^t$. The same mechanism can also be used to relate $\mathbf{I}^0$ to itself and produce features that can be used for static reconstruction. This can be

similarly written as

$$V^0_{\text{static}} = AFA(F^0_{2d}, F^0_{2d}) = \text{MHAttention}(Q, K, V) \text{, where } Q = F^0_{2d}, \ K = V = \text{NN-Select}_{F^0_{2d}, k_1, k_2}(F^0_{2d}) \ .$$

Generating Final Features. After obtaining the volumetric static and motion features V^0_{static} and V^t_{motion}, we use a U-Net-like Ronneberger et al. (2015); Isensee et al. (2021) architecture for further encoding. Just as standard U-Net architecture, it also consists of a downsampling and an upsampling stream. The last level features of the upsample stream are taken as the final features, which we denoted as F^0_{static} and F^t_{motion}.

Note that the initial 2D encoding block and subsequent 3D convolutions are shared for static and motion feature extraction. This unified design allows motion prediction to benefit from the rich 3D spatial information learned in the static reconstruction task, facilitating more accurate motion estimation—an advantage overlooked by some previous works Meng et al. (2023); Yuan et al. (2023). As shown in the experiments, this design leads to superior motion reconstruction performance.

3.3 Initial Heart Model Reconstruction

Given F^0_{static} computed as discussed above, our first task is to reconstruct the initial tetrahedral mesh $\mathcal{G}^0$. To this end, we start from tetrahedral grid $\mathcal{G}^{init}$ obtained by uniformly sampling a unit cube. We then trilinearly interpolate F^0_{static} at the vertices of $\mathcal{G}^{init}$. An SDF value and an offset for each vertex is predicted using a GCN to create $\mathcal{G}^0$. We write

$$\begin{aligned}
\mathcal{G}^0 &= (\{\mathbf{v}'_i\}, \{s_i\}, T), \\
\mathbf{v}'_i &= \mathbf{v}_i + \Delta\mathbf{v}_i, \\
(\Delta\mathbf{v}_i, s_i) &= \text{GCN}([F^0_{\text{static}}(\mathbf{v}_i), \mathbf{v}_i]) \ ,
\end{aligned} \tag{2}$$

where $\mathbf{v}_i$ is a grid vertex of $\mathcal{G}^{init}$ and $\mathbf{v}'_i$ the same grid vertex after deformation. $F^0_{\text{static}}(\mathbf{v})$ denotes feature extraction at $\mathbf{v}$ from F^0_{static} with trilinear interpolation, and $[\cdot, \cdot]$ denotes concatenation. To enhance spatial sensitivity, the vertex location is appended before it is passed to the GCN. For datasets with C classes, we set the output channel to $C \times 4$, 3 for deformation and 1 for SDF. Following prior works Wickramasinghe et al. (2020); Kong et al. (2021); Bongratz et al. (2022); Cruz et al. (2021), our network also has an auxiliary segmentation prediction branch during training time.

3.4 Motion Recovery

The static reconstruction scheme described above delivers comparable or even better reconstruction performance to that of approaches based on pure 3D U-Net architectures Wickramasinghe et al. (2020); Kong et al. (2021); Bongratz et al. (2022); Cruz et al. (2021), with the added benefit that it also allows the independent encoding of individual 2D slices. We now exploit this mechanism to compute deformations for the initial static shape from sets of slices acquired later. Given $\mathcal{G}^0$, F^0_{static}, and F^t_{motion} computed as described above, we use a combination of GCN and GRU Cho et al. (2014) layers to effectively aggregate spatial information and instantiate the deformation model $\mathcal{D}$ of Eq. 1. We write

$$F^{cat}(\mathbf{v}^{(s)}_i) = [F^0_{\text{static}}(\mathbf{v}^{(s)}_i), F^t_{\text{motion}}(\mathbf{v}^{(s)}_i)] \ , \tag{3}$$

$$F^{gcn}(\mathbf{v}^{(s)}_i) = \text{GCN}(F^{cat}(\mathbf{v}^{(s)}_i)) \ , \tag{4}$$

$$\mathbf{h}^{(s+1)}_i = \text{GRU}([F^{gcn}(\mathbf{v}^{(s)}_i), \mathbf{v}^{(s)}_i], \mathbf{h}^{(s)}_i) \ , \tag{5}$$

$$\mathbf{v}^{(s+1)}_i = \mathbf{v}^{(s)}_i + \text{MLP}(\mathbf{v}^{(s)}_i, \mathbf{h}^{(s+1)}_i) \ . \tag{6}$$

In Eq. 3, the vertex features are extracted by trilinear interpolating F^0_{static} and F^t_{motion} at their locations. They are then concatenated and passed through Eq. 4-6 to gradually update the vertex position. Eqs. 3-6 are repeated twice.

3.5 Two-Stage Weakly Supervised Training Using Only Keyframe Annotations

Creating high-quality annotations for an entire sequence is costly. Therefore, many datasets provide annotations only for keyframes, such as end-diastolic (ED) and end-systolic (ES) phases. This rules out full supervision. To address this, we propose a two-stage weakly-supervised pipeline, which makes training the reconstruction and motion branches feasible using only the available keyframe annotations. The core idea is to first train a static reconstruction branch on labeled frames, and then repurpose it to initialize shapes for unlabeled frames, from which motion dynamics can be recovered.

Training the Reconstruction Branch. We start by training the reconstruction branch of Section 3.3. At each training iteration, we randomly select a labeled image $\mathbf{I}$, that is, ED or ES frame, with ground truth segmentation and mesh annotation $L^{gt}, \mathcal{M}^{gt}$. Let $L^p, \mathcal{G}^p$ be our model's segmentation and tetrahedra prediction. We minimize the loss

$$\mathcal{L}_{shape}(\mathbf{I}) = \lambda_{cd}\mathcal{L}_{cd}(\mathbf{MT}(\mathcal{G}^p), \mathcal{M}^{gt}) + \lambda_{sdf}\mathcal{L}_{sdf}(\mathcal{G}^p, \mathcal{M}^{gt}) + \lambda_{ce}\mathcal{L}_{ce}(L^p, L^{gt}) ,$$

where $\mathcal{L}_{cd}$ is chamfer distance, $\mathcal{L}_{sdf}$ is an L1-loss used to supervise the predicted SDF values at tetrahedral grid vertices with the SDF value queried from the ground truth mesh. $\mathcal{L}_{ce}$ is the cross-entropy loss defined on the segmentation map. $\mathbf{MT}$ denotes the marching tetrahedra algorithm Doi & Koide (1991) to convert $\mathcal{G}^p$ into surface mesh. $\lambda_{cd}, \lambda_{sdf}, \lambda_{ce}$ are set to $1.0, 0.1, 0.1$ respectively. The data is diverse enough for our model to generalize to arbitrary timesteps even though annotations are only available for ED/ES frames.

Training the Motion Branch. We then train the motion branch. At each iteration, we randomly select a sequence $\{\mathbf{I}^t\}_{t=0}^T$, and choose an unlabeled frame $\mathbf{I}^u$ and a labeled frame $\mathbf{I}^l$ from it. We first predict $\mathcal{G}^u$ from $\mathbf{I}^u$ using the trained reconstruction branch. Next, assuming that $\mathbf{I}^l$ comprises a total of D slices, we randomly pick 1 to D 2D slices to form $\hat{\mathbf{I}}^l$ to simulate the online few-slice case. Crucially, $\mathbf{I}^l$ and $\hat{\mathbf{I}}^l$ undergo random downsampling/rotation/noise jittering to simulate image quality degradation encountered in the interventional setting. The corresponding ground truth mesh $\mathcal{M}^l$ undergoes the same transformation. V_{motion}^l and $\hat{V}_{\text{motion}}^l$ are generated from $\mathbf{I}^l$ and $\hat{\mathbf{I}}^l$ using the AFA module. We write

$$V_{\text{motion}}^l = AFA(F_{2d}^u, F_{2d}^l), \quad \hat{V}_{\text{motion}}^l = AFA(F_{2d}^u, \hat{F}_{2d}^l) .$$

We use these 3D features to generate two versions of deformed tetrahedra using Eq. 1: $\mathcal{G}^{u \to l}$ from full-slice encoded feature V_{motion}^l, and $\hat{\mathcal{G}}^{u \to l}$ from few-slice encoded feature $\hat{V}_{\text{motion}}^l$. The network is trained with

$$\mathcal{L}_{motion}(\mathbf{I}^l, \mathbf{I}^u) = \mathcal{L}_{distill}(V_{\text{motion}}^l, \hat{V}_{\text{motion}}^l) + \mathcal{L}_{cd}(\mathrm{MT}(\mathcal{G}^{u \to l}), \mathcal{M}^l) + \mathcal{L}_{cd}(\mathrm{MT}(\hat{\mathcal{G}}^{u \to l}), \mathcal{M}^l),$$

$$\mathcal{L}_{distill}(V_{\text{motion}}^l, \hat{V}_{\text{motion}}^l) = ||\hat{V}_{\text{motion}}^l - \mathtt{sg}(V_{\text{motion}}^l)||_2 ,$$

where $\mathtt{sg}$ denotes the stop-gradient operation. Minimizing the *distillation loss* $\mathcal{L}_{distill}$ encourages features from few-slice input $\hat{\mathbf{I}}^l$ to align with the one from full-slice input $\mathbf{I}^l$, while minimizing the chamfer losses enables the deformation model to infer accurate motion from the feature map encoded from any number of 2D slices, while retaining the ability to infer motion from complete 3D full-slice feature map. As a result, minimizing $\mathcal{L}_{motion}$ guarantees good reconstruction results for both full and few-slice inputs. This training strategy allows training a 4D model without dense temporal annotations.

Once trained, our model can be used for new patients without retraining. Note that our training strategy actually enables the model to predict cardiac deformation starting from any phase, not just the ED phase. Additionally, by employing an aggressive sampling and augmentation scheme, the model learns to reconstruct motion from a highly diverse set of input configurations. As a result, even though the online scenario is simulated during training, our model shows strong extendibility. It can robustly handle cases where $\mathbf{O}^t$ originates from a different sequence or where there is a scanning angle discrepancy relative to $\mathbf{I}^0$, consistently producing reliable motion predictions despite these challenges as will be shown in the experiment section.

4 Experiments

4.1 Experimental Settings

Datasets and Metrics. We train and validate our method on a unified dataset constructed from three large and diverse publicly available datasets: *ACDC, M&Ms*, and *M&Ms-2* Bernard et al. (2018); Campello

				ACDC						M&Ms						M&Ms-2					
				CD (mm²) ↓			Dice (%) ↑			CD (mm²) ↓			Dice (%) ↑			CD (mm²) ↓			Dice (%) ↑		
Method	GT-Free	Optim-Free	Repre.	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV
										1-Slice											
4DMR	✗	✗	SDF	29.53	-	-	75.63	-	-	19.87	-	-	76.12	-	-	19.37	-	-	74.11	-	-
MR-Net	✗	✓	Mesh	28.61	40.70	46.15	77.43	82.12	73.33	14.59	16.46	25.11	77.59	80.35	73.94	15.61	16.68	25.23	76.29	82.66	76.23
MulViMotion	✓	✓	Mesh	29.13	-	-	75.02	-	-	18.43	-	-	72.12	-	-	19.51	-	-	70.21	-	-
DeepMesh	✓	✓	Mesh	27.47	-	-	76.89	-	-	14.29	-	-	76.15	-	-	15.35	-	-	73.75	-	-
Ours-Mesh	✓	✓	Mesh	17.42	25.11	42.33	80.14	84.90	74.32	11.26	14.17	23.19	79.88	82.11	75.96	11.98	15.11	22.63	77.66	84.97	77.99
Ours-SDF	✓	✓	SDF	18.65	26.45	43.78	79.55	83.43	74.00	12.16	14.88	23.91	79.09	81.62	74.23	12.15	15.72	23.34	77.01	84.37	77.15
Ours	✓	✓	Tet	**15.24**	**23.65**	**40.64**	**82.25**	**86.20**	**75.42**	**9.76**	**12.37**	**21.42**	**81.00**	**84.45**	**77.32**	**10.12**	**13.89**	**20.27**	**79.33**	**86.06**	**79.65**
										5-Slice											
4DMR	✗	✗	SDF	24.56	-	-	77.03	-	-	16.16	-	-	77.33	-	-	15.52	-	-	76.67	-	-
MR-Net	✗	✓	Mesh	18.82	24.22	36.54	79.54	84.25	74.67	10.60	12.16	20.28	78.70	83.31	76.91	10.76	12.05	19.25	77.84	86.03	78.79
MulViMotion	✓	✓	Mesh	23.05	-	-	77.03	-	-	15.55	-	-	74.03	-	-	15.03	-	-	73.42	-	-
DeepMesh	✓	✓	Mesh	18.34	-	-	78.05	-	-	11.19	-	-	77.59	-	-	10.98	-	-	75.33	-	-
Ours-Mesh	✓	✓	Mesh	11.66	19.39	33.91	82.92	85.88	75.73	9.14	11.07	18.62	80.99	85.11	78.67	8.68	10.96	16.82	79.82	87.99	80.13
Ours-SDF	✓	✓	SDF	12.21	20.03	34.52	82.35	85.32	75.11	9.63	11.69	19.26	80.45	84.56	78.05	9.22	11.55	17.42	79.21	87.41	79.53
Ours	✓	✓	Tet	**10.22**	**18.31**	**32.61**	**84.38**	**87.32**	**77.23**	**7.93**	**9.72**	**17.39**	**82.83**	**86.76**	**79.27**	**7.30**	**9.96**	**15.58**	**81.24**	**89.37**	**81.48**

Table 1: *Few-slice quantitative evaluation on unified dataset.* We compared the predicted 3D meshes by deforming from ED frame to ES frame to the ground-truth 3D meshes. **1- / 5-Slice**: Number of slices used to recover the motion. GT-Free: no need of ED GT mesh at inference time to predict motion. Optim-Free: no need of test-time optimization.

Method	LVESV MAE (ml)↓	LVEF MAE (%) ↓	RVESV MAE (ml)↓	RVEF MAE (%) ↓
		1-Slice		
MR-Net Chen et al. (2021)	9.76	7.87	16.40	11.75
Ours-Mesh Kong et al. (2021)	8.92	6.37	16.41	9.83
Ours-SDF Cruz et al. (2021)	9.35	6.82	17.04	10.21
Ours	**7.51**	**4.83**	**15.03**	**8.51**
		5-Slice		
MR-Net Chen et al. (2021)	7.34	6.29	13.45	8.87
Ours-Mesh Kong et al. (2021)	7.64	5.11	12.81	7.33
Ours-SDF Cruz et al. (2021)	7.93	5.32	13.12	7.59
Ours	**6.86**	**4.30**	**11.88**	**6.55**
		Full-Slice		
FFD Rueckert et al. (2002)	23.53	14.95	24.06	15.87
dDemons Vercauteren et al. (2007)	33.97	22.11	63.03	38.52
MR-Net Chen et al. (2021)	6.75	4.16	12.05	7.01
Ours-Mesh Kong et al. (2021)	5.07	2.23	9.85	4.96
Ours-SDF Cruz et al. (2021)	5.22	2.34	9.98	5.01
Ours	**4.84**	**2.04**	**9.65**	**4.77**

Table 2: *Clinical indexes evaluation on M&Ms dataset.* 1- / 5- / Full-Slice denotes the number of slices used to recover motion.

et al. (2021); Martín-Isla et al. (2023). Segmentation for the left ventricle (LV), right ventricle (RV), and myocardium (MYO) are provided, but only for the ED and ES frames. The unified dataset comprises 835 CMR sequences. To prevent data leakage, we use a subject-level split to randomly split it into training, validation, and testing sets in a ratio of 60%, 20%, and 20%. Dataset details are provided in Appendix A.1.

We use two additional datasets for external evaluation in a clinical context where model retraining is not feasible. The first dataset, referred to as *iCMR*, was collected on the interventional MRI system depicted by Fig. 1(c) under both rest and elevated-heart-rate conditions. It is designed with the target online scenario in mind. For each subject, full short-axis cine stacks were acquired at rest, while additional single mid-ventricular slices were collected at rest and during mild in-scanner exercise to simulate heart-rate variations. The dataset thus reflects realistic cardiac motion dynamics encountered in online interventional MRI. Details on the scanning protocol, equipment, and exercise procedure are provided in Appendix A.2. To evaluate full-sequence performance across the entire cardiac cycle beyond the keyframes, we use the second external dataset, *4DM* Yuan et al. (2023), which focuses on left ventricular myocardium reconstruction and provides mesh annotations for every frame. This dataset allows rigorous evaluation of the proposed weakly supervised motion learning scheme.

For evaluation, we use the L2-norm Chamfer Distance (CD) to measure 3D reconstruction quality and the Dice score to quantify intra-slice segmentation results. We also report the Mean-Absolute Error of LV/RV ES Volume (LVESV/RVESV), LV/RV ejection fraction (LVEF/RVEF) derived from our predicted meshes. LV/RVEF are representative cardiac function indexes used for diagnosis. See Appendix A.3 for details.

Implementation Details. In all experiments, the slices are resampled by linear interpolation to a spacing of 1.25×1.25 mm. From this, the AFA module constructs 3D volume with an effective spacing of $1.25 \times 1.25 \times 2$ mm. See Appendix B for model and optimization details.

Figure 3: *1-Slice Comparison.* (Left) We predict meshes by deforming from ED to ES frame. Color indicates the magnitude of point-to-surface error. (Right) Segmentation results on slices at the apex and base location. Note 4DMR can only output motion for myocardium.

Baselines. We compare our method against several existing approaches for 4D shape reconstruction. For the online interventional setting, as noted earlier, most methods assume access to a complete CMR stack, making them unsuitable for few-slice inputs. Therefore, we adapted several baselines to operate under sparse inputs, including modified versions of **4DMR** Yuan et al. (2023),**MR-Net** Chen et al. (2021), **MulViMotion** Meng et al. (2022b) and **DeepMesh** Meng et al. (2023). We also include two variants derived from shape reconstruction methods: MeshDeformNet Kong et al. (2021) and DeepCSR Cruz et al. (2021). We denote them as **Ours-Mesh** and **Ours-SDF** since their main differences with ours lie in the shape representation. For the traditional offline setting, in addition to the above methods, we further compare against Free Form Deformation (**FFD**) Rueckert et al. (2002), diffeomorphic Demons (**dDemons**) Vercauteren et al. (2007), **MR-Net w. nnU-Net**. Details about the modified baselines are provided in Appendix C.

4.2   Motion Estimation from a Few Slices

We present results under the online few-slice scenario that represents the primary motivation of this work. Fig. 3 shows a representative motion reconstruction using only the central slice as input: one from a normal (NOR) subject and another from a case of dilated cardiomyopathy (DCM). Due to stronger contractile function, the normal heart exhibits larger deformations, making motion prediction inherently more challenging, as reflected by higher reconstruction errors. Nevertheless, our method delivers accurate predictions in both cases. Moreover, segmentation results on slices near the apex and base demonstrate that our model produces reliable deformations even in regions distant from the input slice. Complete reconstructed sequences and their corresponding volume–time curves are shown in Fig. 4.

Quantitative results on the full test set are reported in Tab. 1 and the top section of Tab. 2, including Chamfer Distance, Dice scores, and cardiac function indices. For consistency, all experiments were performed using the central 1 or 5 slices, though slices at arbitrary positions could have been used, as discussed in Section 4.4. *TetHeart* consistently outperforms all baselines without requiring ground-truth contours or segmentations as input at inference time. It infers cardiac motion directly from raw CMR images, eliminating the need for time-consuming test-time optimization, and achieves an inference speed of 12 FPS on an NVIDIA V100 GPU. Further acceleration could be achieved through model quantization or lightweight encoders to meet strict real-time constraints. For model size, ours has 23M parameters and the best competitor MR-Net has 22M, runs at similar speed and performs worse. Notably, even when provided with a single slice, *TetHeart* produces meaningful physiological indices, underscoring its suitability for clinical deployment.

Although Ours-Mesh and Ours-SDF incorporate the same sparse feature handling strategy as our full model, *TetHeart* still achieves superior performance. This confirms that the tetrahedral representation effectively preserves spatial coherence and serves as a more expressive basis for 3D motion inference under sparse

Method	GT-Free	Optim-Free	Repre.	ACDC						M&Ms						M&Ms-2					
				CD (mm²) ↓			Dice (%) ↑			CD (mm²) ↓			Dice (%) ↑			CD (mm²) ↓			Dice (%) ↑		
				Myo	LV	RV	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV	Myo	LV	RV
FFD	✗	✗	Mesh	21.84	38.85	51.04	69.46	79.56	72.04	42.01	63.23	88.77	62.84	76.65	66.23	40.79	66.42	95.66	63.13	79.80	76.02
dDemons	✗	✗	Mesh	15.72	22.97	35.85	77.06	86.87	73.60	32.16	44.39	61.15	65.83	79.87	65.09	34.08	48.77	66.22	67.25	83.44	72.39
4DMR	✗	✗	SDF	10.89	-	-	80.03	-	-	8.42	-	-	82.33	-	-	8.25	-	-	82.67	-	-
MR-Net	✗	✓	Mesh	14.02	19.21	24.07	80.02	86.35	78.93	8.86	10.03	16.86	79.40	86.90	79.73	8.39	7.24	12.99	80.72	87.86	80.60
MR-Net w. nnU-Net	✓	✓	Mesh	15.11	20.45	26.13	79.66	85.38	76.45	9.88	10.25	18.60	78.56	85.00	77.21	9.83	9.24	13.35	79.56	85.89	78.31
MulViMotion	✓	✓	Mesh	17.21	-	-	78.10	-	-	13.75	-	-	75.31	-	-	12.33	-	-	75.23	-	-
DeepMesh	✓	✓	Mesh	16.42	-	-	79.89	-	-	9.24	-	-	78.51	-	-	9.20	-	-	77.75	-	-
Ours-Mesh	✓	✓	Mesh	6.80	11.84	23.41	86.62	88.11	84.74	4.26	5.08	13.92	84.09	87.12	83.73	4.15	5.12	11.04	85.01	88.67	86.65
Ours-SDF	✓	✓	SDF	7.21	12.65	22.78	86.95	88.43	84.00	4.44	5.26	14.21	83.87	86.91	83.77	4.30	5.44	11.32	84.55	88.37	86.10
Ours	✓	✓	Tet	**6.63**	**11.51**	**22.15**	**87.12**	**88.96**	**84.21**	**4.11**	**4.85**	**13.68**	**84.31**	**87.43**	**83.99**	**4.02**	**5.04**	**10.82**	**85.33**	**89.13**	**86.94**

Table 3: *Full-slice quantitative evaluation on unified dataset.* We compared the predicted 3D meshes by deforming from ED frame to ES frame to the ground-truth 3D meshes. GT-Free: no need of ED GT mesh at inference time to predict motion. Optim-Free: no need of test-time optimization.

Figure 4: *Motion sequence prediction results using 1-/Full-slices by deforming from ED frame to the target frame.* Corresponding Volume-frame curve is given on the right. #1: NOR, #2: DCM.

observations. Fig. 5 illustrates performance as a function of slice count: the fewer slices are used, the larger the gap between *TetHeart* and the ablated variants. Nevertheless, both variants still outperform all prior methods by a clear margin, demonstrating that the AFA module and the training strategy form a robust and generalizable solution for incomplete observations.

The relatively poor performance of 4DMR can be attributed to its simple MLP architecture, which may lack the representational power to generalize well to the large and heterogeneous datasets considered here. Moreover, its test-time optimization strategy does not generalize well to severely under-sampled inputs. The weaker performance of MR-Net likely arises from the lack of reconstruction features and its sole reliance on contour points for deformation prediction, which overlooks rich cine image information.

4.3 Motion Estimation from Full Stacks

Since we had to modify existing techniques to compare them to ours in the few-slice scenario, we now compare our method to the original baselines on the full set of slices they were designed to operate on and report the results in Tab. 3 and the bottom of Tab. 2. In Fig. 4 and 7, we show qualitative results on the same hearts as in Fig. 3, but now using all slices as input instead of only one.

Remarkably, although the main driver of this work was operating on as few slices as possible, we also consistently outperform the baselines in the full-slice setting. We attribute this to several factors. Using predicted

Figure 5: Chamfer distance as a function of the number of slices used to capture the motion.

contours in MR-Net leads to slightly worse performance, indicating such a two-stage automated pipeline is prone to error accumulation, consistent with the conclusions in the original paper Chen et al. (2021). MulVi-Motion estimates voxel-wise motion fields and reconstructs meshes from warped segmentation maps, which is inherently less accurate than directly modeling meshes. Compared to DeepMesh, our model benefits from a shared encoding network, allowing the motion branch to leverage features from the reconstruction branch. Moreover, the AFA module dynamically aggregates spatially adjacent features, improving prediction quality.

4.4 Evaluation on External Datasets without Retraining

Method	Training Dataset			Dice (%) ↑		
	ACDC	M&Ms	M&Ms-2	Myo	LV	RV
4DMR	✓	✓	✓	71.33	-	-
MR-Net	✓	✓	✓	76.11	78.65	72.11
Ours-Mesh	✓	✓	✓	79.17	82.02	78.97
Ours-SDF	✓	✓	✓	78.42	81.11	78.21
Ours	✓	-	-	69.12	71.33	69.81
	-	✓	-	76.99	78.17	77.68
	-	-	✓	76.13	79.67	78.21
	✓	✓	✓	**80.15**	**83.22**	**79.76**

Table 4: *Evaluation results on the iCMR dataset.* We report the Dice Score as the evaluation metric. Check marks denote the dataset used for training.

We now perform external evaluation on the iCMR and 4DM datasets *without* retraining. In Fig. 6, we show sequence predictions on iCMR dataset with the spatial position of the real-time captured 2D slice visualized. As can be seen, *TetHeart* reconstructs plausible heart motion across the cardiac cycle under both rest and exercise situation. Moreover, the predicted volume–time curves confirm the physiological plausibility of our motion reconstruction. In the exercise condition, the heart exhibits a shorter ES–ED recovery period and a reduced duration of each cardiac cycle, aligns well with expectation. The quantitative results of Tab. 4 also proves our model performs best. 4DMR exhibits poor generalization, likely due to significant distribution shifts across datasets. While MR-Net show some generalization ability, it still performs worser than us.

Moreover, training with a larger and more diverse dataset improves generalization. And consistent with the findings from previous evaluations, representing the heart as tetrahedra enables more effective spatial information retention and exchange, particularly in scenarios with limited slice input.

Please refer to Appendix E for more details and the external evaluation results on the 4DM dataset.

4.5 Ablation Studies

We conduct ablation studies to evaluate our design choices in Tab. 5. We first verify the effectiveness of the training strategy proposed in Sec. 3.5, especially the influence of using slice subsampling and random downsampling/rotation/noise jittering when training the motion branch. To quantitatively evaluate this, we simulate interventional artifacts on the Unified dataset by applying down-sampling, noise jitter, and 10° rotation to ES frame and its GT meshes during testing. Tab. 5(left) confirms our training strategy overcomes such challenges that are representative of what could occur in interventional scenarios. The clean ES results from Tab. 1 are given as reference.

Next we ablate the model designs in Tab. 5(right). First, we examine the AFA module. Simply stacking 2D features yields similar performance to AFA in the full-slice setting but fails to generalize when the number or positions of slices vary, highlighting the need for a flexible aggregation mechanism. Removing position encodings moderately degrades performance, confirming the importance of spatial information,

Figure 6: *Motion sequence prediction on the iCMR dataset.* (Left) Spatial configuration of the initial SAX stack and the online observations. (Middle) Cardiac motion sequences for each subject under rest and exercise conditions. (Right) Volume–time curves.

Method (1-Slice)	M&Ms CD (mm^2) ↓		
	Myo	LV	RV
Ours w.o. Sec. 3.5 aug	14.45	18.72	30.67
Ours	10.98	13.76	23.66
Ours (clean input)	9.76	12.37	21.42

Method	M&Ms Myo CD (mm^2) ↓		
	1-Slice	5-Slice	Full-Slice
Ours	9.76	7.93	4.11
- AFA module	-	-	4.22
- position embedding	11.23	9.04	4.67
- image feature in query Q	16.67	12.37	6.44
- shared encoding network	20.66	11.63	6.43
- distillation loss	10.34	8.12	4.03

Table 5: *Ablation study.* (left) Evaluation results with simulated interventional artifacts. We compared the predicted meshes by deforming from ED to interventional-like ES. (right) Ablation results on M&Ms dataset.

while using only position encodings leads to a substantial drop—showing that both image features and spatial coordinates are essential for effective motion prediction. When the motion branch is trained without parameter sharing and randomly initialized, performance decreases significantly, demonstrating that our shared-parameter design effectively transfers 3D knowledge from reconstruction to motion modeling.

Lastly, we ablate the impact of the distillation loss. Without it, the model performs slightly better in the full-slice setting but encodes less discriminative features from sparse 2D inputs, resulting in poorer few-slice performance. Combining the distillation loss with our aggressive sampling strategy consistently improves results across all scenarios, underscoring the importance of this training scheme for improving robustness and generalization. Additional ablation results are provided in Appendix F.

5   Conclusion

We have presented *TetHeart*, a unified framework for 4D cardiac shape and motion reconstruction from 2D CMR slices, operating robustly from full stacks down to a single real-time slice. For the first time, we show that accurate 3D cardiac motion can be inferred from sparse intra-operative CMR data, enabling real-time tracking and guidance during interventions. This has the potential to shift cardiac reconstruction from retrospective analysis to live clinical application, laying the foundation for patient-specific digital twins of the beating heart and paving the way for adaptive, predictive, and personalized image-guided procedures.

As an immediate next step, we plan to deploy *TetHeart* in an interventional MRI suite at our collaborating hospital. Looking ahead, we aim to extend this framework to other dynamic organs, integrate multimodal imaging, and develop predictive digital twins that anticipate motion under both physiological and interventional conditions—advancing toward a new generation of real-time, personalized, image-guided medicine.

References

Rahman Attar, Marco Pereañez, Ali Gooya, Xènia Albà, Le Zhang, Milton Hoz de Vila, Aaron M Lee, Nay Aung, Elena Lukaschuk, Mihir M Sanghvi, et al. Quantitative cmr population imaging on 20,000 subjects of the uk biobank imaging study: Lv/rv quantification pipeline and its evaluation. *Medical Image Analysis*, 2019.

Wenjia Bai, Wenzhe Shi, Antonio de Marvao, Timothy JW Dawes, Declan P O'Regan, Stuart A Cook, and Daniel Rueckert. A bi-ventricular cardiac atlas built from 1000+ high resolution mr images of healthy subjects and an analysis of shape and motion. *Medical Image Analysis*, 2015.

Wenjia Bai, Hideaki Suzuki, Jian Huang, Catherine Francis, Shuo Wang, Giacomo Tarroni, Florian Guitton, Nay Aung, Kenneth Fung, Steffen E Petersen, et al. A population-based phenome-wide association study of cardiac and aortic structure and function. *Nature Medicine*, 2020.

Marcel Beetz, Abhirup Banerjee, and Vicente Grau. Point2mesh-net: Combining point cloud and mesh-based deep learning for cardiac shape reconstruction. In *International Workshop on Statistical Atlases and Computational Models of the Heart*, 2022.

Olivier Bernard, Alain Lalande, Clement Zotti, Frederick Cervenansky, Xin Yang, Pheng-Ann Heng, Irem Cetin, Karim Lekadir, Oscar Camara, Miguel Angel Gonzalez Ballester, et al. Deep learning techniques for automatic mri cardiac multi-structures segmentation and diagnosis: is the problem solved? *IEEE Transactions on Medical Imaging*, 2018.

Pranav Bhagirath, Maurits van der Graaf, Rashed Karim, Kawal Rhode, Christopher Piorkowski, Reza Razavi, Juerg Schwitter, and Marco Götte. Interventional cardiac magnetic resonance imaging in electrophysiology: advances toward clinical translation. *Circulation: Arrhythmia and Electrophysiology*, 2015.

Fabian Bongratz, Anne-Marie Rickmann, Sebastian Pölsterl, and Christian Wachinger. Vox2cortex: fast explicit reconstruction of cortical surfaces from 3d mri scans with geometric deep neural networks. In *Conference on Computer Vision and Pattern Recognition*, 2022.

Adrienne E Campbell-Washburn, Mohammad A Tavallaei, Mihaela Pop, Elena K Grant, Henry Chubb, Kawal Rhode, and Graham A Wright. Real-time mri guidance of cardiac interventions. *Journal of Magnetic Resonance Imaging*, 2017.

Victor M Campello, Polyxeni Gkontra, Cristian Izquierdo, Carlos Martin-Isla, Alireza Sojoudi, Peter M Full, Klaus Maier-Hein, Yao Zhang, Zhiqiang He, Jun Ma, et al. Multi-centre, multi-vendor and multi-disease cardiac segmentation: the m&ms challenge. *IEEE Transactions on Medical Imaging*, 2021.

Xiang Chen, Nishant Ravikumar, Yan Xia, Rahman Attar, Andres Diaz-Pinto, Stefan K Piechnik, Stefan Neubauer, Steffen E Petersen, and Alejandro F Frangi. Shape registration with learned deformations for 3d shape reconstruction from sparse and incomplete point clouds. *Medical Image Analysis*, 2021.

Kyunghyun Cho, Bart Van Merriënboer, Caglar Gulcehre, Dzmitry Bahdanau, Fethi Bougares, Holger Schwenk, and Yoshua Bengio. Learning phrase representations using rnn encoder-decoder for statistical machine translation. *arXiv Preprint*, 2014.

Rodrigo Santa Cruz, Leo Lebrat, Pierrick Bourgeat, Clinton Fookes, Jurgen Fripp, and Olivier Salvado. Deepcsr: A 3d deep learning approach for cortical surface reconstruction. In *IEEE Winter Conference on Applications of Computer Vision*, 2021.

A. Doi and A. Koide. An Efficient Method of Triangulating Equi-Valued Surfaces by Using Tetrahedral Cells. *IEICE Transactions on Information and Systems*, 74(1):214–224, 1991.

Haoran Dou, Seppo Virtanen, Nishant Ravikumar, and Alejandro F Frangi. A generative shape compositional framework to synthesize populations of virtual chimeras. *TNNLS*, 2024.

Jinming Duan, Ghalib Bello, Jo Schlemper, Wenjia Bai, Timothy JW Dawes, Carlo Biffi, Antonio de Marvao, Georgia Doumoud, Declan P O'Regan, and Daniel Rueckert. Automatic 3d bi-ventricular segmentation of cardiac images by a shape-refined multi-task deep learning approach. *IEEE Transactions on Medical Imaging*, 2019.

Xifeng Gao, Nikhil V Navkar, Dipan J Shah, Nikolaos V Tsekos, and Zhigang Deng. Intraoperative registration of preoperative 4d cardiac anatomy with real-time mr images. In *BIBE*, 2012.

MJW Götte, LHGA Hopman, P Bhagirath, MJB Kemme, JL Nelissen, K Lindborg, AJ Krafft, MES Sprengers, SAJ Chamuleau, and CP Allaart. First-in-Human Real-Time MR-Guided Ventricular Ablation for Idiopathic Outflow Tract Premature Ventricular Complexes. *JAMA Cardiology*, 10(11):1195–1200, 2025.

Andreas Hauptmann, Simon Arridge, Felix Lucka, Vivek Muthurangu, and Jennifer A Steeden. Real-time cardiovascular mr with spatio-temporal artifact suppression using deep learning–proof of concept in congenital heart disease. *Magnetic resonance in medicine*, 2019.

Huaifei Hu, Ning Pan, and Alejandro F Frangi. Fully automatic initialization and segmentation of left and right ventricles for large-scale cardiac mri using a deeply supervised network and 3d-asm. *Computer Methods and Programs in Biomedicine*, 2023.

Xishi Huang, Jing Ren, Gerard Guiraudon, Derek Boughner, and Terry M Peters. Rapid dynamic image registration of the beating heart for diagnosis and surgical navigation. *TMI*, 2009.

F. Isensee, P.F. Jaeger, S.A.A. Kohl, J. Petersen, and K.H. Maier-Hein. Nnu-Net: A Self-Configuring Method for Deep Learning-Based Biomedical Image Segmentation. *Nature Methods*, 18(2):203–211, 2021.

Fanwei Kong, Nathan Wilson, and Shawn Shadden. A deep-learning approach for direct whole-heart mesh reconstruction. *Medical Image Analysis*, 2021.

Lei Li, Julia Camps, Blanca Rodriguez, and Vicente Grau. Solving the inverse problem of electrocardiography for cardiac digital twins: A survey. *IEEE Reviews in Biomedical Engineering*, 2024.

Cristian A Linte, John Moore, Chris Wedlake, Daniel Bainbridge, Gérard M Guiraudon, Douglas L Jones, and Terry M Peters. Inside the beating heart: An in vivo feasibility study on fusing pre-and intra-operative imaging for minimally invasive therapy. *International journal of computer assisted radiology and surgery*, 2009.

Ze Liu, Yutong Lin, Yue Cao, Han Hu, Yixuan Wei, Zheng Zhang, Stephen Lin, and Baining Guo. Swin transformer: Hierarchical vision transformer using shifted windows. In *International Conference on Computer Vision*, 2021.

Carlos Martín-Isla, Víctor M Campello, Cristian Izquierdo, Kaisar Kushibar, Carla Sendra-Balcells, Polyxeni Gkontra, Alireza Sojoudi, Mitchell J Fulton, Tewodros Weldebirhan Arega, Kumaradevan Punithakumar, et al. Deep learning segmentation of the right ventricle in cardiac mri: the m&ms challenge. *JBHI*, 2023.

Q. Meng, W. Bai, T. Liu, D. P. O'Regan, and D. Rueckert. Mesh-Based 3D Motion Tracking in Cardiac MRI Using Deep Learning. In *Conference on Medical Image Computing and Computer Assisted Intervention*, 2022a.

Q. Meng, C. Qin, W. Bai, T. Liu, de Marvao, A. A. O'Regan, D. P, and D. Rueckert. Mulvimotion: Shape-Aware 3D Myocardial Motion Tracking from Multi-View Cardiac MRI. *IEEE Transactions on Medical Imaging*, 41(8):1961–1974, 2022b.

Q. Meng, W. Bai, D. P. O'Regan, and D. Rueckert. Deepmesh: Mesh-Based Cardiac Motion Tracking Using Deep Learning. *IEEE Transactions on Medical Imaging*, 2023.

Krishna S Nayak, Yongwan Lim, Adrienne E Campbell-Washburn, and Jennifer Steeden. Real-time magnetic resonance imaging. *Journal of Magnetic Resonance Imaging*, 2022.

Ingo Paetsch, Philipp Sommer, Cosima Jahnke, Sebastian Hilbert, Susanne Loebe, Katharina Schoene, Sabrina Oebel, Sascha Krueger, Steffen Weiss, Jouke Smink, et al. Clinical workflow and applicability of electrophysiological cardiovascular magnetic resonance-guided radiofrequency ablation of isthmus-dependent atrial flutter. *European Heart Journal-Cardiovascular Imaging*, 2019.

Esther Puyol-Antón, Bram Ruijsink, Wenjia Bai, Hélene Langet, Mathieu De Craene, Julia A Schnabel, Paolo Piro, Andrew P King, and Matthew Sinclair. Fully automated myocardial strain estimation from cine mri using convolutional neural networks. In *International Symposium on Biomedical Imaging*, 2018.

Shuang Qian, Devran Ugurlu, Elliot Fairweather, Laura Dal Toso, Yu Deng, Marina Strocchi, Ludovica Cicci, Richard E Jones, Hassan Zaidi, Sanjay Prasad, et al. Developing cardiac digital twin populations powered by machine learning provides electrophysiological insights in conduction and repolarization. *Nature Cardiovascular Research*, 4(5):624–636, 2025.

Mengyun Qiao, Kathryn A McGurk, Shuo Wang, Paul M Matthews, Declan P O'Regan, and Wenjia Bai. A personalized time-resolved 3d mesh generative model for unveiling normal heart dynamics. *Nature Machine Intelligence*, pp. 1–12, 2025.

Chen Qin, Wenjia Bai, Jo Schlemper, Steffen E Petersen, Stefan K Piechnik, Stefan Neubauer, and Daniel Rueckert. Joint learning of motion estimation and segmentation for cardiac mr image sequences. In *Conference on Medical Image Computing and Computer Assisted Intervention*, 2018.

Chen Qin, Shuo Wang, Chen Chen, Huaqi Qiu, Wenjia Bai, and Daniel Rueckert. Biomechanics-informed neural networks for myocardial motion tracking in mri. In *Conference on Medical Image Computing and Computer Assisted Intervention*, 2020.

Toby Rogers, Adrienne E Campbell-Washburn, Rajiv Ramasawmy, D Korel Yildirim, Christopher G Bruce, Laurie P Grant, Annette M Stine, Aravindan Kolandaivelu, Daniel A Herzka, Kanishka Ratnayaka, et al. Interventional cardiovascular magnetic resonance: state-of-the-art. *Journal of Cardiovascular Magnetic Resonance*, 2023.

O. Ronneberger, P. Fischer, and T. Brox. U-Net: Convolutional Networks for Biomedical Image Segmentation. In *Conference on Medical Image Computing and Computer Assisted Intervention*, pp. 234–241, 2015.

Daniel Rueckert, Luke I Sonoda, Carmel Hayes, Derek LG Hill, Martin O Leach, and David J Hawkes. Nonrigid registration using free-form deformations: application to breast mr images. *IEEE Transactions on Medical Imaging*, 2002.

Javier Sanz and Zahi A Fayad. Imaging of atherosclerotic cardiovascular disease. *Nature*, 451(7181):953–957, 2008.

T. Shen, J. Gao, K. Yin, M.-Y. Liu, and S. Fidler. Deep Marching Tetrahedra: A Hybrid Representation for High-Resolution 3D Shape Synthesis. In *Advances in Neural Information Processing Systems*, 2021.

Konstantinos Tampakis, Sokratis Pastromas, Alexandros Sykiotis, Stamatina Kampanarou, Georgios Kourgiannidis, Chrysa Pyrpiri, Maria Bousoula, Dimitrios Rozakis, and George Andrikopoulos. Real-time cardiovascular magnetic resonance-guided radiofrequency ablation: A comprehensive review. *WJC*, 2023.

Hans C Van Assen, Mikhail G Danilouchkine, Alejandro F Frangi, Sebastián Ordás, Jos JM Westenberg, Johan HC Reiber, and Boudewijn PF Lelieveldt. Spasm: a 3d-asm for segmentation of sparse and arbitrarily oriented cardiac mri data. *Medical Image Analysis*, 2006.

A Vaswani. Attention is all you need. In *Advances in Neural Information Processing Systems*, 2017.

Tom Vercauteren, Xavier Pennec, Aymeric Perchant, and Nicholas Ayache. Non-parametric diffeomorphic image registration with the demons algorithm. In *Conference on Medical Image Computing and Computer Assisted Intervention*, 2007.

Benjamin Villard, Vicente Grau, and Ernesto Zacur. Surface mesh reconstruction from cardiac mri contours. *Journal of Imaging*, 2018.

U. Wickramasinghe, E. Remelli, G. Knott, and P. Fua. Voxel2mesh: 3D Mesh Model Generation from Volumetric Data. In *Conference on Medical Image Computing and Computer Assisted Intervention*, 2020.

Yan Xia, Xiang Chen, Nishant Ravikumar, Christopher Kelly, Rahman Attar, Nay Aung, Stefan Neubauer, Steffen E Petersen, and Alejandro F Frangi. Automatic 3d+ t four-chamber cmr quantification of the uk biobank: integrating imaging and non-imaging data priors at scale. *Medical Image Analysis*, 2022.

Jia Xiao, Wen Zheng, Wenji Wang, Qing Xia, Zhennan Yan, Qian Guo, Xiao Wang, Shaoping Nie, and Shaoting Zhang. Slice2mesh: 3d surface reconstruction from sparse slices of images for the left ventricle. *IEEE Transactions on Medical Imaging*, 2024.

Robert Xu, Prashant Athavale, Adrian Nachman, and Graham A Wright. Multiscale registration of real-time and prior mri data for image-guided cardiac interventions. *IEEE Transactions on Biomedical Engineering*, 2014.

Jiancheng Yang, Ekaterina Sedykh, Jason Ken Adhinarta, Hieu Le, and Pascal Fua. Generating anatomically accurate heart structures via neural implicit fields. In *Conference on Medical Image Computing and Computer Assisted Intervention*, 2024.

Meng Ye, Dong Yang, Mikael Kanski, Leon Axel, and Dimitris Metaxas. Neural deformable models for 3d bi-ventricular heart shape reconstruction and modeling from 2d sparse cardiac magnetic resonance imaging. In *International Conference on Computer Vision*, 2023.

Xiaohan Yuan, Cong Liu, and Yangang Wang. 4d myocardium reconstruction with decoupled motion and shape model. In *International Conference on Computer Vision*, 2023.

A Datasets and Metrics

A.1 Unified Dataset

ACDC features CMR sequences of about 30 frames for 100 patients, each one covering at least one cardiac cycle and with a slice thickness of 5-10 mm. M&Ms and M&Ms-2 are made of multi-center and multi-vendor collected CMR sequences for 375 and 360 subjects, respectively. While featuring similar resolutions and sequence lengths as ACDC dataset, these two datasets cover a broader range of cardiac pathologies. For these datasets, segmentation masks for the left ventricle (LV), right ventricle (RV), and myocardium (MYO) are provided, but only for the end-diastolic (ED) and end-systolic (ES) frames. As in earlier studies Meng et al. (2023); Yuan et al. (2023); Xiao et al. (2024); Qiao et al. (2025), we register and fit a Statistic Shape Model (SSM) Duan et al. (2019); Bai et al. (2015) to segmentation mask at the ED and ES frames using non-rigid image registration methods to generate cardiac meshes.

A.2 iCMR Dataset

The iCMR dataset was collected at the Center for Interventional MRI of our collaborating university hospital. The infrastructure used for image acquisition is shown in Fig. 1(c). For each subject, a stack of short-axis (SAX) breath-hold cine slices was acquired at rest on a 1.5T interventional system (Siemens Healthineers, Magnetom Aera) using an accelerated steady-state free precession (SSFP) sequence covering the entire left ventricle. In addition, single SAX cine slices were acquired at mid-ventricular levels with slight angulations relative to the full SAX stack. Because both heart shape and rate may vary during intervention, updating of the 3D heart model is often required to ensure precise intracardiac manipulations. To mimic heart-rate changes, five volunteers exercised inside the interventional MRI (bending and extending their legs in two parallel plastic gutters for 2–3 minutes), thereby increasing their heart rates by approximately 20–30 beats per minute. Once elevated, single SAX cine slices were reacquired at the same mid-ventricular levels as during rest. For real-time acquisition, the online sequences had lower spatial resolution than the offline short-axis CMR stacks to meet the latency constraint.

A.3 Evaluation Metrics

For evaluation, we compare the predicted 3D mesh obtained by deforming the ED frame to the ES frame with the ground-truth 3D mesh at the ES frame. We use L2-norm Chamfer Distance (CD) to measure 3D reconstruction quality, which is calculated as:

$$\mathrm{CD}(P, Q) = \frac{1}{|P|} \sum_{p \in P} \min_{q \in Q} \|p - q\|_2^2 + \frac{1}{|Q|} \sum_{q \in Q} \min_{p \in P} \|q - p\|_2^2.$$

Dice score is also reported to quantify the intra-slice segmentation results, calculated as:

$$\mathrm{DSC}(A, B) = \frac{2|A \cap B|}{|A| + |B|}.$$

Besides these quality measures, we also report the Mean-Absolute Error (MAE) of LV/RV ES Volume (LVESV/RVESV), LV/RV ejection fraction (LVEF/RVEF) derived from our predicted meshes. The ejection fraction is calculated as

$$XEF = \frac{XEDV - XESV}{XEDV}, X \in \{LV, RV\},$$

here EDV denotes the volume of the ED meshes. For our method, we use the predicted ED meshes from the reconstruction branch for calculation. For baselines require ground-truth input meshes, we directly use the input meshes for calculation.

The MAE of ES volume and ejection fraction is then calculated as

$$MAE_{esv} = |ESV_p - ESV_{gt}|, \ MAE_{ef} = |EF_p - EF_{gt}|.$$

Algorithm 1: Adaptive Feature Aggregation (AFA)

Input : Reference features F_{2d}^0, Target features F_{2d}^t (voxels $x \in F_{2d}^0$)
Input : Hyperparameters k_1 (nearest slices), k_2 (nearest positions per slice)
Output: Aggregated motion features V_{motion}^t

---

1 foreach *voxel x in F_{2d}^0* **do**
       // Step 1: Identify spatially nearest 2D slices
2 $S \leftarrow$ Select k_1 nearest 2D slices from F_{2d}^t relative to x using point-to-plane distance
       // Step 2: Retrieve local features within each slice
3 $P \leftarrow \emptyset$
4   **foreach** *slice $s \in S$* **do**
5 $p_s \leftarrow k_2$ spatially closest positions in slice s to voxel x using point-to-point distance;
           $P \leftarrow P \cup$ Features at p_s
6   **end**
       // Step 1&2 together works as the NN-Select(F_{2d}^t, k_1, k_2) operator

       // Step 3: Define Q, K, V for localized attention
7 $Q \leftarrow F_{2d}^0(x)$ // Query is the current voxel feature
8 $K \leftarrow P, V \leftarrow P$ // Keys and Values from collected set P
       // Step 4: Incorporate position embedding and compute MHA
9 $Q, K \leftarrow$ Apply learnable position embedding(Q, K)
10 $V_{\text{motion}}^t(x) \leftarrow$ MHAttention(Q, K, V)
11 end
12 return V_{motion}^t

The ground-truth volume and ejection fraction are derived from the ground-truth meshes. LVEF and RVEF are representative cardiac function indexes which could be used for clinical diagnosis.

B Implementation Details

In all experiments, the slices are resampled by linear interpolation to a spacing of 1.25×1.25 mm. From this, the AFA module constructs 3D volume with an effective spacing of $1.25 \times 1.25 \times 2$ mm. The AFA module uses a 8-head multi-head attention with learnable 3D position embedding. In the reconstruction branch, a modified nnU-Net Isensee et al. (2021) is used as the image encoder. It comprises five downsampling and upsampling blocks with channel sizes [32, 64, 128, 256, 320]. We use the same GCN structure as in Shen et al. (2021) with 3 layers and 128 channels. The initial tetrahedra resolution is 128. For the motion branch, we use the identical encoding network structure as the reconstruction branch. For the deformation model, the hidden dimension of the GCN and GRU are both set to 128. We use SGD optimizer with an initial learning rate of 0.01, a weight decay of $3e - 5$, and momentum of 0.99. We train all models for 300/150 epochs for the reconstruction/motion learning stage. All experiments are conducted using one V100 GPU.

AFA Module. We provide the workflow of how the AFA module is executed in Algorithm 1.

C Baselines

We compare our method against several existing approaches for 4D shape reconstruction. For the online interventional setting, we adapted several baselines so that they can operate under sparse inputs. **4DMR** Yuan et al. (2023) is an implicit representation method that reconstructs myocardium mesh sequence by optimizing decoupled shape and motion latent codes at test time, minimizing the discrepancy between the predicted mesh and ground-truth contour points. We adapt it by restricting the optimization to incomplete contours extracted from the available few slices, while still using full contours for the initial frame. **MR-Net** Chen et al. (2021) is a template-mesh-based approach that takes ground-truth contour points as input and leverages a hybrid PointNet + 3D CNN architecture to predict vertex deformations. Since MR-Net directly

encodes point sets, it can naturally accommodate incomplete inputs, a scenario already considered in the original paper. Although originally designed for static reconstruction, we adapt it for motion estimation by setting the ED-frame mesh as the template. **MulViMotion** Meng et al. (2022b) and **DeepMesh** Meng et al. (2023) are mesh-based approaches that predict deformation fields via 3D CNNs, but only for the myocardium class. They rely on feature concatenation, requiring the initial and target frame to have same number of slices. We add zero-valued slices to the target frame to enable it to handle few-slice input. Note this modification only allow them to handle the issue of inconsistent slice numbers. It can not solve problems like scanning angle discrepancy. For fair comparison, we adopt their SAX-view-only versions as mentioned in the original papers.

In addition, we include two modified baselines derived from shape reconstruction methods: MeshDeform-Net Kong et al. (2021) and DeepCSR Cruz et al. (2021). To adapt them for motion reconstruction, we integrated into both the same nnU-Net backbone, AFA module, deformation model, and training pipeline as our method. Since their main differences with our full framework lie in the shape representation, we denote them as **Ours-Mesh** and **Ours-SDF**.

For the traditional offline setting, in addition to the above methods, we further compare against B-spline Free Form Deformation (FFD) Rueckert et al. (2002), diffeomorphic Demons (dDemons) Vercauteren et al. (2007), MR-Net w. nnU-Net, MulViMotion and DeepMesh. **FFD** and **dDemons** are classical registration algorithms that have been widely used in recent cardiac motion tracking studies Bai et al. (2020); Qin et al. (2020); Puyol-Antón et al. (2018). **MR-Net w. nnU-Net** adopts the MR-Net architecture but replaces ground-truth contours with those predicted by nnU-Net, featuring a two-stage automated pipeline.

Unless explicitly stated otherwise, all methods are jointly trained on the unified dataset.

Figure 7: Full-Slice Comparison. (Left) We predict 3D meshes by deforming from ED to ES frame. Color indicates the magnitude of point-to-surface error in mm. (Right) Segmentation results on slices at the apex and base location. Note 4DMR and DeepMesh can only output for myocardium.

D Complementary Experiment Results

Disease Definition. A normal heart should satisfy the following criteria: LV EF $> 50\%$ and RV EF $> 40\%$. A subject with dilated cardiomyopathy should have LV EF $< 40\%$. This is not a strict definition; a more rigorous classification criterion would require a comprehensive analysis of the ratio between volume and body surface area, along with other indicators. However, verifying whether the predicted ejection fraction from *TetHeart* satisfies this approximate criterion can further demonstrate the practicality of our model.

As could be seen from Fig. 4, even with a single slice, *TetHeart* could predict cardiac function indexes that meet the criteria. This highlights our model's practicality.

Method	Training Dataset				Myocardium CD (mm^2) ↓		
	ACDC	M&Ms	M&Ms-2	4DM	1-Slice	5-Slice	Full-Slice
FFD	-	-	-	-	-	-	11.51
dDemons	-	-	-	-	-	-	12.46
4DMR	✓	✓	✓	-	24.00	19.03	13.97
MR-Net	✓	✓	✓	-	15.87	11.56	7.97
DeepMesh	✓	✓	✓	-	-	-	8.32
Ours-Mesh	✓	✓	✓	-	13.31	10.32	5.97
Ours-SDF	✓	✓	✓	-	13.78	10.78	6.24
	✓	-	-	-	25.43	19.42	13.81
Ours	-	✓	-	-	16.91	12.54	8.23
	-	-	✓	-	17.88	13.03	8.83
	✓	✓	✓	-	**11.33**	**9.08**	**5.51**
Ours (Oracle)	-	-	-	✓	6.24	4.87	2.70

Table 6: *Zero-shot evaluation results on public 4DM dataset.* Check marks denote the dataset used for training. Oracle is trained and evaluated on 4DM.

Method	4DMR	MR-Net	Ours-Mesh	Ours-SDF	Ours
GLS MAE (1-Slice, %)	4.47	3.73	2.65	3.08	**2.18**

Table 7: *Strain analysis results on M&Ms dataset.*

Visualization Results Using Full Slices. In Fig. 7, we show qualitative results on the same hearts as in Fig. 3. But now using full slices rather than using a single slice.

Statistical Significance Evaluation. We perform paired t-test to evaluate the statistical significance of our method. Take the 1-Slice setting of Tab. 1 as reference, our method achieves statistically significant improvements ($p < 0.05$) when compared the average Chamfer Distance to all competitors. Specifically, 4DMR, MR-Net, MulViMotion, DeepMesh have p-values smaller than 0.001, and Ours-Mesh ($p = 0.0347$), Ours-SDF ($p = 0.0213$). The significance test results are similar for other comparisons.

Strain Analysis. Figure 3 and 7 justify the plausibility of cardiac motion by measuring volumetric changes and the myocardial volume curves partially reflect the quasi-incompressibility. We further verify the accuracy of the estimated motion using ED-to-ES strain analysis. The Mean Absolute Error of Global Longitudinal Strain in reported in Tab. 7. As could be seen, ours performs best. This further validates the effectiveness of our method.

E External Evaluation Results

4DM Dataset. As 4DM provides full-sequence annotations, but only for the myocardium. we evaluate the full-sequence prediction performance for the myocardium subclass, rather than focusing solely on the ES frame and report the results in Tab. 6. Here, check marks denote the dataset used for training. *Oracle* refers to training and testing directly on 4DM, serving as an upper bound for performance.

As shown in the table, our model demonstrates the best generalization abilities, achieving top performance in both few-slice and full-slice settings. When compared to the best competitor MR-Net, our model generalizes significantly better, and achieves 30% improvement. In contrast, 4DMR exhibits poor generalization, likely due to significant distribution shifts across datasets and the limited representational capacity of its simple MLP architecture. While MR-Net and DeepMesh show some generalization ability, they still perform worser than us.

Consistent with the findings from previous evaluations, training with a larger and more diverse dataset improves generalization. and representing the heart as tetrahedra enables more effective spatial information retention and exchange, particularly in scenarios with limited slice input.

standard slice, 256x208   interventional slice, 208x168

Figure 8: Normal SAX and interventional slice comparison.

F Extended Ablation Studies

The Impact of AFA Configuration. In Tab. 8, we show the impact of the number of slices in the AFA module and the number of selected positions per slice on both performance and speed (measured on an NVIDIA V100 GPU). As shown in the table, our default configuration achieves a good balance between accuracy and efficiency. Further increasing the number of slices or selected positions brings only limited performance gains but leads to a slowdown. One thing worth mentioning is that to implement our AFA module we use a relatively simple strategy of sorting and selecting the top-k elements. Therefore, for $|S| = 1, |G| = 5^2$ and $|S| = 3, |G| = 3^2$, although the two configurations use a similar number of features, the latter runs slower because it performs sorting multiple times. More advanced strategies or precomputation could potentially narrow this gap for further acceleration.

Model	$|S|$	$|G|$	Myocardium CD & FPS ($mm^2 \downarrow$ & $img/s \uparrow$)					
			1-Slice		5-Slice		Full-Slice	
	1	1^2	10.34	15.1	8.45	15.1	4.56	15.1
	1	3^2	9.98	15.0	8.22	14.9	4.43	14.9
	1	5^2	9.93	14.8	8.13	14.7	4.39	14.6
	3	1^2	10.32	15.0	8.19	13.0	4.37	13.0
TetHeart	3*	3^2*	9.76	14.9	7.93	12.1	4.11	12.0
	3	5^2	9.73	14.7	7.86	10.8	4.05	10.7
	5	1^2	10.27	15.0	8.17	11.9	4.33	11.8
	5	3^2	9.74	14.8	7.81	10.4	4.09	10.4
	5	5^2	9.72	14.7	7.83	8.7	4.10	8.7
	7	5^2	9.76	14.6	7.85	8.6	4.07	7.3

Table 8: *Impact of different AFA configuration evaluated on the M&Ms dataset. $|S|$: the number of slice. $|G|$: the number of positions selected from each slice. *: default setting.*

Category-Wise Ablation Results. In Tab. 9, we present more detailed results on the unified dataset as a supplement to Tab. 5 in the main text. As shown in the table, across all datasets and all categories, the conclusions remain consistent with those in the main text: the design of the AFA module, parameter sharing, and the use of distillation loss all contribute to improving the final performance of our model.

Rationale Behind the Training Strategy. In Sec. 3.5, when training the motion branch, we propose an aggressive augmentation strategy by using slice subsampling and random downsampling/rotation/noise jittering to mimic the interventional scenario. The rationale behind this strategy is given in Fig. 8. As could be seen, in terms of image quality, the interventional slice sample of iCMR dataset differs from standard SAX mainly in resolution and sharpness. So our proposed training strategy could effectively bridge the domain gap between normal SAX slices and interventional slices. This is also validated by the ablation results

Method	Myocardium CD (mm^2) ↓			LV CD (mm^2) ↓			RV CD (mm^2) ↓		
	1-Slice	5-Slice	Full-Slice	1-Slice	5-Slice	Full-Slice	1-Slice	5-Slice	Full-Slice
ACDC									
Ours	15.24	10.22	6.63	23.65	18.31	11.51	40.64	32.61	22.15
- AFA module	-	-	6.55	-	-	11.58	-	-	22.30
- position embedding	16.60	11.33	6.84	25.32	20.14	12.73	42.23	34.19	23.59
- image feature in query Q	24.10	15.96	9.39	36.76	28.39	17.44	58.31	43.21	32.32
- shared encoding network	32.23	15.24	10.01	49.12	28.99	18.59	75.92	44.81	34.44
- distillation loss	16.13	10.48	6.49	25.07	18.78	11.28	43.08	33.43	21.71
M&Ms									
Ours	9.76	7.93	4.11	12.37	9.72	4.85	21.42	17.39	13.68
- AFA module	-	-	4.22	-	-	4.79	-	-	13.74
- position embedding	11.23	9.04	4.67	13.35	10.34	5.67	22.72	18.84	14.93
- image feature in query Q	16.67	12.37	6.44	19.76	14.17	7.82	33.63	23.81	20.60
- shared encoding network	20.66	11.63	6.43	24.56	14.24	7.77	41.80	24.12	20.45
- distillation loss	10.34	8.12	4.03	13.11	9.94	4.75	22.70	17.79	13.41
M&Ms-2									
Ours	10.12	7.30	4.02	13.89	9.96	5.04	20.27	15.58	10.82
- AFA module	-	-	4.17	-	-	5.07	-	-	10.76
- position embedding	11.93	8.41	4.74	14.98	10.40	5.69	21.91	16.72	11.65
- image feature in query Q	17.34	11.38	6.31	22.32	13.57	7.57	32.65	21.40	15.49
- shared encoding network	21.49	11.09	6.54	26.96	13.64	7.85	39.44	20.73	16.07
- distillation loss	10.95	7.47	4.11	15.01	10.19	5.05	21.89	15.94	11.06

Table 9: *More comprehensive ablation study results on the unified dataset.*

in Tab. 5(left), where the proposed training strategy could effectively handle the simulated interventional artifacts.

G   More Visualization Results

We provide video results in the 'video' subfolder. 'single_and_full_slice_comparison.mp4' presents video results for the NOR and DCM samples used in Fig. 4. 'icmr_results.mp4' presents video results for the samples from iCMR samples used in Fig. 6, which uses a single slice with slight angulations relative to the full SAX stack for motion inference. We also provide the video results for more diseases, including Myocardial Infarction (MINF), Abnormal Right Ventricle (ARV), in folder 'video/more_sequences'. Each subject is named 'XX_sYY.mp4', where XX is the disease category and YY is the number of slices used to recover motion.

H   Limitation

Our method achieve promising results on both full-stack and few-slice settings, surpassing previous methods. However, by examining our own model, we can observe that although we have proposed several techniques, there still remains a performance gap between motion reconstruction using only a few slices versus the full stack. We plan to introduce physical constraints as prior knowledge into the model to reduce this gap. Another choice is to use a generative model to generate the mesh sequence progressively to complement our current feed-forward reconstruction process. However we need to strike a good balance between the reconstruction speed and quality if we are targeting real-time applications.

While the distillation training strategy could elevate the performance, it also has a risk of introducing population-level statistics to hallucinate anatomy, producing 3D cardiac shapes with averaged features. Such a tendency might obscure unique pathological anomalies. Our framework's flexibility, i.e. arbitrary inputs, allows a possible mitigation: dynamically adjusting scan planes to capture missing details during intervention. Although effective integration of this temporal data requires a dedicated temporal module, we view this as a valuable future extension. We leave these as future works.