# Generative Adversarial Networks for Coronary CT Angiography Acquisition Protocol Correction with Explicit Attenuation Constraints

## Abstract

The image quality of coronary CT angiography (CCTA) is important for the correct diagnosis of patients with suspected coronary artery disease, which is heavily influenced by image acquisition. Timing of the contrast media injection specifically influences the level of arterial enhancement, and it is aimed to allow optimal assessment of the coronary artery morphology. However, a consensus on an optimal acquisition protocol that can account for the large variety in patient cohorts has not been reached, commonly resulting in suboptimal arterial enhancement. In this work, we propose a generative adversarial network for the retrospective correction of contrast media attenuation in CCTA, thus reducing the dependency on an optimal timing protocol at acquisition. We develop and evaluate the method in a set of 1,179 CCTA scans with varying levels of contrast enhancement. We evaluate the consistency of intensity values in the coronary arteries and evaluate performance of coronary centerline extraction as a commonly performed analysis task. Results show that correction of contrast media attenuation values in CCTA scans is feasible, and that it improves the performance of automatic centerline extraction. The method may allow improved analysis of coronary arteries in CCTA scans with suboptimal contrast enhancement.

**Keywords:** Contrast media timing, coronary attenuation, coronary CT angiography, deep learning, generative adversarial networks

## 1. Introduction

Coronary CT angiography (CCTA) is an effective modality for the non-invasive diagnosis of (suspected) coronary artery disease (Hoffmann et al., 2012; Mortensen et al., 2020). Through contrast enhancement it visualizes the coronary artery lumen, and allows for the evaluation of arterial stenosis and atherosclerotic plaque. These are important for patient management (Cury et al., 2016). However, clinical assessment of CCTA scans heavily relies on scan quality, which is affected by a number of factors, including timing of contrast media injection, cardiac motion, and blooming artifacts (Dey et al., 2008). Insufficient arterial enhancement ($\leq 300$ Hounsfield Units (HU)) typically leads to unreliable and often overestimated evaluation of the stenosis degree (Cademartiri et al., 2006, 2008). Conversely, elevated levels of coronary enhancement (i.e. $\geq 500$ HU) may cause beam hardening and partial volume effects, resulting in an underestimation of stenosis degree (Horiguchi et al., 2007; Fei et al., 2008). Ideally, for the analysis of coronary arteries a contrast medium injection protocol should always result in coronary artery attenuation values approaching 400 HU. Consistency in contrast enhancement of the coronary arteries in CCTA scans depends on the chosen image acquisition protocol. Several contrast media injection protocols

optimizing coronary enhancement were proposed (Fleischmann, 2010; Kawaguchi et al., 2014; Sandfort et al., 2020), but they may be limited in accounting for differences in patient cohort characteristics.

As an alternative to the optimization during image acquisition, recent advances in deep learning have allowed for the retrospective adjustment of previously acquired medical images. Generative modelling has been explored to address image quality in various images and modalities (Ghekiere et al., 2017; Wolterink et al., 2021). Specifically, in CCTA scans, generative models have been used to reduce image noise (Gondara, 2016; Kang et al., 2019), reduce motion artifacts (Jung et al., 2020), and perform super-resolution (Sun and Ng, 2022). However, the correction of the contrast media enhancement levels in CCTA has seen little attention thus far.

Hence, we propose a method for the retrospective adjustment of coronary artery attenuation values in CCTA. We employ a generative adversarial network (GAN) which, given a CCTA scan with suboptimal contract enhancement, synthesizes a CCTA scan with contrast intensity levels in an optimal range (Goodfellow et al., 2020). We develop and evaluate the method with CCTA scans of 1,179 patients suspected of coronary artery disease. To evaluate the performance of the method, we compare the consistency of the intensity values in the main coronary arteries before and after arterial contrast correction. Moreover, we compare performance of coronary artery centerline extraction, a commonly performed downstream task in scans before and after correction.

## 2. Data

### 2.1. Patient and Image Data

This study retrospectively included 1,179 CCTA scans of patients with suspected coronary artery disease from clinical routine in ********************************************** ********. The need for informed consent was waived by the Institutional Ethical Review Board. CCTA exams were acquired on a Siemens Somaton Force CT scanner (Siemens Healthineers, Erlangen, Germany). The tube voltage ranged from 70 to 120 kVp and the tube current ranged from 296 to 644 mAs. Scans were ECG triggered and have an in-plane resolution of 0.26-0.46 mm$^2$ and a slice thickness and increment of 0.6 mm. During acquisition, contrast medium was injected for optimal enhancement of the coronary arteries.

For all patients, coronary artery ostia and centerlines were extracted using the **** **************************** described by (***************). The mean and standard deviation of the attenuation at the aortic root were determined by extracting and evaluating 3D patches of size 9.5 mm × 9.5 mm × 9.5 mm centered around the ostia. This was done by identifying the Gaussian distribution describing the contrast agent in the patch histogram (Figure 1). Based on the mean contrast attenuation $\mu_{HU}$, scans were subsequently divided into a group of *optimal* CCTA scans ($300 < \mu_{HU} < 500$) and a group of *suboptimal* CCTA scans ($\mu_{HU} \leq 300$, $\mu_{HU} \geq 500$). This resulted in a total of 402 optimal and 777 suboptimal scans.

A total of 96 patients with suboptimal levels of contrast enhancement were prospectively set apart for evaluation. For this set, automatic coronary artery lumen and myocardium segmentations were obtained through the respective methods described by (***************) and (***************).

$$HU_{Contrast} \sim \mathcal{N}(\mu, \sigma^2)$$

Figure 1: Suboptimal contrast detection pipeline. Coronary ostia are first extracted from the CCTA through an automated method (***************). A 3D patch is subsequently extracted around both ostia, for which the intensity histogram is computed (right). By representing the histogram as a mixture of Gaussians, the histogram peak describing the contrast media can be identified. The value of this peak is used to label CCTA scan as either suboptimal or optimal.

## 2.2. Reference Standard

To evaluate whether automatic centerline tracking was successful, a trained observer manually placed markers in the three major coronary arteries (i.e. the right coronary artery (RCA), the left anterior descending (LAD), and the left circumflex (LCx)) in CCTA scans from the hold-out test set comprising 96 scans (see Section 2.1). These markers were placed along the centerline of the coronary lumen for each artery starting from the ostia, resulting in a total of 1,148 markers.

## 3. Method

We propose a GAN for the automatic adjustment of contrast in suboptimally acquired CCTA scans. GANs typically consist of two main parts: a generator network and a discriminator network. For our purposes, the generator generates images with optimal contrast from images with suboptimal contrast, while the discriminator tries to distinguish generated images from real optimal images. Once trained, the generator accepts axial CCTA slices as an input and returns a synthesized version of the slice with a contrast level within the optimal range of approximately 400 HU.

Inspired by previous work for coronary artery calcium (CAC) scoring that features decomposition of non-contrast CT with CAC into an image without CAC and a corresponding CAC map (Van Velzen et al., 2022), we assume that suboptimally acquired CCTA scans ($I_{sub}$) can be decomposed into a CCTA scan with optimal contrast ($I_{opt}$) and a super-

Figure 2: Overview of the proposed methodology. A generative model $G$ analyzes CCTA images acquired with suboptimal contrast media timing ($I_{sub}$), and predicts a contrast offset mask ($M_{off}$) that is subtracted from the input image to obtain a synthetic optimally acquired image ($\hat{I}_{opt}$). Optimal contrast is enforced by constraining contrast values in a coronary artery centerline mask ($M_{ctl}$) to operate within a certain range of HU values ($\mathcal{L}_{HU}$). Realistic images are encouraged by training a discriminator $D$ to distinguish real optimally acquired CCTA images ($I_{opt}$) from generated ones with $\mathcal{L}_G$ and $\mathcal{L}_D$, while similarity between generator input and output is encouraged through $\mathcal{L}_{sim}$.

imposed contrast offset mask ($M_{off}$), i.e. $I_{sub} = I_{opt} + M_{off}$. As such, the task of the generator is simplified to only compute the difference between the two CT scans. The proposed decomposition is illustrated in Figure 2.

The generator architecture used to produce $M_{off}$ employs a CNN backbone consisting of 6 ResNet-blocks, which has been shown to be capable of performing high-quality image-to-image translation (He et al., 2016; Zhu et al., 2017). For the discriminator network we employ a PatchGAN, which learns to classify whether $70 \times 70$ overlapping patches are real ($I_{opt}$) or fake ($\hat{I}_{opt}$) images with an optimal contrast level (Isola et al., 2017). The fully convolutional nature of PatchGAN allows it to operate on arbitrary image sizes while requiring fewer parameters than a full-image discriminator. Since the discriminator produces an output matrix for the overlapping patches, image labels are projected to match the output size of the discriminator network before computation of the loss.

The loss function employed in this work comprises several terms for the generation of realistic optimally acquired CCTA scans, and is defined as:

$$\mathcal{L} = \mathcal{L}_G + \mathcal{L}_D + \mathcal{L}_{sim} + \mathcal{L}_{HU} \tag{1}$$

Here, $\mathcal{L}_G$ and $\mathcal{L}_D$ describe the adversarial loss used in GAN optimization. The similarity loss $\mathcal{L}_{sim}$ directly minimizes differences between $I_{sub}$ and $\hat{I}_{opt}$, and is defined as the negative normalized cross-correlation between the two images (De Vos et al., 2019). Hence,

$$\mathcal{L}_G, \mathcal{L}_D = \min_G \max_D \mathbb{E}_{I_{opt} \sim p_{\text{data}}(opt)}[\log D(I_{opt})] + \mathbb{E}_{I_{sub} \sim p_{\text{data}}(sub)}[1 - \log D(G(I_{sub}))] \tag{2}$$

$$\mathcal{L}_{sim} = -NCC(I_{sub}, \hat{I}_{opt}) \tag{3}$$

The combination of these losses ensures the generation of images that feature the high-frequency details derived from the input scan, yet appear similar to optimally acquired CCTA scans.

Since we are interested in enforcing a specific range of contrast attenuation in the coronary arteries and by extension the aortic root, the final constituent term of the loss function $\mathcal{L}_{HU}$ explicitly penalizes outliers in the optimal contrast range along coronary artery centerlines. This is achieved by providing a mask for each input CCTA slice in which centerline point locations are marked. The mask of $2 \times 2$ voxels is placed around the identified centerline point. By applying this mask to $\hat{I}_{opt}$, predicted attenuation values along coronary artery centerlines are extracted ($\hat{I}_{ctl}$) and subsequently penalized when contrast values are not within an optimal range. This results in the following loss component:

$$\mathcal{L}_{HU} = \sum_{j=1}^{N}(\min(\hat{I}_{ctl}^{j}, HU_{min}) - HU_{min})^2 + \sum_{j=1}^{N}(\max(\hat{I}_{ctl}^{j}, HU_{max}) - HU_{max})^2 \tag{4}$$

where $N$ is the number of masked centerline voxels in a training batch, and $HU_{min}$ and $HU_{max}$ describe a lower and upper bound for the expected contrast range. The implementation of the presented method is publicly available[1].

## 4. Evaluation

The proposed method is evaluated qualitatively by observing the shifts in frequency distribution of attenuation values in the CT image. This is performed for three cardiac structures: the ostia, the coronary artery lumen, and the myocardium. We further evaluate the method quantitatively by assessing the effect of contrast enhancement correction on automatic coronary artery centerline tracking. Centerline tracking typically fails when contrast levels are outside of an expected range, and should therefore benefit from contrast correction. We perform centerline tracking in the original and in the corrected images and evaluate the completeness of the extracted coronary artery trees. The recall of manually placed markers

---

1. https://github.com/withheld/withheld

Figure 3: Attenuation distributions in the coronary ostia, coronary arteries, and the myocardium, before and after applying the method. The top row describes the test set images for which contrast level was below the optimum range ($\leq 300$ HU) in the original scan, while the bottom row describes images which had a contrast level above the optimum range ($\geq 500$ HU).

along the major coronary arteries is evaluated before and after GAN correction. For recall, a true positive is defined as a manually identified marker being within a 5 mm radius of an automatically extracted centerline point. Finally, we compare attenuation values at marker locations before and after the correction.

## 5. Experiments and Results

### 5.1. Experimental Setting

The 1,083 images in the training set were divided into two subsets: 80% for training and 20% for validation. Performance on the correction of suboptimal contrast levels in the validation set was tracked during training, allowing for hyperparameter tuning and the selection of the best-performing generative model.

The generator network was trained on minibatches, which were balanced to contain axial patches cropped around coronary artery centerlines and patches cropped around random locations in the image. Batches were further balanced to contain both patches below and above the optimal range of arterial enhancement. Training patches were of size $128 \times 128$ voxels, and patches cropped around centerlines were randomly rotated or flipped with $p = 0.5$. Batches for training the discriminator were sampled similarly, but were balanced with respect to suboptimal and optimal acquisition. Image values were divided by a factor of 600 to normalize GAN inputs. In tandem with the tanh activation function of the generator, this allows for a maximum contrast offset of 600 HU.

During training, a batch size of 64 was employed for both the generator and the discriminator. Both networks were optimized using the Adam optimizer with a learning rate of 2e-4 and parameters $\beta_1 = 0.5$ and $\beta_2 = 0.999$ (Kingma and Ba, 2014). Networks were trained

Figure 4: Results for coronary artery centerline extraction. (a) Recall of automatic coronary artery centerline extraction for manually placed markers along the three major coronary arteries in the test set. Bars indicate centerline extraction recall before (blue) and after (orange) correction of suboptimal contrast enhancement. (b) Attenuation values at marker locations in HU, before (blue) and after (orange) correction.

for a total of 100,000 iterations, with the learning rate being multiplied with a factor of $\gamma = 0.1$ at iterations [60,000; 80,000]. We further defined $HU_{min}$ and $HU_{max}$ to be 350 HU and 450 HU respectively to ensure adequate attenuation along coronary arteries. At test time, the generator processes full axial CCTA input slices. All methods and optimizations were implemented with the PyTorch deep learning library (Paszke et al., 2019).

### 5.2. Results

The frequency distribution shifts of attenuation values in three cardiac structures are shown by the histogram plots in Figure 3, displaying distributions before and after applying the method. For both the coronary ostia and arteries, a shift of intensity values resulting from the correction is observed towards the contrast distribution in optimally acquired scans. According to expectation, a shift is not observed in the myocardium, which is largely unaffected by contrast enhancement. Two examples of CCTA slice correction are presented in the Appendix (Figure 5).

Figure 4 displays the impact of our contrast correction on automatic coronary artery centerline extraction performance. The overall recall of manually identified markers in the test set was 88% before correction and 90% after correction. For both the LAD and the LCx, an increase in recall is observed as well, while a slight decrease in performance is observed for the RCA. Average and standard deviation of the attenuation values at marker locations were $423 \pm 182$ HU before and $405 \pm 101$ HU after correction.

## 6. Discussion and Conclusion

In this work, a method for the correction of contrast media enhancement in CCTA scans was proposed. The method leverages a GAN to learn a difference map between suboptimal and optimal levels of contrast enhancement, and is constrained to force attenuation values along coronary artery centerlines within a specific range. This allows for the adjustment of previously acquired CCTA scans, thus rendering a realistic image within the optimal range of contrast attenuation.

The qualitative evaluation reveals that the method learns to accurately target volumes affected by contrast enhancement, and does not influence the attenuation level in tissue which is not directly impacted by angiography (Figure 3). Furthermore, the results indicate that attenuation value distribution shifts in the coronary ostia and coronary arteries result in a distribution which closely resembles real CCTA scans with an optimal level of contrast enhancement. Though the myocardium is eventually affected by contrast enhancement as well, training did not differentiate between early and late enhancement scans, which may be an interesting topic for future research.

Results on the extraction of coronary artery centerlines indicate a better overall recall for contrast-corrected scans. The largest improvement is observed in the LCx, which may be attributed to the fact that only a small percentage of the population has a left-dominant coronary circulation. As such, the LCx is typically smaller and more difficult to distinguish, therefore benefiting more from contrast correction. The RCA on the other hand is typically subject to motion artifacts, which our method does not correct for. This may explain why no improvement is found for the recall of RCA markers.

A limitation of the proposed method may be that it only processes 2D axial slices. Consequently, the network is only able to learn the appearance of arterial structures in a 2D plane, therefore missing potentially valuable 3D information on the connectivity of coronary arteries. Furthermore, a larger field of view may help the model to identify other cardiac structures relevant for attenuation correction. Nonetheless, results indicate that explicit penalization of values along coronary artery centerlines helps the model in finding an optimal distribution of contrast enhancement that ignores most 2D noise. Future work could investigate whether a 3D approach may help the method in creating even better corrections for contrast enhancement.

Given that the method employs a GAN, it allows for training with unpaired data. Obtaining paired data for training would be impossible because voxel level spatial alignment of a patient is infeasible. As such, it offers advantages over standard CNN architectures requiring end-to-end data. Future work might investigate whether advanced synthesis methods may contribute to performance improvement (Ho et al., 2020; Rombach et al., 2022).

In conclusion, this study presented a method for the retrospective correction of CCTA scans with suboptimal levels of coronary artery contrast enhancement by generating contrast offset masks using a GAN. Contrast enhancement correction is shown to be feasible and produce realistic images within an optimal range of attenuation values. The method may extend applicability of automatic coronary artery analysis to CCTA scans with suboptimal contrast enhancement.

## References

Filippo Cademartiri, Nico R Mollet, Pedro A Lemos, Fransesco Saia, Massimo Midiri, Pim J de Feyter, and Gabriel P Krestin. Higher intracoronary attenuation improves diagnostic accuracy in MDCT coronary angiography. *American Journal of Roentgenology*, 187(4): 1115, 2006.

Filippo Cademartiri, Erica Maffei, Anselmo Alessandro Palumbo, Roberto Malagò, Ludovico La Grutta, W Bob Meiijboom, Annachiara Aldrovandi, Michele Fusaro, Luigi Vignali, Alberto Menozzi, et al. Influence of intra-coronary enhancement on diagnostic accuracy with 64-slice CT coronary angiography. *European radiology*, 18(3):576–583, 2008.

Ricardo C Cury, Suhny Abbara, Stephan Achenbach, Arthur Agatston, Daniel S Berman, Matthew J Budoff, Karin E Dill, Jill E Jacobs, Christopher D Maroules, Geoffrey D Rubin, et al. CAD-RADSTM coronary artery disease–reporting and data system. An expert consensus document of the Society of Cardiovascular Computed Tomography (SCCT), the American College of Radiology (ACR) and the North American Society for Cardiovascular Imaging (NASCI). Endorsed by the American College of Cardiology. *Journal of cardiovascular computed tomography*, 10(4):269–281, 2016.

Bob D De Vos, Floris F Berendsen, Max A Viergever, Hessam Sokooti, Marius Staring, and Ivana Išgum. A deep learning framework for unsupervised affine and deformable image registration. *Medical image analysis*, 52:128–143, 2019.

Damini Dey, Cynthia J Lee, Muneo Ohba, Ariel Gutstein, Piotr J Slomka, Victor Cheng, Yasuyuki Suzuki, Shoji Suzuki, Arik Wolak, Ludovic Le Meunier, et al. Image quality and artifacts in coronary CT angiography with dual-source CT: initial clinical experience. *Journal of cardiovascular computed tomography*, 2(2):105–114, 2008.

Xiaolu Fei, Xiangying Du, Qi Yang, Yun Shen, Pengyu Li, Jingmin Liao, and Kuncheng Li. 64-MDCT coronary angiography: phantom study of effects of vascular attenuation on detection of coronary stenosis. *American Journal of Roentgenology*, 191(1):43–49, 2008.

Dominik Fleischmann. CT angiography: injection and acquisition technique. *Radiologic Clinics*, 48(2):237–247, 2010.

Olivier Ghekiere, Rodrigo Salgado, Nico Buls, Tim Leiner, Isabelle Mancini, Piet Vanhoenacker, Paul Dendale, and Alain Nchimi. Image quality in coronary CT angiography: challenges and technical solutions. *The British journal of radiology*, 90(1072):20160567, 2017.

Lovedeep Gondara. Medical image denoising using convolutional denoising autoencoders. In *2016 IEEE 16th international conference on data mining workshops (ICDMW)*, pages 241–246. IEEE, 2016.

Ian Goodfellow, Jean Pouget-Abadie, Mehdi Mirza, Bing Xu, David Warde-Farley, Sherjil Ozair, Aaron Courville, and Yoshua Bengio. Generative adversarial networks. *Communications of the ACM*, 63(11):139–144, 2020.

Kaiming He, Xiangyu Zhang, Shaoqing Ren, and Jian Sun. Deep residual learning for image recognition. In *Proceedings of the IEEE conference on computer vision and pattern recognition*, pages 770–778, 2016.

Jonathan Ho, Ajay Jain, and Pieter Abbeel. Denoising diffusion probabilistic models. *Advances in Neural Information Processing Systems*, 33:6840–6851, 2020.

Udo Hoffmann, Quynh A Truong, David A Schoenfeld, Eric T Chou, Pamela K Woodard, John T Nagurney, J Hector Pope, Thomas H Hauser, Charles S White, Scott G Weiner, et al. Coronary CT angiography versus standard evaluation in acute chest pain. *New England Journal of Medicine*, 367(4):299–308, 2012.

Jun Horiguchi, Chikako Fujioka, Masao Kiguchi, Yun Shen, Christian E Althoff, Hideya Yamamoto, and Katsuhide Ito. Soft and intermediate plaques in coronary arteries: how accurately can we measure CT attenuation using 64-MDCT? *American Journal of Roentgenology*, 189(4):981, 2007.

Phillip Isola, Jun-Yan Zhu, Tinghui Zhou, and Alexei A Efros. Image-to-image translation with conditional adversarial networks. In *Proceedings of the IEEE conference on computer vision and pattern recognition*, pages 1125–1134, 2017.

Sunghee Jung, Soochahn Lee, Byunghwan Jeon, Yeonggul Jang, and Hyuk-Jae Chang. Deep learning cross-phase style transfer for motion artifact correction in coronary computed tomography angiography. *IEEE Access*, 8:81849–81863, 2020.

Eunhee Kang, Hyun Jung Koo, Dong Hyun Yang, Joon Bum Seo, and Jong Chul Ye. Cycle-consistent adversarial denoising network for multiphase coronary CT angiography. *Medical physics*, 46(2):550–562, 2019.

Naoto Kawaguchi, Akira Kurata, Teruhito Kido, Yoshiko Nishiyama, Tomoyuki Kido, Masao Miyagawa, Akiyoshi Ogimoto, and Teruhito Mochizuki. Optimization of coronary attenuation in coronary computed tomography angiography using diluted contrast material. *Circulation Journal*, 78(3):662–670, 2014.

Diederik P Kingma and Jimmy Ba. Adam: A method for stochastic optimization. *arXiv preprint arXiv:1412.6980*, 2014.

Martin Bødtker Mortensen, Omar Dzaye, Flemming Hald Steffensen, Hans Erik Bøtker, Jesper Møller Jensen, Niels Peter Rønnow Sand, Kristian Hay Kragholm, Henrik Toft Sørensen, Jonathon Leipsic, Michael Mæng, et al. Impact of plaque burden versus stenosis on ischemic events in patients with coronary atherosclerosis. *Journal of the American College of Cardiology*, 76(24):2803–2813, 2020.

Adam Paszke, Sam Gross, Francisco Massa, Adam Lerer, James Bradbury, Gregory Chanan, Trevor Killeen, Zeming Lin, Natalia Gimelshein, Luca Antiga, et al. Pytorch: An imperative style, high-performance deep learning library. *Advances in neural information processing systems*, 32, 2019.

Robin Rombach, Andreas Blattmann, Dominik Lorenz, Patrick Esser, and Björn Ommer. High-resolution image synthesis with latent diffusion models. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition*, pages 10684–10695, 2022.

Veit Sandfort, Younhee Choi, Rolf Symons, Marcus Y Chen, and David A Bluemke. An optimized test bolus contrast injection protocol for consistent coronary artery luminal enhancement for coronary ct angiography. *Academic Radiology*, 27(3):371–380, 2020.

Zhonghua Sun and Curtise KC Ng. Artificial Intelligence (Enhanced Super-Resolution Generative Adversarial Network) for Calcium Deblooming in Coronary Computed Tomography Angiography: A Feasibility Study. *Diagnostics*, 12(4):991, 2022.

Sanne GM Van Velzen, Bob D de Vos, Julia MH Noothout, Helena M Verkooijen, Max A Viergever, and Ivana Išgum. Generative models for reproducible coronary calcium scoring. *Journal of Medical Imaging*, 9(5):052406, 2022.

Jelmer M Wolterink, Anirban Mukhopadhyay, Tim Leiner, Thomas J Vogl, Andreas M Bucher, and Ivana Išgum. Generative adversarial networks: A primer for radiologists. *Radiographics*, 41(3):840–857, 2021.

Jun-Yan Zhu, Taesung Park, Phillip Isola, and Alexei A Efros. Unpaired image-to-image translation using cycle-consistent adversarial networks. In *Proceedings of the IEEE international conference on computer vision*, pages 2223–2232, 2017.

## 7. Appendix

Figure 5: Example outputs of the proposed method. Note how the highlighted regions in the offset mask strongly affect image contrast in the aortic root and coronary arteries, with blue causing an increase in attenuation and red causing a decrease.