On the benefits of deep RL in accelerated MRI sampling

Abstract

Deep learning approaches have shown great promise in accelerating magnetic resonance imaging (MRI) by reconstructing high quality images from highly under-sampled data. While previous sampling methods relied on heuristics, recent work has improved the state-of-the-art (SotA) with deep reinforcement learning (RL) sampling policies, which offer the possibility of long term planning and of adapting to the observations at test time. In this work, we perform a careful reproduction and comparison of SotA RL sampling methods. Surprisingly, we find that i) a greedily trained fixed policy can match or outperform deep RL methods and ii) find and resolve subtle variations in the preprocessing and reporting which previously made results incomparable across different works. Our results cast doubt on the added value of current RL approaches over fixed masks in MRI sampling and highlight the importance of leveraging strong fixed baselines, standardized reporting as well as isolating the source of improvement in a given work via ablations. We conclude with recommendations for the training and evaluation of deep reconstruction and sampling systems for adaptive MRI based on our findings.

1 Introduction

Magnetic resonance imaging (MRI) is a non-invasive, non-ionizing medical imaging method which has been widely adopted in clinical settings due to its with unmatched quality in soft tissue contrast. However, MRI suffers from long scanning times, which limits patient comfort, imaging quality as well as throughput (Zbontar et al., 2019). These issues have been a longstanding research direction in the community, and one prominent way to shorten scanning times is acquiring less measurements and use a reconstruction method to recover the full image, a setting known as *accelerated MRI*.

Spurred by the great gains in reconstruction quality (Muckley et al., 2020), recent works have moved away from using generic, non-learning based heuristics, such as variable-density sampling (VDS) (Lustig et al., 2007) to select the measurements to acquire. Instead, there has been a growing interest in strong, learning-based methods that tailor sampling policies to the reconstruction method and achieve even greater accelerations, and as a result, further speeding up imaging.

The state-of-the-art approaches rely mostly on the idea of pairing a reconstruction model with a patient-adaptive sampling model, where the former estimates a clean image from partial measurements, and the latter selects measurements that it predicts as likely to improve the reconstruction quality. The selection policy can be trained relying on techniques from reinforcement-learning (Jin et al., 2019; Pineda et al., 2020; Bakker et al., 2020), heuristics such as a model trying to estimate the Fourier space error of the locations to be acquired (Zhang et al., 2019) or end to end via the straight-through estimator (Van Gorp et al., 2021; Yin et al., 2021). Other approaches choose to directly parameterize a fixed mask and do not learn a policy via neural networks (Bahadir et al., 2019; Huijben et al., 2020b; Weiss et al., 2019).

In this work, we focus our investigation on the benefits of the policies trained using reinforcement learning, namely the contributions of Pineda et al. (2020) and Bakker et al. (2020). We reproduce their results and extend their evaluation with additional baselines and experimental settings, i.e. change of hyperparameters and preprocessing. Deep RL approaches have drawn attention because the policies can in principle offer two important benefits: i) long term planning and ii) adapting to the currently acquired image at test time. The work of Pineda et al. (2020) however seems to indicate that long term planning could be the most important component in deep RL, as their results show that a non-adaptive, long term planning policy model trained on the dataset can perform as well as

Sampling Method	Learning-based	Adaptive	Long horizon
VDS (Lustig et al., 2007)	✗	✗	✗
LBCS (Gözcü et al., 2018)	✓	✗	✗
Jin et al. (2019)	✓	✓	✓
Bakker et al. (2020)	✓	✓	✓/✗
Pineda et al. (2020)	✓	✓/✗	✓

Table 1: Methods that will be considered in the paper. Bakker et al. (2020) considered fixed vs. adaptive and greedy vs. non-greedy, Pineda et al. (2020) considered data specific vs. subject specific policies and also compared against greedy methods. VDS stands for Variable-Density Sampling (Lustig et al., 2007) a common heuristic which is not learning-based.

an adaptive, long-term planning policy. On the contrary, the contribution of Bakker et al. (2020) highlights the importance of adaptivity, as a greedy policy, that does not do long-term planning is found to closely match policies that do long term planning.

Our results synthesize these seemingly contradictory statements: *We observe that a simple, easy-to-train method that does not rely on deep RL, does not attempt long term planning and is by definition not adaptive can perform as well as the state-of-the-art approaches of Pineda et al. (2020); Bakker et al. (2020).* The fixed sampling policies are obtained by optimizing a fixed policy greedily optimization on a training set, following the Learning-based Compressed Sensing (LBCS) approach of (Gözcü et al., 2018; Sanchez et al., 2019).

This trend can be consistently observed on the fastMRI single-coil knee dataset (Zbontar et al., 2019), where we carry out evaluations across a variety of settings, whether full scale images, cropping the field of view, various mask designs used for training and various architectures. We observe that such small changes in the experimental pipeline can easily lead to *reversals*, where the RL approach is outperformed by LBCS. Such reversals can even occur when reporting the same experiment slightly differently, for instance, when reporting the performance at the last step of the trajectory, or summarizing the performance of the whole trajectory by its area under curve.

Together with the observation that current SotA RL methods only add marginal value over the simple baseline, we hope that this work will highlight the need for further discussions in the community about standardized metrics, strong baselines, and careful design of experimental pipelines to evaluate MRI sampling policies fairly.

To ensure reproducibility and facilitate the use of strong baselines in future work we will release our code upon acceptance of this paper.

2 BACKGROUND

This work focuses on the following inverse problem, where we seek to recover a signal $x \in \mathbb{C}^P$ from partial observations $y \in \mathbb{C}^N$, $N \ll P$ obtained by subsampling a unitary transform matrix $A \in \mathbb{C}^{P \times P}$:

$$y_\omega = P_\omega A x + \eta, \tag{1}$$

where $\eta \in \mathbb{C}^N$ is a noise vector, $\omega \subseteq [P] := \{1, \ldots, P\}$ is an index set of allowable sampling locations with cardinality N, P_ω is a diagonal matrix such that $(P_\omega)_{ii} = 1$ if $i \in \omega$, 0 otherwise. P_ω and ω are referred to as the *(sampling) mask*, as they control what locations are acquired in the original signal. This problem is inherently ill-posed, due to $N \ll P$, and our first goal will be to construct an estimate $\hat{x}_{\omega;\theta} = f_\theta(y_\omega, \omega)$ of the original signal x, where f_θ is a reconstruction method parameterized by θ.

2.1 MRI FUNDAMENTALS

In MRI, observations are obtained in the Fourier space, also referred to as *k-space*. The acquisition of data happens sequentially, but the physical constraints of an MRI acquisition make it more efficient to observe entire columns or rows at once, a setting known as *Cartesian sampling* (Zbontar et al., 2019). Non-Cartesian sampling is also possible (see for instance Weiss et al. (2019); Lazarus et al. (2019) for recent references), but some sampling methods like e.g. the pixel level sampling in Bahadir

Figure 1: Overview of the accelerated MRI pipeline. Acquisition happens sequentially in Fourier space, where a policy decides on the next location to acquire. Dashed lines indicate optional relations: not all policies rely on training data, and not all policies are data adaptive, as shown in Table 1.

et al. (2019); Yin et al. (2021) are not practical for 2D MRI, as they do not allow for a reduction of scanning time compared sampling full trajectories.

The acquisition of a column or row is known as a *readout*, and the complete procedure consists in acquiring N readouts sequentially. A full acquisition would require P readouts, and acquiring only N of P lines accelerates the scan by a factor P/N, adequately named *acceleration factor*, the inverse of which is known as the *sampling rate*, and the spectrum obtained with N out of P readouts is referred to as *undersampled*. The observations obtained by taking an inverse Fourier transform lead to an aliased image, and require processing through a reconstruction method.

In MRI, the Fourier space is typically structured, containing the bulk of energy in the low-frequencies located around the center of the space, and less around the high ones. While Compressed Sensing (CS) prescribes uniform sampling Donoho (2006); Candès et al. (2006), the structure of Fourier space made it necessary to leverage heuristics that assign more probability to low-frequencies in order to reflect the energy distribution. This approach is known as variable density-sampling (VDS) (Lustig et al., 2007).

2.2 SAMPLING OPTIMIZATION

The success of deep-learning approaches to MRI have led the broader medical imaging community to re-think the problem of optimizing the undersampling patterns in a data-driven fashion as well (Gözcü et al., 2018; Sanchez et al., 2019; Zhang et al., 2019; Jin et al., 2019; Pineda et al., 2020; Huijben et al., 2020a), instead of relying on heuristics such as variable density sampling (Lustig et al., 2007).

The ideal sampling algorithm would tailor the mask to each instance of $\boldsymbol{x} \sim p(\mathbf{x})$ solving

$$\min_{\omega:|\omega|\leq N} \ell(\boldsymbol{x}, \hat{\boldsymbol{x}}_\theta(\boldsymbol{y}_\omega = \mathbf{P}_\omega \boldsymbol{A}\boldsymbol{x})), \tag{2}$$

which is not realizable since this requires using the unknown ground truth signal $\boldsymbol{x}$ at testing time and is computationally intractable due to the combinatorial nature of the problem. Two main approaches have been explored to circumvent this problem.

Fixed (open-loop) sampling. A majority of data-driven mask design approaches use fixed masks (Ravishankar & Bresler, 2011; Gözcü et al., 2018; Sanchez et al., 2019; Bahadir et al., 2019; Wu et al., 2019; Huijben et al., 2020b). The subsampling mask is constructed ahead of time - either via a heuristic or by using training data - and kept fixed at inference time. Formally, the problem of choosing the subsampling pattern corresponds to finding a subset ω that satisfies

$$\arg\min_{\omega:|\omega|\leq N} \mathbb{E}_{\boldsymbol{x}\sim p(\mathbf{x})} \left[\ell(\boldsymbol{x}, \hat{\boldsymbol{x}}_\theta(\boldsymbol{y}_\omega = \boldsymbol{P}_\omega \boldsymbol{A}\boldsymbol{x}))\right], \tag{3}$$

where we are constrained with a maximal sampling budget N and want to find a mask that minimizes a given loss function ℓ. The true risk is substituted with the empirical one, estimates from training samples, and $\hat{\boldsymbol{x}}_\theta(\boldsymbol{y}_\omega = \boldsymbol{P}_\omega \boldsymbol{A}\boldsymbol{x})$ is an approximation of the ground truth obtained by reconstruction.

Adaptive (closed-loop) sampling. In contrast to Equation 3, adaptive sampling generates a dynamic sampling mask using a heuristic $\mathcal{H}_\phi$ evaluated at test time. For a fixed, unknown data sample $\boldsymbol{x}$,

we use information of the previously obtained measurements $\boldsymbol{y}_{\omega_{t-1}}$ to determine which candidate $v \subset [P]$, $v \notin \omega_{t-1}$ should be acquired at time t. These methods solve at test time, for each individual target $\boldsymbol{x}$, the following sequence of problems:

$$v_t \in \underset{v:v \in [P]}{\arg\min} \mathcal{H}_\phi(v, \hat{\boldsymbol{x}}_{\omega_{t-1};\theta}, \omega_{t-1}) \quad \text{for } t = 1, \ldots, N \tag{4}$$

where $\omega_t = \{\omega_{t-1}, v_t\}$ is defined in a nested fashion at each step. The heuristic score can be specifically trained for sampling leveraging reinforcement-learning framework (Bakker et al., 2020; Pineda et al., 2020; Jin et al., 2019), trained to directly estimate the current error and simply acquiring measurements where the error is estimated to be currently largest (Zhang et al., 2019), or be derived from available heuristics like posterior variance as in (Ji et al., 2008; Haupt et al., 2009; Sanchez et al., 2020).

3 THE MRI DATA PROCESSING PIPELINE

Section 2 outlined the mathematical setting underlying accelerated MRI, but does not capture practical considerations that can greatly affect the performance of methods as well as their applicability in the real world. In this section, we will step through the stages of the MRI data processing pipeline as shown in Figure 2. Note that this pipeline is independent of the sampling method and simply operates on a given mask.

Figure 2: Illustration of the the data processing pipeline of MRI subsampling. Diamonds represent data in Fourier domain (k-space) and circles represent data in image domain. The postprocessed observation and ground truth are the data that are subsequently used for training the reconstruction and policy models. The pipeline features three main blocks, namely *preprocessing*, *subsampling* and *postprocessing*.

As we will show in Section 4.1, seemingly trivial changes at even a single of these stages can affect the results of the evaluation and lead to changes in the ordering of the performance of different policies, an observation that we will refer to as *reversal*. We mark sections where we observed changes leading to reversals with * and sections here they only shifted the results but do not lead to reversals with *italics*. We also discuss some caveats at these different steps in Appendix B

Data sources and storage. In all cases, the ground truth signal $\boldsymbol{x}$ is initially acquired as a complex signal in k-space, generally using multiple coils. The fastMRI dataset (Zbontar et al., 2019) provides this raw multicoil data, as well as a simulated single-coil k-space which we use throughout this work.

Preprocessing*. Due to computational constraints, it is common to resize the images by cropping and/or resizing or use magnitude images over the raw data. Cropping and resizing changes the ground truth distribution, which as seen in Table 2 can also lead to reversals in the final results.

Sampling. While in the real world the data is actually acquired by sampling k-space (prospective sampling), in practice, the acquisition is generally simulated by retrospectively undersampling fully sampled Cartesian data following equation equation 1.

When training reconstructor and sampling method separately, training data are generally constructed with random masks that sample a certain fraction of center frequencies, and then the rest from a random distribution. There has been considerable variation in defining the parameters of these distributions, but to this day, no systematic study of their effect on the reconstruction quality have been carried out.

Post processing and reconstruction. After sampling, the observation will be processed according to the implementation details of the reconstruction algorithm (e.g., normalized or standardized) and the reconstruction is computed. We discuss the impact of preprocessing in Appendices B and C.2.

Evaluation metrics *. Finally to judge the results, the most common metrics are peak signal-to-noise ratio (PSNR), mean squared error (MSE), normalized mean squared error (NMSE) and structural similarity index metric (SSIM), reported as a curve plotted against sampling rate (Zhang et al., 2019), its reciprocal, the acceleration rate (Pineda et al., 2020), at the end of sampling or as an area under curve (AUC). As can be seen in Table 3, the choice of metric here can greatly impact results, especially when the examined methods are close in performance.

4   RE-EXAMINING DEEP RL FOR MRI SAMPLING

Our initial impetus is the observation that the non-adaptive oracle used in Bakker et al. (2020) is highly reminiscent of the LBCS method (Gözcü et al., 2018), which has been used as a strong baseline in the literature (Jin et al., 2019; Sanchez et al., 2020). To our surprise, evaluating this methods against the greedy RL method in the data processing pipeline used in Zhang et al. (2019) resulted in the fixed method not only coming close to the non-adaptive oracle, but actually closely matching the RL method (see Table 2). Wanting to perform a fair comparison as well as to asses the impact of the variability in the pipelines used across the literature, we closely replicated the pipeline of Bakker et al. (2020)[1] as well as an extensive ablation study to understand this reversal and identify its source.

In the sequel, the term *setting* then refers to a particular choice of preprocessing, subsampling and postprocessing of the data, independently of the data source (we always use the same dataset) and the model or sampling policy used. Preprocessing and postprocessing are relevant throughout the training of the reconstruction and the policy models, whereas changes in the sampling masks only directly affect the pretraining of the reconstructor, as the training the policy is done by successive rollouts using the policy model itself.

Dataset and preprocessing. Like the original paper of Bakker et al. (2020), we used the fastMRI (Zbontar et al., 2019) single-coil knee dataset for the experiments. We slightly modify follow their preprocessing (using complex data instead of magnitude data, different data normalization) that do not affect the relative ordering of the different methods. We provide ablations and a detailed discussion of these changes in Appendix C.2.

Specifically, they crop the data to 128×128, which we refer to as (c)). Zhang et al. (2019) instead resizes the data to 128×128 which we replicate by first cropping to 256×256 and then resizing them to 128×128. This preprocessing results in images with different fields of view. We refer to this preprocessing as (c+r) and integrate it to our ablations. We also evaluate horizontal (h) sampling masks, used in Gözcü et al. (2018); Jin et al. (2019) in addition to the vertical (v) sampling masks used by Bakker et al. (2020); Zhang et al. (2019) and Pineda et al. (2020).

The deep reconstructors used in Pineda et al. (2020); Bakker et al. (2020) pre-trained by randomly sampling a mask from a set of distributions with different parameters. We ablated over two pretraining regimes which we abbreviate as b and z, respectively. We describe them in more detail in Appendix A

Reconstruction models. We ablate over the two reconstruction models used in the RL SotA (Pineda et al., 2020) and (Bakker et al., 2020), namely the cResNet architecture from Zhang et al. (2019), used in Pineda et al. (2020) and the U-Net baseline provided in the fastMRI dataset, used in Bakker et al. (2020). Hyperparameters and training details are discussed in Appendix A.3.

Sampling methods In each ablation setting we compared Bakker et al. (2020) to the following baselines and oracles (citations in squared brackets here refer to prior works that also evaluate them):

- Random sampling (**Random**): Acquire a fixed proportion of low-frequency lines in Fourier and then randomly sample the remaining lines [Jin et al., 2019; Pineda et al., 2020; Bakker et al., 2020].
- Low-to-high frequencies (**LtH**): select k-space lines from low-to-high frequencies lines [Zhang et al., 2019; Pineda et al., 2020; Jin et al., 2019].
- (Stochastic) Learning-based Compressive Sampling (**LBCS**) (Gözcü et al., 2018; Sanchez et al., 2019): This method trains a non-adaptive, greedy sampling policy that selects as a measurement candidate in each acquisition step the column that leads to the greatest average improvement over a sample from the training dataset. We use the stochastic version that scales better to large dataset

---

[1]We thank the authors for providing the original checkpoints and general help and responsiveness during this replication

and images [Jin et al., 2019].

- Non-adaptive Oracle (**NA Oracle**) (Bakker et al., 2020): This oracle is computed by training and evaluating LBCS directly on the test set, and can serve to illustrate the benefit of adaptivity in greedy methods. This is the instance of a non-adaptive, greedy sampling method [Bakker et al., 2020].

For the training of the policy models of Bakker et al. (2020), we use the parameters of the greedy model in their paper, which we refer to simply as **RL** in the sequel. We excluded the non-greedy version, as Bakker et al. (2020) notice that the performance of the non-greedy model with discount factor $\gamma = 0.9$ is always close to within one standard deviation of the greedy model, but significantly more computationally demanding. We studied both the short and long horizon sampling regimes but only report on the long horizon for conciseness, with the short horizon results summarized in Appendix C.1.

Except for the deterministic LtH and NA Oracle,we report the performance of each method averaged on three runs/separately trained RL policies, along with the standard deviation.

Summarizing, we start in the setting of Bakker et al. (2020) except for using complex data and not magnitude data and then ablate over: **i)** cropped (c) vs cropped+resized (c+r), **ii)** vertical (v) vs horizontal (h) Cartesian sampling, **iii)** reconstruction with a pretrained UNet (Ronneberger et al., 2015; Bakker et al., 2020) or a cResNet (Zhang et al., 2019) and **iv)** the training regime for the reconstructor proposed in Bakker et al. (2020) (b) or in Zhang et al. (2019) (z). The base setting is referred to as cvb, for cropped (preprocessing) + vertical (mask) + Bakker et al. (sampling parameters) using a both a UNet and a cResNet reconstructor. Our main results will compare the difference between cvb and c+rhz to illustrate the impact of largely different data processing. A more complete ablation, using the setting cvz to ablate over the impact of different mask parameters, c+rvz to study the impact of different field of views and mask orientation is carried out in Appendix C.3.

4.1 RESULTS ON BAKKER ET AL. (2020)

A consistent trend that can be observed from Table 2 as well as the ablations in Appendix C.3 is that the return on investment (ROI) of adopting RL over LBCS is generally marginal compared to what changes in the modeling pipeline can bring. Improving the reconstruction architecture or using masking regimes adapted to the sampling horizon yield much more significant gains, e.g., moving from a UNet to a cResNet brings an order of magnitude more improvement (around 0.01 SSIM difference) than what RL brings over LBCS (around 0.0015 at best).

Additional observations can be made from Table 2. We see the performance of LBCS always remains close to the NA Oracle, testifying to the generalization ability of the fixed LBCS masks as indicated by theory. Results also show the superiority of cResNet over UNet across all settings, and show two occurrences of *reversals* when using a different data processing. Moving from cvb to c+rhz, we see that LBCS is outperformed by the RL in the former, and matches it in the latter, while the opposite happens for the cResNet. Note that the large gap in performance between the cvb and the c+rhz settings is not due to a significant difference in performance of the reconstruction methods, but rather originates from the difference in field of view. This can be seen by comparing the SSIM value of the observations (before reconstruction) using the deterministic LtH policy in both cases: in the cvb setting, one gets an AUC of 0.4989 for the observation SSIM, whereas in the c+rhz setting, the AUC is 0.6534.

Policy	cvb		c+rhz	
	UNet	cResNet	UNet	cResNet
Random	0.4348 ± 0.0001	0.4432 ± 0.0002	0.5860 ± 0.0002	0.5885 ± 0.0002
LtH	0.4849	0.5107	0.6678	0.6902
LBCS	0.5134 ± 0.0004	$\mathbf{0.5243 \pm 0.0003}$	$\mathbf{0.7035 \pm 0.0001}$	0.7096 ± 0.0001
RL	$\mathbf{0.5142 \pm 0.0001}$	0.5242 ± 0.0002	$\mathbf{0.7035 \pm 0.0002}$	$\mathbf{0.7111 \pm 0.0009}$
NA Oracle	0.5140	0.5247	0.7038	0.7099

Table 2: AUC on the test set using SSIM, on knee data comparing two models in the long horizon setting of Bakker et al. (2020).

4.2 Examining long range adaptivity with Pineda et al. (2020)

Given these results, we can observe that if there is indeed a benefit to adaptive but still greedy policies, it is highly sensitive to the parameters used, and the improvement over the fixed greedy baseline is marginal. One might wonder whether there will be a significant gain from adaptive RL policies if they are trained to perform long term planning on a longer horizon? To investigate this, we replicated the second SotA RL method described in Pineda et al. (2020), using the pretrained checkpoints due to computational constraints. This work uses a discount factor of $\gamma = 0.5$ and operates on the full fastMRI image of size 640×332, meaning it is both trained for a longer time horizon as well as having a lot more leeway for decision making. We replicated the *extreme* setting which starts with 2 center frequencies (0.6% undersampling rate or $166\times$ acceleration) and is evaluated up to 100 frequencies (30.1% undersampling rate or $3.32\times$ acceleration)[2]. As can be seen in Table 3, the LBCS mask is very close to the performance of both the data and sample specific DDQN policies provided by Pineda et al. (2020).

Interestingly, as seen in Figure 3 and Table 3, this is another example of a simple change in metric leading to a possible reversal of interpretation: visually in Figure 3a and when considering the AUC over sampling rate, LBCS seems competitive with the DDQN policies. In Figure 3b and the next column, we simply adopt the convention of Pineda et al. (2020) of using the acceleration rate (reciprocal of sampling rate) and we observe another reversal.

This puts more emphasis on the sub 5% range (i.e. below 17 lines) where LBCS has not yet caught up on DDQN (see Figure 7 in Appendix E for a more expensive large-scale version of LBCS trained on MSE which *does* catch up). Finally, if we report only the final sampling rate as is the convention in Bakker et al. (2020), the presentation gives the win to LBCS again as can be seen in the rightmost column. We also performed a more detailed comparison of the masks given by the policies, using a subset of the first 200 test set images, keeping the order fixed in across methods. The results can be found in Appendix D.1 but can be summarized as **i)** the adaptivity of SS-DDQN mainly affects the ordering of frequencies, a large section of acquired frequencies being shared across samples and in similar regions to the LBCS mask at the final sampling rate **ii)** it confirms that the adaptive masks have a small edge only until about a sampling rate of 5%, after which LBCS catches up and overtakes the RL policy.

Policy	SSIM			PSNR		
	Samp. rate	Acc. factor	Final rate	Samp. rate	Acc. factor	Final rate
Random	0.5801	0.4497	0.6723	26.489	22.327	28.962
LtH	0.5636	0.4506	0.6686	27.169	23.133	29.360
LBCS	0.6079	0.4787	**0.6886**	**28.491**	23.799	**30.211**
DS-DDQN	0.6101	**0.4797**	0.6855	28.240	**23.978**	29.652
SS-DDQN	**0.6139**	**0.4797**	0.6882	28.424	23.918	29.929
Adaptive Oracle	0.6341	0.4910	0.7131	29.013	24.498	30.683

Table 3: AUC on the test set when calculated against *sampling rate* and *acceleration* factor ($1/$sampling rate), as well as performance at the final sampling rate (100 lines acquired out of 332) on the knee dataset, using the processing of Pineda et al. (2020).

4.3 Discussion

As we have seen in Section 4 and Section 4.2, the fixed mask of LBCS is competitive with or outperforms both the greedy policy gradient methods of Bakker et al. (2020) and the non-greedy DDQN methods of Pineda et al. (2020). Our ablations in Section 4 and Appendices C.3 and E, as well as the example of Figure 3 highlight that the benefit of current RL methods over the fixed baseline, if it exists at all, is so small that a change of field of view, architecture or mask distribution (cf. Tables 2 and 5) can lead to a reversal in the relative performance ordering.

---

[2]We thank the authors for providing us with the original scores and general responsiveness and helpfulness during this replication.

(a) PSNR vs sampling rate   (b) PSNR vs acceleration rate

Figure 3: PSNR performance plot on the test set, using the knee dataset with the processing from Pineda et al. (2020). The plots feature two ways to report the same result, which are also displayed in Table 3. The SSIM performance plot can be found in Figure 6 in the Appendix.

Table 3 confirms the expectation that methods trained on SSIM can underperform with respect to PSNR and vice versa. In general agreement with the literature, we see that using several metrics is important to capture these tradeoffs, as reporting only one particular metric might hide underperformance in the other.

There have been various ways of reporting the metrics, ranging from showing sampling curves as in Figure 6a, reporting the area under curve (AUC) tables or simply reporting the metric at the final sampling rate. We see that reporting only the summarizing statistic can be deceitful and give relatively little information on the overall performance of the method. For instance, using only the last sampling rate, one could not distinguish between a method that performs well at all sampling rates leading to the final one, and one that has overall mediocre results and quickly improves on the performance at the end.

For this reason we consider sampling curves the gold standard for assessing the quality of accelerated MRI sampling methods, as a good performance at all sampling rates coupled with some stopping criterion could mean that MRI scans could be stopped when sufficient information has been acquired, further increasing their efficiency. If one wishes to report a single number, we recommend using AUC unless only the final performance is of interest.

There is also the question of reporting acceleration factors or sampling rates, which put focus on very different regimes of sampling: acceleration factors, as shown in (Pineda et al., 2020) focus on extreme undersampling rates. For instance, in the 2L scenario of Pineda et al. (2020), almost 90% of the plot and of the corresponding AUC consists of 20 lines out of the 100 acquired. While is true that the high acceleration regimes are where it is most desirable to make improvements, results on sampling rates have the advantage of uniformly distributing the performance throughout a range of interest instead of a single acquisition determining a third of the metric as in fig. 3b.

Alternatives to using acceleration factors could consist in displaying sampling rates over a region of interest or computing the AUC by explicitly assigning more weight to low sampling rates instead of doing it implicitly and nonlinearly through acceleration factors.

Finally, a close study of the masks given by the SS-DDQN in Figure 5, the edge of the SS-DDQN over the LBCS mask in Figure 4a and the comparison of the LBCS and SSDQN edge over a fixed heuristic in Figures 4b and 4c reveals that the frequencies selected by the adaptive policy are very similar to those selected by the LBCS mask and that the variability between samples is concentrated at very early sampling rates.

4.4 RELATION TO OTHER WORKS

While we focused strictly on RL trained sampling policies leveraging pre-trained reconstructors, there are other paradigms for accelerated MRI sampling present in the literature.

The first line is that of using end-to-end training of masks using stochastic relaxations of the sampling mask to be able to differentiate it and optimize the reconstructor and sampling method jointly. Bahadir

et al. (2019); Huijben et al. (2020) directly optimize for a single sampling rate and sampling mask. More recent work extend this technique for jointly training an adaptive policy, doing away with the pretraining and making more efficient use of the reconstructor's capacity (Yin et al., 2021; Van Gorp et al., 2021). Another line of joint training relies on self-supervised learning using Monte-Carlo Tree Search, as was done in Deepmind's AlphaGo Silver et al. (2017); Jin et al. (2019)

An open question regarding this line of research is whether the joint training enables to achieve a better sampling policy or simply serves as a curriculum and a way to specialize the sampling policy and reconstructor onto each other. To our knowledge, neither Van Gorp et al. (2021) nor Yin et al. (2021) investigated this. An interesting experiment could be to train a LBCS sampling mask on the co-trained reconstructor to see whether it can recover the sampling performance of the co-trained sampling network.

More similar to the spirit of our work, the recent study of Shimron et al. (2021) tackles subtle biases induced by the use of stored data and improper processing, Edupuganti et al. (2020) investigated specifically uncertainty methods for MRI reconstruction.

For works focusing on RL, Henderson et al. (2018) performed extensive ablation studies showing the sensitivity of RL methods to even minute variations in parameters. Engstrom et al. (2020); Gronauer et al. (2021) showed that the improvements of SotA RL methods can be traced to the exploitation of a subset of implementation details and algorithmic improvements and that even simple algorithms leveraging this subset can achieve SotA performance, similar to our work. This showcases that adapting to the data distribution seems to be sufficient to reach SotA in accelerated MRI sampling.

5 CONCLUSION

Taken together, our observations lead us to conclude that at least *in their current state*, RL SotA methods do not offer benefits over fixed masks in proportion to the complexity and training cost they introduce. Whether this limitation originates from a lack of structure in the k-space of MRI images or stems from algorithmic or computational considerations remains to be determined. It is also important to highlight that both RL methods do slightly outperform the fixed baseline in specific settings and sampling regions, meaning that they might be an avenue for maximizing performance out of a reconstructor trained on extreme undersampling rate.

We also want to emphasize that conducting the experiments in this paper would have been impossible if the RL methods had not been exemplary in terms of openness and reproducibility. Without access to the checkpoints and code and the authors responsiveness we would not have been able to reproduce both works and add the missing baseline, and despite the theoretical generalization guarantees of LBCS, we were surprised that it matched and sometimes simply outperformed more sophisticated methods. We therefore do not view our work as criticism of these works but rather as an extension and a synthesis, as stated in the introduction: Bakker et al. (2020) already speculated that non-greedy training might not be worth the increased training cost, while Pineda et al. (2020) already noted the long-range, non-adaptive DS-DDQN could be competitive with the sample specific variant. Our results confirm both of their observations that neither adaptivity nor long-term planning methods are *currently* worth the increased cost.

Finally, we hope that our ablations will help fellow researchers to navigate safely through the sea of confounders that arise in evaluating models for accelerated MRI sampling and avoid the common caveats and pitfalls that we highlight.

REFERENCES

Cagla Deniz Bahadir, Adrian V Dalca, and Mert R Sabuncu. Learning-based optimization of the under-sampling pattern in MRI. In *International Conference on Information Processing in Medical Imaging*, pp. 780–792. Springer, 2019.

Tim Bakker, Herke van Hoof, and Max Welling. Experimental design for mri by greedy policy search. *Advances in Neural Information Processing Systems*, 33, 2020.

Emmanuel J Candès, Justin Romberg, and Terence Tao. Robust uncertainty principles: Exact signal reconstruction from highly incomplete frequency information. *IEEE Trans, on Inf. Theory*, 52(2): 489–509, 2006.

David L Donoho. Compressed sensing. *IEEE transactions on Information Theory*, 52(4):1289–1306, 2006.

Vineet Edupuganti, Morteza Mardani, Shreyas Vasanawala, and John Pauly. Uncertainty quantification in deep MRI reconstruction. *IEEE Transactions on Medical Imaging*, pp. 1–1, 2020. ISSN 1558-254X. doi: 10.1109/TMI.2020.3025065. URL 10.1109/TMI.2020.3025065. Conference Name: IEEE Transactions on Medical Imaging 4 citations (Semantic Scholar/DOI) [2020-12-13].

Logan Engstrom, Andrew Ilyas, Shibani Santurkar, Dimitris Tsipras, Firdaus Janoos, Larry Rudolph, and Aleksander Madry. Implementation matters in deep rl: A case study on ppo and trpo. In *International Conference on Learning Representations*, 2020. URL https://openreview.net/forum?id=r1etN1rtPB.

Baran Gözcü, Rabeeh K. Mahabadi, Yen-Huan Li, Efe Ilıcak, Tolga Çukur, Jonathan Scarlett, and Volkan Cevher. Learning-based compressive MRI. *IEEE Transactions on Medical Imaging*, 2018.

Sven Gronauer, Martin Gottwald, and Klaus Diepold. The successful ingredients of policy gradient algorithms. In Zhi-Hua Zhou (ed.), *Proceedings of the Thirtieth International Joint Conference on Artificial Intelligence, IJCAI-21*, pp. 2455–2461. International Joint Conferences on Artificial Intelligence Organization, 8 2021. doi: 10.24963/ijcai.2021/338. URL https://doi.org/10.24963/ijcai.2021/338. Main Track.

Jarvis Haupt, Robert Nowak, and Rui Castro. Adaptive sensing for sparse signal recovery. In *2009 IEEE 13th Digital Signal Processing Workshop and 5th IEEE Signal Processing Education Workshop*, pp. 702–707. IEEE, 2009.

Peter Henderson, Riashat Islam, Philip Bachman, Joelle Pineau, Doina Precup, and David Meger. Deep reinforcement learning that matters. In *Proceedings of the AAAI conference on artificial intelligence*, volume 32, 2018.

I. A. M. Huijben, B. S. Veeling, and R. J. G. van Sloun. Learning sampling and model-based signal recovery for compressed sensing mri. In *2020 IEEE International Conference on Acoustics, Speech and Signal Processing (ICASSP)*, pp. 8906–8910, 2020.

Iris AM Huijben, Bastiaan S Veeling, Kees Janse, Massimo Mischi, and Ruud JG Van Sloun. Learning sub-sampling and signal recovery with applications in ultrasound imaging. *IEEE Transactions on Medical Imaging*, 2020a.

Iris A.M. Huijben, Bastiaan S. Veeling, and Ruud J.G. van Sloun. Deep probabilistic subsampling for task-adaptive compressed sensing. In *International Conference on Learning Representations*, 2020b. URL https://openreview.net/forum?id=SJeq9JBFvH.

Shihao Ji, Ya Xue, and Lawrence Carin. Bayesian compressive sensing. *IEEE Transactions on signal processing*, 56(6):2346–2356, 2008.

Kyong Hwan Jin, Michael Unser, and Kwang Moo Yi. Self-supervised deep active accelerated MRI. *arXiv preprint arXiv:1901.04547*, 2019.

Diederik P Kingma and Jimmy Ba. Adam: A method for stochastic optimization. *arXiv preprint arXiv:1412.6980*, 2014.

Carole Lazarus, Pierre Weiss, Nicolas Chauffert, Franck Mauconduit, Loubna El Gueddari, Christophe Destrieux, Ilyess Zemmoura, Alexandre Vignaud, and Philippe Ciuciu. Sparkling: variable-density k-space filling curves for accelerated t2*-weighted mri. *Magnetic resonance in medicine*, 81(6): 3643–3661, 2019.

Michael Lustig, David Donoho, and John M Pauly. Sparse MRI: The application of compressed sensing for rapid MR imaging. *Magnetic Resonance in Medicine*, 58(6):1182–1195, 2007.

Matthew J Muckley, Bruno Riemenschneider, Alireza Radmanesh, Sunwoo Kim, Geunu Jeong, Jingyu Ko, Yohan Jun, Hyungseob Shin, Dosik Hwang, Mahmoud Mostapha, et al. State-of-the-art machine learning mri reconstruction in 2020: Results of the second fastmri challenge. *arXiv preprint arXiv:2012.06318*, 2020.

Luis Pineda, Sumana Basu, Adriana Romero, Roberto Calandra, and Michal Drozdzal. Active mr k-space sampling with reinforcement learning. *arXiv preprint arXiv:2007.10469*, 2020.

Saiprasad Ravishankar and Yoram Bresler. Adaptive sampling design for compressed sensing MRI. In *Engineering in Medicine and Biology Society, EMBC, 2011 Annual International Conference of the IEEE*, pp. 3751–3755. IEEE, 2011.

Olaf Ronneberger, Philipp Fischer, and Thomas Brox. U-net: Convolutional networks for biomedical image segmentation. In *International Conference on Medical image computing and computer-assisted intervention*, pp. 234–241. Springer, 2015.

Thomas Sanchez, Baran Gözcü, Ruud B van Heeswijk, Efe Ilıcak, Tolga Çukur, et al. Scalable learning-based sampling optimization for compressive dynamic MRI. *arXiv preprint arXiv:1902.00386*, 2019.

Thomas Sanchez, Igor Krawczuk, Zhaodong Sun, and Volkan Cevher. Uncertainty-driven adaptive sampling via {gan}s. In *NeurIPS 2020 Workshop on Deep Learning and Inverse Problems*, 2020. URL https://openreview.net/forum?id=lWLYCQmtvW.

Efrat Shimron, Jonathan I Tamir, Ke Wang, and Michael Lustig. Subtle inverse crimes: Naively training machine learning algorithms could lead to overly-optimistic results. *arXiv preprint arXiv:2109.08237*, 2021.

David Silver, Julian Schrittwieser, Karen Simonyan, Ioannis Antonoglou, Aja Huang, Arthur Guez, Thomas Hubert, Lucas Baker, Matthew Lai, Adrian Bolton, et al. Mastering the game of go without human knowledge. *nature*, 550(7676):354–359, 2017.

Hans Van Gorp, Iris Huijben, Bastiaan S Veeling, Nicola Pezzotti, and Ruud JG Van Sloun. Active deep probabilistic subsampling. In *International Conference on Machine Learning*, pp. 10509–10518. PMLR, 2021.

Tomer Weiss, Sanketh Vedula, Ortal Senouf, Alex Bronstein, Oleg Michailovich, and Michael Zibulevsky. Learning fast magnetic resonance imaging. *arXiv preprint arXiv:1905.09324*, 2019.

Yan Wu, Mihaela Rosca, and Timothy Lillicrap. Deep compressed sensing. In Kamalika Chaudhuri and Ruslan Salakhutdinov (eds.), *Proceedings of the 36th International Conference on Machine Learning*, volume 97 of *Proceedings of Machine Learning Research*, pp. 6850–6860, Long Beach, California, USA, 2019. PMLR. URL http://proceedings.mlr.press/v97/wu19d.html.

Tianwei Yin, Zihui Wu, He Sun, Adrian V Dalca, Yisong Yue, and Katherine L Bouman. End-to-end sequential sampling and reconstruction for mr imaging. *arXiv preprint arXiv:2105.06460*, 2021.

Jure Zbontar, Florian Knoll, Anuroop Sriram, Tullie Murrell, Zhengnan Huang, Matthew J. Muckley, Aaron Defazio, Ruben Stern, Patricia Johnson, Mary Bruno, Marc Parente, Krzysztof J. Geras, Joe Katsnelson, Hersh Chandarana, Zizhao Zhang, Michal Drozdzal, Adriana Romero, Michael Rabbat, Pascal Vincent, Nafissa Yakubova, James Pinkerton, Duo Wang, Erich Owens, C. Lawrence Zitnick, Michael P. Recht, Daniel K. Sodickson, and Yvonne W. Lui. fastMRI: An Open Dataset and Benchmarks for Accelerated MRI, 2019. URL http://arxiv.org/abs/1811.08839.

Hongyi Zhang, Yann N. Dauphin, and Tengyu Ma. Fixup Initialization: Residual Learning Without Normalization. *ICLR*, 2018. URL https://openreview.net/forum?id=H1gsz30cKX.

Zizhao Zhang, Adriana Romero, Matthew J Muckley, Pascal Vincent, Lin Yang, and Michal Drozdzal. Reducing uncertainty in undersampled MRI reconstruction with active acquisition. In *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition*, pp. 2049–2058, 2019.

A   IMPLEMENTATION DETAILS

A.1   PRETRAINING REGIMES

The deep reconstructors used in Pineda et al. (2020); Bakker et al. (2020) are pre-trained by randomly sampling a mask from a set of distributions with different parameters. Bakker et al. (2020) uses a discrete set of sampling rates at (25%,25%,25%, 16.7%,16.7%, and 12.5%) with (25%,16.7%,12.5%, 16.7%,12.5%,12.5%) of the total selected frequencies being allocated to center frequencies and the rest sampled uniformly,while Zhang et al. (2019) uses a distribution where 10 center frequencies are always acquired, and between 0 and 38 additional frequency lines are acquired following a uniform distribution (as a reminder, the total number of possible sampling locations is 128 lines or columns).

A.2   COMPUTATIONAL HARDWARE

We performed all of your experiments on a DGX-2 server using $A100$ GPUs. On this machine the reconstrction model of Pineda et al. (2020) fits into GPU memory with an effective batchsize of 50 (we use subbatching to enable arbitrary batch sizes) and it took ≈ 3 hours to train each of the reported LBCS masks. Meanwhile masks for the comparisons of Bakker et al. (2020) could be trained in ≈ 20 minutes, while training the RL policy took on the order of days as reported in the authors original paper. While one might be able to obtain a certain speedup using a more optimised or parallel RL algorithm, a key bottleneck is the sequential nature of the optimization.

A.3   RECONSTRUCTOR ARCHITECTURES

We experiment on two main generator architectures, a UNet(Ronneberger et al., 2015) as used in Zbontar et al. (2019) and Bakker et al. (2020) and the cascade of residual networks, called cResNet introduced by Zhang et al. (2019).

For the cResNet in the comparison with Pineda et al. (2020) we used the checkpoint provided at `https://facebookresearch.github.io/active-mri-acquisition/misc.html` by the authors. All UNet models of the ablations of the Bakker et al. (2020) setting were trained following the hyperparameters described in the paper, i.e. using Adam Kingma & Ba (2014) and training to 50 epochs, otherwise using the hyperparameters from Zbontar et al. (2019) (except for the $10\times$ learning rate drop 40 epochs, we instead kept the initial 10^{-3} rate throughout). We used a batchs ize of 32.

For the cResNet models used in the ablations we used Adam as an optimizer with learning rate 10^{-3},betas of $(0.9, 999)$, trained for 50 epochs and a batch size of 32. The architecture is the same as in Zhang et al. (2019) except only outputting the mean of the image (no uncertainty channel), and using 72 channels for the 3 residual blocks, with $18, 36, 72$ channels in the encoder and $72, 36, 18$ channels in the decoder.

The images for both were undersampled using masks according to the settings described in appendix A.1. For the UNet,like Bakker et al. (2020) we take a single channel for magnitude only data as input, while for the C-ResNet we take the 2 complex channels as inputs. All models are trained using ℓ_1 loss and directly output the reconstruction without a final nonlinearity. The UNet has $837'635$ parameters in total, while the c-ResNet is larger with $1'093'479$ parameters.

B   EXTENDED DISCUSSION ON THE MRI DATA PROCESSING PIPELINE

We extend the discussion of Section 4, focusing on some caveats and common pitfalls to avoid.

Caveat #1 (Preprocessing). Converting the raw ground truth to or starting from a magnitude induces a conjugate symmetry in Fourier space that is not present in real data, a distortion that makes our modeling less faithful to the physical model.

Postprocessing details. Typical normalization, as used for instance in Pineda et al. (2020), consist of dividing every input by a fixed value equal to the average energy of the dataset. Bakker et al. (2020) standardized their data and then clamped them to have standardized data in the range $[-6, 6]$.

We found that standardization used by Bakker et al. (2020) was necessary only when paired with non-residual models, which tracks with observations in the literature that residual networks can work without normalization, although usually this is studied for network internal normalization which we also leverage (see e.g. (Zhang et al., 2018) which removes normalization entirely by using a proper initialization).

Caveat #2 (Postprocessing). Normalization should occur in a consistent fashion between ground truth and observation data, as failing to do this can lead to inconsistent values in the reconstruction, especially in parts where data consistency occurs. In (Bakker et al., 2020), the authors normalize observations and ground truth using their respective statistics and denormalize the reconstructed image using the *ground truth* statistics, which is not compatible with the use of data consistency in the reconstructor. If one wishes to get a realistic estimate of the performance at deployment, it's also advisable to use only statistics available at test time (i.e. observation statistics, not ground truth statistics).

C   FURTHER EXPERIMENTS ON BAKKER'S SETTING

C.1   DATA PROCESSING DIFFERENCES

In section 4, we mentioned that we slightly deviated from the setting of Bakker et al. (2020). We now detail these changes and show that they do not change the *relative* performance of the different methods.

1. **Train-test split:** we used a different train-test split than Bakker et al. (2020), as we randomly splitted 10% of the training set as a test set, and used the fastMRI validation set for test. We used the 50% more central slices, resulting in 15599 training slices, 1743 validation slices and 3564 test slices.

2. **Complex data undersampling:** , Bakker et al. (2020) use magnitude ground truth images as the reference that is undersampled. We discussed in Appendix B the issue with this approach. We chose to use complex preprocessing of data, followed by taking magnitude of the observation obtained *after* undersampling.

3. **Data range:** PSNR and SSIM need to be provided with a maximal data range in their computation: Bakker et al. (2020) used the maximal intensity in the ground truth *volume*, while we used the maximal intensity of each ground truth *slice* or image.

4. **Data standardization:** Bakker et al. (2020) used a unique data standardization, where observations and ground truth data are standardized using their respective statistics, but denormalized, after reconstruction, using only the ground truth statistics. While this ensures a more closely matching data range, this introduces a mismatching data range that biases models leveraging data consistency, used in most state-of-the-art models. We performed *matched* data normalization and denormalization using the observation statistics for the observation, and the ground truth ones for the ground truth.

The third and fourth changes are related to postprocessing, and do *not* require retraining a model. We see on the three first columns of Table 4 that these changes do not alter the relative ordering of the methods. However, in absolute numbers, the worse performing method in the initial setting outperforms the best method the best performing one after these changes (highlighted in italic on Table 4). While comparing these numbers would be a mistake, this highlights the impact of subtle postprocessing changes and the need for care in comparing methods, especially across different papers.

C.2   ABLATION OVER MAGNITUDE VS COMPLEX DATA

We first illustrate the impact of doing the appropriate preprocessing by evaluating the model of Bakker et al. (2020) using their data processing pipeline against a pipeline with complex preprocessing. The SSIM performance AUC is reported in Table 4.

This would correspond to a simulation of what would happen if the model of Bakker et al. (2020) was to be used in deployment, as one would receive undersampled complex observations that are

Policy	Bakker's setting	Individual data range	Matched standardization	Mismatched setting	Complex preprocessing (matched - ours)
Random (SH)	*0.6251*	0.5964	0.5798	0.5595	0.5608
LtH (SH)	0.6073	0.5762	0.5650	0.5617	0.5719
LBCS (SH)	0.6375	0.6099	0.5928	0.5684	0.5732
RL (SH)	**0.6388**	**0.6112**	*0.5941*	**0.5698**	**0.5738**
NA Oracle (SH)	0.6383	0.6107	0.5937	0.5717	0.5738

Table 4: AUC on the test set, using SSIM, on various data processing on Bakker's knee setting. The ablation is carried out over The mismatched setting refers to the evaluation of the model trained in the setting of Bakker et al. (2020), but evaluated complex data preprocessing instead of magnitude data preprocessing. This makes the reconstruction and policy models to be out of the distribution. A model trained on the complex preprocessing is reported in the complex preprocessing (matched - ours) column.

only then transformed to magnitude images. While this does not induce any *reversal* in this case, we see that in the matched setting, the gap between LBCS and RL shrinks. We see also again that training and evaluating on the matched setting leads to larger performance improvements than the ones obtained by training sophisticated policies.

C.3 FURTHER RESULTS IN THE BAKKER EXPERIMENT

We provide the results of our full ablation study on the components described in section 4. We report observations on the `cvb`, `cvz`, `c+rvz` and `c+rhz` settings. The summary of the full trajectory with an AUC is presented on Table 5 and the performance at the end of sampling is shown in Table 6.

It is difficult to consistently establish a trend for when reversarls will happen, but several important observations can be made from the results.

First of all, by comparing Tables 5 and Table 6, we see that reversals can happen by considering one way of reporting over another (see `c+rvz` cResNet in both tables). This highlights again the impact of the way results are reported.

The effect of the masks used for pretraining the reconstructor is also interesting. Comparing the `cvb` and `cvz` results, once can consistently see the following trend. In the short horizon setting (LHS of the tables), using `cvb` leads to consistent improvements over `cvz`. In the long horizon setting, the opposite is true. Recall that the `b` masks are discretely distributed from sampling rates 12.5% to 25%, which matches the short horizon experiment range. The `z` masks span a continuous range, from roughly 7% to 37.5%, whereas the long horizon experiment spans a range from 3% to 25% sampling rate. These results suggest that matching the pretraining regime to the regime on which evaluation will be carried out has a significant influence on the performance of the sampling policy, an observation which has, to our knowledge, never been quantified before.

D DIGGING INTO LONG RANGE ADAPTIVITY

Since we had to update the codebase to the current Pytorch FFT API we used the Low-To-High baseline as confirmation that we indeed replicate their setting exactly and then compare their method against the Low-To-High and random sampling baselines. We also use the fact that the code provided by the authors reports the images used for evaluation to evaluate LBCS on *exactly* the same subset of the validation set as the original codebase (we always train on the training set only). Similar Pineda et al. (2020), we train a separate LBCS mask for the reporting of SSIM and PSNR/MSE/NMSE (we found the latter 3 to be highly correlated in performance).

D.1 ADAPTIVITY

As can be seen in fig. 5a., d. and e, LBCS puts more emphasis on the center frequencies, but acquires similar sections of kspace as the SSDDQN. It also creates a more symmetric masks, which is in line with Pineda et al. (2020) observations that SSIM creates more asymmetric masks. More interestingly, as can be seen in fig. 5c., most of the variation is concentrated at the early sampling rates (left of

Policy	Short Horizon				Long Horizon			
	cvb		c+rhz		cvb		c+rhz	
	UNet	CResNet	UNet	CResNet	UNet	CResNet	UNet	CResNet
Random	0.5608	0.5659	0.7292	0.7338	0.4348	0.4432	0.5860	0.5885
LtH	0.5719	0.5764	0.7309	0.7412	0.4849	0.5107	0.6678	0.6902
LBCS	0.5732	0.5828	**0.7430**	**0.7526**	0.5134	**0.5243**	**0.7035**	0.7096
RL	**0.5738**	**0.5830**	**0.7430**	0.7524	**0.5142**	0.5242	**0.7035**	**0.7111**
NA Oracle	0.5738	0.5832	0.7430	0.7527	0.5140	0.5247	0.7038	0.7099

Policy	Short Horizon				Long Horizon			
	cvz		c+rvz		cvz		c+rvz	
	UNet	CResNet	UNet	CResNet	UNet	CResNet	UNet	CResNet
Random	0.5580	0.5661	0.6851	0.6984	0.4483	0.4621	0.5252	0.5276
LtH	0.5663	0.5759	0.6739	0.7028	0.5122	0.5218	0.6292	0.6444
LBCS	0.5712	0.5822	0.7034	0.7235	0.5174	0.5328	0.6568	0.6712
RL	**0.5717**	**0.5829**	**0.7042**	**0.7239**	**0.5183**	**0.5334**	**0.6582**	**0.6733**
NA Oracle	0.5718	0.5835	0.7038	0.7236	0.5180	0.5336	0.6568	0.6717

Table 5: AUC on the test set, using SSIM, for the full ablation study using the model of Bakker et al. (2020). The short and long horizon results are *not* comparable with each other, as AUCs are integrated on the whole range of sampling rates. The top right part of the table (long horizon) replicated the results of table 2, excluding the standard deviation for legibility. The rest of the ablation were *not* averaged on several seeds due for computational reasons.

Policy	Short Horizon				Long Horizon			
	cvb		c+rhz		cvb		c+rhz	
	UNet	CResNet	UNet	CResNet	UNet	CResNet	UNet	CResNet
Random	0.607	0.6141	0.7565	0.7667	0.5376	0.5488	0.6652	0.6745
LtH	0.6267	0.6313	0.7602	0.7786	0.6267	0.6313	0.7602	0.7786
LBCS	0.6288	0.6413	**0.7751**	**0.7923**	0.6275	0.6411	**0.7771**	**0.7941**
RL	**0.6298**	**0.6417**	0.7745	**0.7921**	**0.6302**	**0.6421**	0.7758	0.7924
NA Oracle	0.6301	0.6421	0.7751	0.7926	0.6301	0.6428	0.7771	0.7942

Policy	Short Horizon				Long Horizon			
	cvz		c+rvz		cvz		c+rvz	
	UNet	CResNet	UNet	CResNet	UNet	CResNet	UNet	CResNet
Random	0.6053	0.6148	0.7201	0.7401	0.5502	0.5693	0.6191	0.6359
LtH	0.619	0.6309	0.7052	0.7487	0.6190	0.6309	0.7052	0.7487
LBCS	0.6282	0.6413	0.7411	0.7710	0.6293	**0.6451**	0.7453	**0.7746**
RL	**0.6288**	**0.6416**	**0.7417**	**0.7714**	**0.6304**	0.6452	**0.7470**	0.7738
NA Oracle	0.6292	0.643	0.741	0.7717	0.6304	0.6464	0.7457	0.7755

Table 6: SSIM at 25% sampling rate, using SSIM, with the model of Bakker et al. (2020). This is the counterpart of the results shown in Table 5. Here, the results across short and long horizon are comparable.

the plot) with the std and especially the coefficient of variation ($\frac{\sigma}{\mu}$) decaying towards zero in most locations. This implies that while SSDDQN is indeed adapting to each image individually, this mainly affects ordering early on and after 20 -40 samples ($6 - 12\%$ sampling rate) LBCS starts to catch up. This is also supported by two observations:

(a) LBCS-SSDDQN (b) LBCS-LtH (c) SSDDQN-LtH

Figure 4: Per Image distribution of SSIM differences between sampling methods across the sampling process. Each shade corresponds to a 10 percent region, with the lightest shade indicating max and min regions.

Figure 5: SS-DDQN variability and comparison with the LBCS mask. CoV stands for Coefficient of Variation here.

1. LBCS underperforms SSDQN by a larger margin than on the full evaluation, which we interpret as the mask being good *on average* and with a smaller sample there is a higher chance of individual suboptimality

2. the per-image difference between LBCS and SSDQN grows in favor of SSDDQN until about 40 sampels where LBCS starts to slowly recover. The distribution of the difference however becomes much wider, implying there are images where LBCS performs wildly different from SSDDQN

(a) SSIM vs sampling rate (b) SSIM vs acceleration rate

Figure 6: Two ways to report the same result, SSIM version of fig. 3

E   DETAILS ON LBCS USED FOR THE FIXED MASK DESIGN

Algorithm 1 shows the pseudocode of the algorithm we used for designing the fixed masks used for the comparisons throughout the paper, while table 7 summarizes the hyper parameters used and table 8 gives an overview of the computational complexity. Note that LBCS is almost completely parallelizable, which leads to the stark runtime differences noted in appendix A. As can be seen when comparing fig. 3 and fig. 7, once a certain data batch size l and sampling candidate set cardinality k is reached, LBCS performance saturates although it continues to benefit especially in the low sampling regime. This means that one could very feasibly reduce l for the comparison with Bakker et al. (2020), we simply went with a larger batchsize because we could.

Setting	Num. lines $\|\mathcal{S}\|$	Max. cardinality N	Candidate set size k	Data batch size l
Short horizon (Bakker et al., 2020)	128	16	128	256
Long horizon (Bakker et al., 2020)	128	28	128	256
(Pineda et al., 2020)	332	100	64	20
(Pineda et al., 2020) "big"	332	48	200	256

Table 7: The hyperparameters used for the stochastic LBCS masks throughout comparisons

---

Algorithm 1 Stochastic LBCS mask design
(G) refers to the greedy algorithm Gözcü et al. (2018)
(SG) refers to the stochastic greedy algorithm of Sanchez et al. (2019)

Input: Training data $\{\boldsymbol{x}\}_{i=1}^m$, reconstructor f_θ, sampling set $\mathcal{S}$, max. cardinality N, samp. batch size k, train. batch size l, performance metric $\eta(\cdot,\cdot)$
Output: Sampling pattern ω

1:
2: **while** $|\omega| \leq N$ **do**
3:     **(G)** $\begin{cases} \text{Pick } \mathcal{S}_{iter} = \mathcal{S} \\ \text{Pick } \mathcal{L} = \{1,\ldots,m\} \end{cases}$
4:     **(SG)** $\begin{cases} \text{Pick } \mathcal{S}_{iter} \subseteq \mathcal{S} \text{ at random, with } |\mathcal{S}_{iter}| = k \\ \text{Pick } \mathcal{L} \subseteq \{1,\ldots,m\}, \text{ with } |\mathcal{L}| = l \end{cases}$
5:     **for** $S \in \mathcal{S}_{iter}$ such that $|\omega \cup S| \leq \Gamma$ **do**
6:         $\omega' = \omega \cup S$
7:         For each $\ell \in \mathcal{L}$ set $\hat{x}_\ell \leftarrow f_\theta(\boldsymbol{P}_{\omega'} \boldsymbol{A} \boldsymbol{x}_\ell, \omega')$
8:         $\eta(\omega') \leftarrow \frac{1}{|\mathcal{L}|} \sum_{\ell \in \mathcal{L}} \eta(x_\ell, \hat{x}_\ell)$
9:     $\omega \leftarrow \omega \cup S^*$, where $S^* = \underset{S:|\omega \cup S| \leq N}{\arg\max} \ \eta(\omega \cup S)$
10: **return** ω

Method	Forward	Backward	Total
(Bakker et al., 2020)	$q(8)E(50)B(16)H(n_r + n_p) = 6'400H(n_r + n_p)$	$6400Hn_p$	$6'400H(n_r + 2n_p)$
LBCS for Bakker et al. (2020)	$n(H)k(128)l(256)n_r = 32'768Hn_r$	0	$32'768Hn_r$
(Pineda et al., 2020)	$5e6(n_r + n_p)$	$5e6(n_p)$	$5e6(n_r + 2n_p)$
LBCS for Pineda et al. (2020)	$n(100)k(64)l(20)n_r = 128e3n_r$	0	$128e3n_r$

Table 8: Approximate computational cost of the compared methods. Note that at test time, LBCS is basically free while the RL policies will still need to be deployed. n_r, n_p are the parameter counts of the reconstruction and sampling policies respectively

(a) PSNR vs sampling rate   (b) SSIM vs sampling rate

Figure 7: PSNR and SSIM of LBCS with larger batch size (LBCS - big) trained on MSE, zoomed to the region where the smaller batch size LBCS underperformed. Note that it fully matches or outperforms both versions of DDQN on PSNR.