A deep learning method trained on synthetic data for digital breast tomosynthesis reconstruction

Abstract

Digital Breast Tomosynthesis (DBT) is an X-ray imaging modality enabling the reconstruction of 3D volumes of breasts. DBT is mainly used for cancer screening, and is intended to replace conventional mammography in the coming years. However, DBT reconstructions are impeded by several types of artefacts induced by the geometry of the device itself, degrading the image quality and limiting its resolution along the thickness of the compressed breast. In this study, we propose one of the first deep-learning-based pipelines to address the highly challenging DBT reconstruction problem, focusing on the removal of sparse-view and limited-angle artefacts. Specifically, this procedure is composed of two steps: a classic reconstruction algorithm is first applied on normalised projections, then a deep neural network is tasked with erasing the artefacts present in the obtained reconstructed volumes. A major difficulty to address this problem is the lack of real conditions artefact-free data. To overcome this complication, we resort to a new dataset comprised of synthetic breast texture phantoms. We then show that our training method and database strategy are promising to tackle the problem as they improve the informational value of planes orthogonal to the detector, whose imaging quality is currently too low to be used by radiologists. Eventually, we assess the impact of removing the bias components from the network and using stacks of slices as inputs, with regard to the generalisation ability of our approach on both synthetic and clinical data.

Keywords: DBT reconstruction, inverse problem, deep learning, limited angle, sparse view, synthetic phantoms, 2.5D, bias-free network.

1. Introduction

Digital breast tomosynthesis (DBT) is an X-ray imaging modality used for breast cancer screening (Niklason et al., 1997). It relies on several low-dose cone-beam acquisitions that are performed at different angles. The resulting projections can then, in principle, be used to reconstruct a 3D volume modelling the breast. DBT devices are however impeded by several physical constraints, such as the X-ray source rotation being limited to a certain angular range (limited angle), and the number of projections not exceeding a dozen (sparse view). Consequently, the 3D volume reconstruction quality is generally dramatically reduced.

Several reconstruction algorithms have been designed to recover volumes from DBT projections (Sechopoulos, 2013). They can be divided into two main families, the analytical (*e.g.*, Filtered Back-Projection – FBP) and iterative methods (*e.g.*, Simultaneous Iterative Reconstruction – SIRT). Nevertheless, when applied to projections which are generated by a system suffering from the aforementioned physical limitations, these techniques do not allow for high precision reconstructions (Vedantham et al., 2015). The resolution on the source-to-detector axis (vertical direction) is severely limited, and a number of artefacts

Figure 1: Examples of phantoms. (a) and (b) are coronal slices of a geometric phantom and its FBP reconstruction: as can be seen, the objects are difficult to distinguish in the reconstruction due to limited-angle and sparse-view artefacts along the vertical direction. (c) is an axial slice of a breast texture phantom used in this work and (d) a projection of the whole phantom.

emerge due to the geometrical limitations (see Figure 1 (a) and (b)): objects are propagated through planes and tend to disappear too slowly, restricting the ability to determine their true boundaries. This spreading effect results in breast tissues being poorly separated along the vertical direction. As a consequence, only the planes parallel to the detector are interpreted by radiologists during an examination.

In this work, we tackle limited-angle sparse-view DBT reconstruction. Specifically, we propose a deep learning reconstruction method for a realistic DBT device with 9 projections acquired over a 25° angular range. To our knowledge, no author has yet proposed a pipeline of this kind to address a geometrical setting as constrained as the one studied here. The method comprehends two steps: specific data is first processed by a classic analytical reconstruction method, namely FBP, enabling to obtain an initial, albeit imperfect reconstruction of the 3D volume. A deep neural network is then applied in order to correct the artefacts mentioned above. Due to the lack of datasets with ground truth volumes, the network is trained on meticulously crafted reconstructions of synthetic data. Such a 2-step approach is empirically shown to greatly improve the image quality of DBT reconstructions compared to FBP. In addition, we further analyse the interest of giving more spatial context as input to the model (2.5D) and investigate the use of bias-free neural networks.

2. Related works

Over the last years, several works have been using deep learning to tackle X-ray imaging reconstruction (Wang et al., 2020). Most of these are however dedicated to computed tomography (CT), and few of them address both sparse-view and limited-angle artefacts. Two main workflows exist to address X-ray reconstruction, depending on how deep-learning is used inside the reconstruction pipeline (Liang et al., 2019; Wang et al., 2020).

First, some authors rely on a single neural network to perform the whole reconstruction from the projections. Nevertheless, merely using generic neural network architectures requires many high-quality training datasets to obtain good results, and, as pointed out before, no DBT data with the corresponding ground-truth is available. As such, some works rely on designing network architectures better taking into account the nature of the

problem at hand by *unrolling* iterative reconstruction methods. Deep learning algorithms for DBT reconstruction mainly use this strategy (Wu et al., 2020; Teuwen et al., 2021; Su et al., 2021). Toward the same objective, others use *plug-and-play* approaches, replacing some mathematical operators of an iterative method by a neural network (Liu et al., 2022; Hu et al., 2022). Additionally, hybrid reconstruction pipelines combine two neural networks operating respectively in the projection and volume domains (Liang et al., 2019; Zhou et al., 2022). Although having shown some good results for several inverse problems, both *unrolled* and *plug-and-play* methods yield network architectures that might not be flexible enough to benefit from the whole training set information, as their performance is inherently restricted to specific prior forms (Zhang et al., 2017).

Secondly, some works propose to combine neural networks with classic reconstruction algorithms. In pre-processing pipelines, a neural network (Liang et al., 2019) works in the projection domain to simulate the missing projections, which are then fed into a classic reconstruction algorithm. In post-processing pipelines (Jin et al., 2017; Qiao and Du, 2022), a first imperfect reconstruction is obtained with a conventional algorithm before being processed by a neural network to remove the artefacts.

3. Methodology

3.1. Mathematical problem statement

Let $X \in \mathbb{R}^{i \times j \times k}$ be a vector representing a 3D object to image with size (i, j, k), and $Y \in \mathbb{R}^{w \times h \times m}$ the vector of the corresponding m DBT projections on a detector of width w and height h. The X-ray acquisition can be modelled as $Y = \mathcal{P}X + \eta$, where $\mathcal{P}$ is the projection operator, and η is the realisation of a random variable modelling the noise. DBT reconstruction consists in retrieving X from the noisy measurements Y. However, because of the physical constraints explained in the introduction, measurements Y do not carry enough information to compute a perfect estimation of X. Said differently, DBT is an ill-posed inverse problem.

3.2. Algorithm

To solve our DBT reconstruction problem, we use a modification of the model developed by Jin et al. (2017) to address CT reconstruction. This architecture is often used as baseline for posterior works, and we assess how it performs on our DBT dataset. As pointed out in Section 2, the authors of this article propose a post-processing deep learning pipeline on top of a conventional FBP reconstruction step. We use the same approach here and detail in the following of this subsection the deep learning part.

The model is based upon U-Net (Ronneberger et al., 2015). In the decoder part of the network, we make sure that deconvolution operations consist of an upsampling followed by a convolution to avoid chequerboard artefacts (Odena et al., 2016). The depth of the network is set to 5 levels, with feature maps size increasing from 64 to 1024, as in the original paper. Furthermore, in contrast to Jin et al. (2017), we study in Section 4 the relevance of removing bias terms from the convolutions and the batch normalisation steps of the network. This adjustment makes the network homogeneous (*i.e.*, $f(\alpha x) = \alpha f(x)$), which has been shown to be more efficient on some inverse problems such as denoising (Mohan et al., 2020).

In addition, processing a 3D data volume with the above 2D network requires slicing it as explained in Section 3.3. Such an approach can however lead to deteriorated results, as all the spatial information present in the full 3D volume is not leveraged. Therefore, we propose here to resort to a so called 2.5D approach. 2.5D consists in stacking neighbour slices of the 3D data volume in a sliding window with the aim of inferring the central image. Turning to such inputs can drastically improve the visual aspect of output images, sometimes as much as using heavier 3D convolutions (Ziabari et al., 2018). Besides, implementing this technique requires very few modifications to the initial 2D network: only the first convolutional layer is altered, where the number of input features is now taken as the width of the sliding window. In Section 4, we compare the results obtained with the 2D and 2.5D versions of our model and show that the latter enables improvements.

To train our deep learning model, we use a region of interest (ROI) supervised L2 loss inspired by Wu et al. (2020), yet with a different purpose. In fact, we want to focus the loss on the texture area rather than the padding because the latter does not represent useful information to optimise the network weights. Hence, the ROI is centred on the texture:

$$J = \frac{1}{N} \sum_{i=1}^{N} \|x_i^* - x_i\|_2^2 + \lambda \frac{1}{N} \sum_{i=1}^{N} \|\mathcal{M}(x_i^*) - \mathcal{M}(x_i)\|_2^2, \tag{1}$$

where x_i^* and x_i are respectively the output slice i and the corresponding ground truth (*i.e.*, the phantom), $\mathcal{M}$ is a binary mask of the texture area computed on x_i, N is the batch size, and λ is a hyperparameter to balance the two terms. We observed that removing the whole-image term from the loss function worsen the results, as the network is unable to retrieve the boundaries of the ROI due to the gradients not being updated in the masked-out regions.

3.3. Simulated dataset

Training the neural network in a supervised fashion requires data of sufficient quantity and quality. In the following, we create and normalise a novel database to suit our needs.

Synthetic phantoms   We have at our disposal a stochastic model that can generate synthetic phantoms whose texture mimics the one inside a breast (Li et al., 2016). They are made up of two materials, namely simulating glandular and adipose tissues. These textures follow 12 sets of different tissues distribution extracted from the inner part of real breast CT images. An example of a phantom can be seen in Figure 1 (c) and (d).

The 108 generated phantoms have a depth and a width of 5 cm, while their heights are taken randomly between 2.5 and 7 cm to mimic a various range of thicknesses. The digital twin of a commercial DBT device is then used to compute the X-ray projections of those phantoms (Sánchez de la Rosa et al., 2019). No noise and no photon scatter are simulated, and phantoms are lesion-free.

Normalisation step   Pixel values of projections correspond to photon counts. This quantity I can be expressed thanks to the Beer-Lambert law, whose discretised expression is:

$$I_p = I_p^0 e^{-\sum_j \Delta_{pj} \mu_j},$$

where I_p^0 is the number of photons emitted by the X-ray source towards the detector pixel p, μ_j is the attenuation coefficient of voxel j, the latter being taken along the X-ray beam

joining the source to detector pixel p, and Δ_{pj} is the length travelled by the same X-ray beam inside of voxel j. Taking the natural logarithm of the projections makes our problem linear. To get the linear attenuation coefficients, we then need to subtract I^0 from the equation. Blank projections (*i.e.*, without any object placed on the detector) are used to estimate I^0 as in this set-up, photons only go through air, which has a negligible attenuation.

Images are further normalised by the breast thickness, an information which is known from the acquisition phase, in order to retrieve the mean attenuation for each pixel. Furthermore, because the detector of the simulated device does not rotate, photons are colliding with it at different angles. To counterbalance this effect, pixel values are scaled according to their distance to the source.

Reconstruction DBT volumes are reconstructed from normalised projections thanks to a custom FBP-based algorithm. The resolution used to perform this reconstruction is set on all three axes to the detector pixel size, resulting in isotropic voxels of 100 µm^3. The volumes are reconstructed with a margin of 10 mm below the breast support and above the phantom height computed during the acquisition phase.

Data pairing Our network is trained in a supervised way, therefore, a reference is required. However, such an image does not exist for DBT as it would need a device with a wide angular range, which is not available. To overcome this difficulty, we resort to synthetic phantoms whose composition is known and can be used as ground truth. The size in memory of full-resolution reconstructed volumes does not facilitate their handling for the creation of a simple proof of concept. Hence, both phantoms and reconstructions are downsampled to a resolution of 500 µm^3 after application of a low-pass filter to avoid aliasing issues. We sample oblique coronal images from the reconstructed volumes along the chest wall to nipple axis, so that the artefacts we want to address are coplanar to the obtained slices. As phantoms are generated with different thicknesses, 2D slices are zero-padded to give them a constant size of 256 × 256 pixels. Eventually, we end up with 8 856 coronal images. These slices are then separated into three subsets according to the phantom they are created from. Among the initial 108 phantoms, 24 are selected to be equally divided in a validation and test set. The training set is made up of the remainder of the phantoms. We make sure that each one of the 12 texture sets used for generation is represented in all the sets. Finally, the number of slices in training, validation, and test sets amounts respectively to 6 888, 984 and 984.

4. Experiments

Implementation details To train our model, we used Adam optimiser with a learning rate of 10^{-4}, which decreases exponentially at each epoch so that it is divided by 10 every 100 epochs. Each batch contains 8 images, as we observed that both too small and too large batches have a negative impact on the final results. The training dataset is virtually augmented with vertical and horizontal flips to simulate more data. To further increase the variability of the input images, we also apply a constant random shift to the pixel values. In the end, the training database contains 27 552 images.

Evaluation measures To evaluate our method, we use three measures: peak signal-to-noise ratio (PSNR), structural similarity index (SSIM, Wang et al. (2004)), and relative error

Table 1: Evaluation measures averaged on the test dataset. Standard deviations are given in Appendix A.

	FBP	**2D**		**2.5D, 7 slices**		**2.5D, 13 slices**	
		Bias-free	*With bias*	*Bias-free*	*With bias*	*Bias-free*	*With bias*
PSNR ↑	24.28	30.08	30.10	31.51	**31.78**	31.00	31.08
SSIM ↑	0.68	0.86	0.86	**0.89**	**0.89**	0.88	0.88
RE ↓	0.026	0.009	0.008	**0.006**	**0.006**	0.007	0.007

(RE). They are computed only in the texture area, with the same mask as in Equation (1). Results are gathered in Table 1. Note that due to the lack of methods addressing the difficult DBT problem, our approach is compared to FBP only. We can observe that for all configurations of the network, the bias-free version mentioned in Section 3.2 performs slightly worse than the classic one with bias. This behaviour is interesting and demonstrates that even if our model has been empirically shown to be homogeneous, the findings of Mohan et al. (2020) might not be applicable to our problem. Next, we look successively at the results of the 2D and the 2.5D networks, as well as the relevance of bias-free approaches.

2D reconstructions   Test set volumes inferred with the best 2D model are displayed in Figure 2 (c). We can observe that glandular and adipose tissues match pretty well with the ground truth. However, the horizontal edges, where information is most lacking, are the ones where we observe the majority of errors. The boundaries between the two materials on the output reconstructions are quite blurred. Indeed, because of its smoothness around 0, the L2 loss tends to even out the regions having a steep gradient between close pixels. Moreover, the model seems to perform worse on slices with a high heterogeneity. We think that this is due to the blobs of materials being too small in these images for the network to take them into account correctly. The downsampling step applied during the dataset creation might be responsible for this phenomenon. Stacking output 2D slices from our model along the chest wall to nipple axis, we can retrieve a 3D volume. Figure 3 (c) shows an example of axial slice taken from this kind of volume. The image displays horizontal hatchings, also observed by other authors (Teuwen et al., 2021). These degradations can be linked to a lack of spatial continuity between 2D slices in the stacking direction. On this basis, we below assess the relevance of 2.5D approaches to lessen these artefacts.

2.5D results   2.5D inputs are a simple way to add continuity without resorting to 3D convolutions (see Section 3.2). Figure 2 (d) and (e) show coronal slices reconstructed with the best 2.5D models. As displayed in Figure 3 (d) and (e), hatchings on axial planes are smoothed out, but wrong decisions of the network can be propagated through the volume. Nonetheless, as shown in Table 1, all results benefit from more depthwise spatial context. We observe that the best performance is obtained with 7 input slices, and that additional context does not further improve the results. This demonstrates both the benefit and limitation of the 2.5D approach.

Figure 2: Examples of reconstructed coronal slices of 3 different synthetic phantoms from the test set. (a) FBP, (b) ground truth, outputs from (c) 2D network, (d) 2.5D with 7 slices, (e) 2.5D with 13 slices. The arrows are pointing to horizontal edges the network struggles to retrieve. Above images are computed with the best model for each configuration.

Figure 3: Example of a reconstructed axial slice. (a) FBP, (b) ground truth, outputs from (c) 2D network, (d) 2.5D with 7 slices, (e) 2.5D with 13 slices. Hatchings are visible on (c) but are smoothed out on (d) and (e). However, artefacts marked by the arrows are spread.

Test on a clinical image Our deep learning pipeline is able to reconstruct breast texture phantoms on which it has been trained. To test its generalisation ability, we apply our best 2.5D model to a real DBT clinical image. Results obtained with the best 2.5D bias-free model are quite similar and available in Appendix B. Figure 4 tends to show that fibres and gland are better separated than with FBP. Although trained on synthetic images without lesions, our method is able to reconstruct plausible volumes containing a suspect mass. The aforementioned hatchings are however clearly visible. To assess the capacity of the

Figure 4: Example reconstruction of a clinical case with a mass. The two images on the upper left are axial slices of the FBP reconstruction and the result of our best 2.5D model. Bottom left are the corresponding coronal slices taken at the depth of the red dotted line. Orange and green dotted lines in these images are associated respectively with slices at height 102 and 142, and slices at height 112 and 132. Images on the right show FBP axial slices at different heights, with assessment on their actual belonging to the mass.

neural network to reconstruct the depth limits of lesions inside the breast, an image quality expert blindly evaluated the dimensions of a mass on a FBP reconstruction. We then compared the findings with the visual result outputted by the model (Figure 4). The mass location is precisely recovered by our algorithm, as its reconstructed boundaries lie within the expected range. While the behaviour of our method in the inner part of the breast is rather impressive, it does not succeed in recreating the rounded shape of the breast. This is surely due to the phantoms used for training not having a breast shape.

5. Conclusion

In this paper, we proposed one of the first deep learning pipelines for DBT reconstruction. Great care was used to generate the training set, in particular for the meticulous normalisation step. In addition, we also compared the utility of using 2.5D inputs rather than 2D, and showed that introducing more spatial context results in better reconstructions. Thanks to all these improvements, the proposed method reaches good performances, both on synthetic and real datasets.

Our approach might be enhanced with regard to several aspects. First, we strongly believe that our results would be improved using 3D convolutions, as it would give control of the spatial information to the model. Furthermore, as shown by this study, training on realistic synthetic data enables to generalise to clinical images. Developing the realism of the latter, to be even more faithful to a real breast, with a greater variability of materials like fibres and lesions, would doubtlessly be beneficial. Last but no least, the proposed method lacks a parameter to ensure that the projections of the final reconstructed volumes match the ones that were acquired thanks to the device.

Acknowledgments

Acknowledgments withheld.

References

Dianlin Hu, Yikun Zhang, Jin Liu, Shouhua Luo, and Yang Chen. DIOR: Deep Iterative Optimization-Based Residual-Learning for Limited-Angle CT Reconstruction. *IEEE Transactions on Medical Imaging*, 41(7):1778–1790, July 2022. ISSN 1558-254X. doi: 10.1109/TMI.2022.3148110.

Kyong Hwan Jin, Michael T. McCann, Emmanuel Froustey, and Michael Unser. Deep Convolutional Neural Network for Inverse Problems in Imaging. *IEEE Transactions on Image Processing*, 26(9):4509–4522, September 2017. ISSN 1057-7149, 1941-0042. doi: 10.1109/TIP.2017.2713099.

Zhijin Li, Agnès Desolneux, Serge Muller, and Ann-Katherine Carton. A Novel 3D Stochastic Solid Breast Texture Model for X-Ray Breast Imaging. In Anders Tingberg, Kristina Lång, and Pontus Timberg, editors, *Breast Imaging*, Lecture Notes in Computer Science, pages 660–667, Cham, 2016. Springer International Publishing. ISBN 978-3-319-41546-8. doi: 10.1007/978-3-319-41546-8_82.

Kaichao Liang, Hongkai Yang, and Yuxiang Xing. Comparison of projection domain, image domain, and comprehensive deep learning for sparse-view X-ray CT image reconstruction. *arXiv:1804.04289 [physics]*, May 2019.

Jiaming Liu, Rushil Anirudh, Jayaraman J. Thiagarajan, Stewart He, K. Aditya Mohan, Ulugbek S. Kamilov, and Hyojin Kim. DOLCE: A Model-Based Probabilistic Diffusion Framework for Limited-Angle CT Reconstruction, November 2022. arXiv:2211.12340 [cs, eess].

Sreyas Mohan, Zahra Kadkhodaie, Eero P. Simoncelli, and Carlos Fernandez-Granda. Robust and interpretable blind image denoising via bias-free convolutional neural networks. In *International Conference on Learning Representations*, 2020.

L. T. Niklason, B. T. Christian, L. E. Niklason, D. B. Kopans, D. E. Castleberry, B. H. Opsahl-Ong, C. E. Landberg, P. J. Slanetz, A. A. Giardino, R. Moore, D. Albagli, M. C. DeJule, P. F. Fitzgerald, D. F. Fobare, B. W. Giambattista, R. F. Kwasnick, J. Liu, S. J. Lubowski, G. E. Possin, J. F. Richotte, C. Y. Wei, and R. F. Wirth. Digital tomosynthesis in breast imaging. *Radiology*, 1997. doi: 10.1148/radiology.205.2.9356620.

Augustus Odena, Vincent Dumoulin, and Chris Olah. Deconvolution and checkerboard artifacts. *Distill*, 2016. doi: 10.23915/distill.00003. URL http://distill.pub/2016/deconv-checkerboard.

Zhiwei Qiao and Congcong Du. RAD-UNet: A Residual, Attention-Based, Dense UNet for CT Sparse Reconstruction. *Journal of Digital Imaging*, July 2022. ISSN 1618-727X. doi: 10.1007/s10278-022-00685-w.

Olaf Ronneberger, Philipp Fischer, and Thomas Brox. U-Net: Convolutional Networks for Biomedical Image Segmentation. In Nassir Navab, Joachim Hornegger, William M. Wells, and Alejandro F. Frangi, editors, *Medical Image Computing and Computer-Assisted Intervention – MICCAI 2015*, Lecture Notes in Computer Science, pages 234–241, Cham, 2015. Springer International Publishing. ISBN 978-3-319-24574-4. doi: 10.1007/978-3-319-24574-4_28.

R. Sánchez de la Rosa, A.-K. Carton, P. Milioni de Carvalho, I. Bloch, and S. Muller. Analysis of CEDBT and CESM Performance Using a Realistic X-Ray Simulation Platform. In *2019 IEEE 16th International Symposium on Biomedical Imaging (ISBI 2019)*, pages 1070–1073, April 2019. doi: 10.1109/ISBI.2019.8759527.

Ioannis Sechopoulos. A review of breast tomosynthesis. part II. image reconstruction, processing and analysis, and advanced applications. *Medical Physics*, 40(1):014302, 2013. ISSN 0094-2405. doi: 10.1118/1.4770281.

Ting Su, Xiaolei Deng, Zhenwei Wang, Jiecheng Yang, Jianwei Chen, Hairong Zheng, Dong Liang, and Yongshuai Ge. DIR-DBTnet: Deep iterative reconstruction network for 3D digital breast tomosynthesis imaging. *Medical Physics*, 48(5):2289–2300, May 2021. ISSN 0094-2405, 2473-4209. doi: 10.1002/mp.14779.

Jonas Teuwen, Nikita Moriakov, Christian Fedon, Marco Caballo, Ingrid Reiser, Pedrag Bakic, Eloy García, Oliver Diaz, Koen Michielsen, and Ioannis Sechopoulos. Deep learning reconstruction of digital breast tomosynthesis images for accurate breast density and patient-specific radiation dose estimation. *Medical Image Analysis*, 71:102061, July 2021. ISSN 13618415. doi: 10.1016/j.media.2021.102061.

Srinivasan Vedantham, Andrew Karellas, Gopal R. Vijayaraghavan, and Daniel B. Kopans. Digital breast tomosynthesis: State of the art. *Radiology*, 277(3):663–684, 2015. ISSN 0033-8419. doi: 10.1148/radiol.2015141303.

Ge Wang, Jong Chul Ye, and Bruno De Man. Deep learning for tomographic image reconstruction. *Nature Machine Intelligence*, 2(12):737–748, December 2020. ISSN 2522-5839. doi: 10.1038/s42256-020-00273-z.

Zhou Wang, A.C. Bovik, H.R. Sheikh, and E.P. Simoncelli. Image quality assessment: From error visibility to structural similarity. *IEEE Transactions on Image Processing*, 13(4):600–612, April 2004. ISSN 1941-0042. doi: 10.1109/TIP.2003.819861.

Dufan Wu, Kyungsang Kim, and Quanzheng Li. Digital Breast Tomosynthesis Reconstruction with Deep Neural Network for Improved Contrast and In-Depth Resolution. In *2020 IEEE 17th International Symposium on Biomedical Imaging (ISBI)*, pages 656–659, April 2020. doi: 10.1109/ISBI45749.2020.9098661.

Kai Zhang, Wangmeng Zuo, Yunjin Chen, Deyu Meng, and Lei Zhang. Beyond a gaussian denoiser: Residual learning of deep cnn for image denoising. *IEEE transactions on image processing*, 26(7):3142–3155, 2017.

Bo Zhou, Xiongchao Chen, S. Kevin Zhou, James S. Duncan, and Chi Liu. DuDoDR-Net: Dual-domain data consistent recurrent network for simultaneous sparse view and metal artifact reduction in computed tomography. *Medical Image Analysis*, 75:102289, January 2022. ISSN 1361-8415. doi: 10.1016/j.media.2021.102289.

Amirkoushyar Ziabari, Dong Hye Ye, Somesh Srivastava, Ken D. Sauer, Jean-Baptiste Thibault, and Charles A. Bouman. 2.5D Deep Learning For CT Image Reconstruction Using A Multi-GPU Implementation. In *2018 52nd Asilomar Conference on Signals, Systems, and Computers*, pages 2044–2049, October 2018. doi: 10.1109/ACSSC.2018. 8645364. ISSN: 2576-2303.

Appendix A. Evaluation measures

Table 2: Evaluation measures computed with 2D models averaged on the test dataset.

	FBP	**2D**	
		Bias-free	*With bias*
PSNR ↑	24.28 ± 1.79	30.08 ± 3.41	30.10 ± 3.23
SSIM ↑	0.68 ± 0.07	0.86 ± 0.06	0.86 ± 0.06
RE ↓	0.026 ± 0.007	0.009 ± 0.006	0.008 ± 0.005

Table 3: Evaluation measures computed with 2.5D models averaged on the test dataset.

	2.5D, 7 slices		**2.5D, 13 slices**	
	Bias-free	*With bias*	*Bias-free*	*With bias*
PSNR ↑	31.51 ± 3.27	31.78 ± 3.44	31.00 ± 3.41	31.08 ± 3.40
SSIM ↑	0.89 ± 0.05	0.89 ± 0.05	0.88 ± 0.06	0.88 ± 0.06
RE ↓	0.006 ± 0.004	0.006 ± 0.004	0.007 ± 0.004	0.007 ± 0.004

Appendix B. Results of the bias-free network on the clinical image

Figure 5: Example reconstruction of a clinical case with a mass, computed with the best bias-free 2.5D model. Placement of images and legend for the dotted lines are the same as Figure 4. The bias-free model seems to infer less gland on this slice than its counterpart with bias. However, the mass has the same shape and boundaries with both configurations.