# ViTaL-Diff: Video-Token Latent Diffusion for Contactless Respiratory Monitoring

**Shadman Sakib** [1]  **Nirmalya Roy** [1]

## Abstract

Contactless respiratory monitoring from video is a challenging inverse problem because breathing is observed indirectly through subtle body and clothing motion, which can be corrupted by illumination changes, occlusion, pose variation, and non-respiratory movement. We propose **ViTaL-Diff**, a video-token latent diffusion framework that reconstructs respiratory waveforms from short upper-body videos without contact sensors at inference. ViTaL-Diff learns a compact latent space from belt-derived respiratory waveforms and trains a video-conditioned diffusion transformer to generate respiratory latent tokens from spatiotemporal visual evidence. By modeling a distribution over plausible respiratory waveforms, ViTaL-Diff supports both respiratory-rate estimation and uncertainty quantification. Across three datasets, including an in-house RGB dataset, AIR-125, and a low-light Sleep dataset, ViTaL-Diff achieves the lowest error among classical, deterministic, and diffusion baselines, with up to $28.9\%$ mean absolute error (MAE) reduction over deterministic deep baselines. Its uncertainty estimates further identify visually ambiguous clips, providing practical reliability signal for contactless monitoring.

## 1. Introduction

Respiratory rate (RR) is a core vital sign for cardiopulmonary assessment, sleep monitoring, neonatal care, and remote health monitoring (Cretikos et al., 2008; Majumder et al., 2017; Massaroni et al., 2020). Although contact sensors and wearables support continuous RR measurement, they can be uncomfortable, motion-sensitive, and difficult to deploy long term (Massaroni et al., 2019a). Video-based monitoring offers a scalable contactless alternative by capturing respiration-induced cues from the chest, abdomen, shoulders, face, or clothing (Massaroni et al., 2019b; Queiroz et al., 2020).

Existing video-based RR methods use handcrafted motion extraction or learned spatiotemporal representations. Classical approaches rely on optical flow, frame differencing, phase-based analysis, motion magnification, and filtered ROI signals (Siam et al., 2020; Tan et al., 2010a; Massaroni et al., 2019a; Manne et al., 2023), but are fragile under illumination changes, occlusion, weak respiratory motion, and non-respiratory movement. Recent deep models learn video time-series features from RGB, NIR, thermal, or flow-based representations (Khanam et al., 2021; Sakib et al., 2025b; Figueroa et al., 2019; Hasan et al., 2025), while remote-physiology models such as DeepPhys (Chen & McDuff, 2018), TS-CAN (Liu et al., 2020), PhysNet (Yu et al., 2019), and EfficientPhys (Liu et al., 2023) show that spatiotemporal networks can recover physiological signals from video. However, most methods remain deterministic, mapping each clip to a single waveform or RR estimate even when visual evidence is weak or ambiguous. Contactless RR estimation is therefore an ill-posed *inverse problem*: different respiratory waveforms can produce similar visual motion, and nuisance motion can mimic breathing. Diffusion models are well suited to this setting because they learn conditional time-series distributions rather than point estimates (Ho et al., 2020). They have shown strong results in speech generation (Ghosal et al., 2023), time-series generation (Yuan & Qiao, 2024), imputation (Tashiro et al., 2021), biomedical denoising (Liu et al., 2025), and respiratory waveform estimation from PPG (Miao et al., 2025). Yet prior respiratory diffusion methods mainly use contact-sensor inputs or operate directly in waveform space.

We propose ViTaL-Diff, a video-token latent diffusion framework for respiratory waveform reconstruction from upper-body video time series. It learns a compact respiratory latent space from belt-derived waveforms and generates respiratory latent tokens from spatiotemporal video evidence, enabling both RR estimation and uncertainty quantification. ViTaL-Diff handles weak, occluded, or ambiguous breathing cues by modeling multiple plausible respiratory trajectories.

[1]Department of Information Systems, University of Maryland, Baltimore County, Baltimore, MD, USA. Correspondence to: Shadman Sakib <ssakib1@umbc.edu>.

*Proceedings of the Workshop on Structured Data for Health at the $43^{rd}$ International Conference on Machine Learning*, Seoul, South Korea. Copyright 2026 by the author(s).

Our contributions are summarized as follows:

- We formulate video-based respiratory monitoring as an uncertainty-aware conditional latent generative modeling problem, enabling reconstruction of plausible respiratory waveforms from ambiguous visual evidence.

- We introduce ViTaL-Diff, a video-conditioned latent diffusion transformer that generates respiratory time-series tokens from spatiotemporal video evidence.

- We demonstrate consistent improvements across an in-house RGB dataset, AIR-125, and a low-light Sleep dataset, outperforming classical signal-processing, deterministic deep learning, and waveform-space diffusion baselines.

## 2. Problem Setup

We study contactless respiratory waveform reconstruction from short upper-body video clips. During training, each video clip $V = \{F_1, \ldots, F_N\}$ is paired with a synchronized respiratory-belt waveform $y \in \mathbb{R}^T$; at inference time, only $V$ is available. Given paired samples $\mathcal{D} = \{(V_i, y_i)\}_{i=1}^{N_d}$, the goal is to learn a conditional generative model $p_\theta(y \mid V)$ that reconstructs a physiologically plausible respiratory waveform from video alone. This mapping is ambiguous under weak respiratory motion and visual artifacts. We therefore adopt a generative formulation rather than a single deterministic estimate. ViTaL-Diff learns a latent conditional model $p_\theta(z_0 \mid V)$, where the belt waveform is encoded as respiratory latent tokens $z_0 = E_R(y) \in \mathbb{R}^{L \times d}$, and the waveform is recovered by a decoder $D_R$. This latent formulation reduces diffusion dimensionality and enables uncertainty estimation through sample variability.

## 3. Methodology

ViTaL-Diff first learns a compact latent representation of respiratory waveforms, then trains a video-conditioned diffusion model to generate respiratory latent time-series tokens from visual evidence. As shown in Figure 1, the framework contains three modules: (i) a respiratory waveform tokenizer, (ii) a video evidence encoder, and (iii) a video-conditioned latent diffusion transformer.

### 3.1. Respiratory Waveform Tokenizer

The tokenizer maps each belt-derived respiratory waveform into a compact sequence of $L$ latent tokens, each with dimension $d$:

$$z_0 = E_R(y), \qquad z_0 \in \mathbb{R}^{L \times d}. \tag{1}$$

*Figure 1.* Overview of **ViTaL-Diff**: video evidence tokens condition latent diffusion over respiratory waveform tokens.

The decoder reconstructs the waveform from this latent sequence, $\hat{y}_{tok} = D_R(z_0)$. The tokenizer is trained with

$$\mathcal{L}_{tok} = \|y - D_R(E_R(y))\|_1. \tag{2}$$

After training, $E_R$ and $D_R$ are frozen. The encoder provides clean latent targets for diffusion training, while the decoder maps generated latent tokens back to waveform space during inference. The tokenizer is important because respiratory waveforms are low-dimensional but temporally structured. Direct diffusion over waveform samples requires the denoising model to learn both signal structure and sample-level noise removal in the original time domain. By first learning a compact respiratory latent space, ViTaL-Diff separates representation learning from conditional generation. The frozen tokenizer provides a stable target space for diffusion training and constrains generated samples to lie near physiologically plausible respiratory trajectories.

### 3.2. Video Evidence Encoder

The video clip is treated as a visual time series and divided into spatiotemporal tubelets, which are embedded as visual tokens and processed by a video encoder:

$$C = E_V(V), \qquad C \in \mathbb{R}^{M \times d}. \tag{3}$$

Here, $C$ is a sequence of $M$ time-ordered video evidence tokens projected to the same token dimension as the respiratory latent space. Temporal positional embeddings preserve frame order, and validity masks suppress missing or corrupted frames. The evidence tokens $C$ are the only conditioning input to the diffusion model, allowing ViTaL-Diff to learn respiratory cues directly from video time series rather than frequency features.

### 3.3. Video-Conditioned Latent Diffusion

The diffusion model is trained in the respiratory latent space. For each pair $(V, y)$, the frozen respiratory encoder produces

$z_0 = E_R(y)$, and the video encoder produces $C = E_V(V)$. The forward process corrupts $z_0$ with Gaussian noise:

$$q(z_t \mid z_0) = \mathcal{N}\left(z_t; \sqrt{\bar{\alpha}_t} z_0, (1 - \bar{\alpha}_t)I\right), \qquad (4)$$

where $\bar{\alpha}_t = \prod_{s=1}^{t}(1 - \beta_s)$. Equivalently,

$$z_t = \sqrt{\bar{\alpha}_t} z_0 + \sqrt{1 - \bar{\alpha}_t}\epsilon, \quad \epsilon \sim \mathcal{N}(0, I). \qquad (5)$$

The denoising network predicts the clean latent tokens from noisy tokens, timestep, and video evidence:

$$\hat{z}_{0,\theta} = f_\theta(z_t, t, C). \qquad (6)$$

We use clean latent prediction instead of noise-only prediction to better preserve temporal structure. The diffusion objective is

$$\mathcal{L}_{diff} = \mathbb{E}\left[\|z_0 - f_\theta(z_t, t, C)\|_2^2\right]. \qquad (7)$$

Training is therefore two-stage: first train the tokenizer with $\mathcal{L}_{tok}$, then freeze it and train the video-conditioned diffusion model with $\mathcal{L}_{diff}$.

### 3.4. Latent Diffusion Transformer

The denoising network $f_\theta$ is a conditional transformer over respiratory latent time-series tokens. Each block first applies self-attention over the noisy latent sequence,

$$H' = \text{SelfAttn}(H), \qquad (8)$$

followed by cross-attention from respiratory latent tokens to video evidence tokens,

$$H'' = \text{CrossAttn}(H', C). \qquad (9)$$

Self-attention captures temporal dependencies across the respiratory time series, while cross-attention aligns each respiratory latent token with relevant time-ordered video evidence.

### 3.5. Robust Conditioning and Uncertainty Estimation

To improve robustness to weak motion, occlusion, illumination variation, and dropped frames, we apply evidence dropout during training:

$$\tilde{C} = \text{DropEvidence}(C, \rho), \qquad (10)$$

where $\rho$ controls the fraction of evidence tokens removed. With probability $p_{\text{uncond}}$, the evidence tokens are replaced by a learned null condition, enabling classifier-free guidance:

$$\hat{z}_0 = (1 + w)f_\theta(z_t, t, C) - wf_\theta(z_t, t, \varnothing). \qquad (11)$$

The scalar $w$ controls the strength of video conditioning.

At inference time, only video is used. We encode $V$ into evidence tokens $C$, sample $z_T \sim \mathcal{N}(0, I)$, and iteratively denoise to obtain $\hat{z}_0$. The waveform is reconstructed as

$$\hat{y} = D_R(\hat{z}_0). \qquad (12)$$

RR is computed from $\hat{y}$ using a fixed post-processing estimator such as peak counting or dominant-frequency estimation. Since the model is generative, we draw $K$ samples and compute the posterior mean waveform,

$$\bar{y} = \frac{1}{K}\sum_{k=1}^{K} \hat{y}^{(k)}. \qquad (13)$$

RR uncertainty is estimated as:

$$u_{RR} = \text{Var}\left(RR^{(1)}, \dots, RR^{(K)}\right). \qquad (14)$$

A high value of $u_{RR}$ indicates weak visual evidence; thus, ViTaL-Diff provides both a respiratory estimate and a reliability signal. This uncertainty estimate is computed without adding a separate confidence head or auxiliary supervision. Instead, it is derived directly from the posterior variability of sampled respiratory trajectories. When the video contains clear periodic chest or abdominal motion, generated samples tend to agree and produce low $u_{\text{RR}}$. When the visual evidence is weak, occluded, or corrupted by non-respiratory movement, samples become more variable, producing higher $u_{\text{RR}}$. This makes uncertainty estimation naturally aligned with the generative formulation.

## 4. Experiments

### 4.1. Datasets and Preprocessing

We evaluate ViTaL-Diff on three datasets: an in-house 50-video RGB dataset, AIR-125 (Manne et al., 2023), and a Sleep dataset (Hu et al., 2018). The in-house dataset was collected under an institutionally approved IRB protocol and contains synchronized upper-body RGB recordings with a Vernier respiration belt (VER) used to obtain ground-truth respiratory waveforms. AIR-125 contains 125 pediatric videos from diverse real-world settings, with RR ranging from 18 to 42 bpm and varied resolutions and frame rates. The Sleep dataset contains 28 synchronized IR/NIR recordings from 12 adults under low-light sleep conditions, enabling evaluation in low-motion nighttime settings. Across datasets, videos are paired with available respiratory references. For the in-house dataset, the Vernier belt signal is temporally aligned to each video window and normalized per recording. Frames are decoded, invalid frames are removed, resized, normalized, and segmented into 6-second windows with 3-second overlap, producing paired samples $(V_i, y_i)$, where $y_i$ is the ground-truth waveform or RR reference. Classical ROI, optical-flow, FFT, and peak-based

processing are used only for baselines and diagnostics. Additional dataset construction, Vernier belt processing, video–belt alignment, diagnostic motion-signal extraction, and frequency-domain RR estimation details are provided in Sections A.1 to A.5.

## 4.2. Baselines and Metrics

We compare ViTaL-Diff with five baselines: (i) Peak/FFT estimates RR from filtered video-derived motion signals using peak counting or dominant-frequency estimation, following classical video-based respiratory monitoring pipelines (Tan et al., 2010b; Siam et al., 2020; Massaroni et al., 2019b; 2020); (ii) E2RespUNet (Sakib et al., 2025b) and (iii) RespFormer (Sakib et al., 2025a) represent recent contactless respiratory monitoring models for waveform/RR estimation; (iv) latent regression is an internal deterministic ablation that uses the same respiratory tokenizer as ViTaL-Diff but predicts $z_0$ directly from video evidence without diffusion; and (v) waveform-space diffusion applies conditional diffusion directly to respiratory waveforms instead of latent respiratory tokens, following diffusion-based generative modeling for time series and respiratory waveform estimation (Ho et al., 2020; Tashiro et al., 2021; Yuan & Qiao, 2024; Miao et al., 2025).

## 4.3. Implementation Details

ViTaL-Diff is trained in two stages. We train the respiratory tokenizer with $\mathcal{L}_{tok}$ for 50 epochs, freeze it, and then train the video-conditioned latent diffusion transformer with $\mathcal{L}_{diff}$ for 100 epochs. We use AdamW with learning rate $1 \times 10^{-4}$, weight decay $1 \times 10^{-4}$, cosine decay, and batch size 8. Since diffusion is performed in a compact respiratory latent space, we use 200 training timesteps and 20 DDIM sampling steps. Evidence dropout uses $\rho = 0.15$, and classifier-free conditioning dropout uses $p_{\mathrm{uncond}} = 0.1$. At inference, only video is used; latent tokens are decoded into waveforms, and RR is computed from the posterior mean using $K = 10$ samples per clip. Experiments are run on an NVIDIA H200 GPU server, with subject-independent splits when subject identifiers are available.

## 5. Results and Analysis

### 5.1. Overall Respiratory-Rate Estimation

Table 1 reports respiratory-rate MAE across the three datasets. ViTaL-Diff achieves the lowest error in every setting, showing consistent gains over classical signal processing, deterministic neural models, and diffusion-based variants. Against RespFormer, a strong deterministic baseline, ViTaL-Diff reduces MAE by 18.2%, 12.5%, and 22.2% on the in-house RGB, AIR-125, and Sleep datasets, respectively. Compared with E2RespUNet, ViTaL-Diff reduces

*Table 1.* Respiratory-rate MAE across datasets. ViTaL-Diff consistently outperforms classical, deterministic, and diffusion baselines.

| Method | In-house RGB | AIR-125 | Sleep |
|---|---|---|---|
| Peak/FFT | 4.6 | 6.1 | 5.4 |
| E2RespUNet | 3.8 | 4.9 | 4.1 |
| RespFormer | 3.3 | 4.8 | 4.5 |
| Latent regression | 3.6 | 4.4 | 4.2 |
| Waveform-space diffusion | 3.2 | 4.7 | 4.1 |
| **ViTaL-Diff** | **2.7** | **4.2** | **3.5** |

*Table 2.* Aggregate MAE across datasets. The relative reduction is computed with respect to ViTaL-Diff using the average MAE across the three datasets.

| Method | Avg. MAE↓ | Reduction by ViTaL-Diff |
|---|---|---|
| Peak/FFT | 5.37 | 35.4% |
| E2RespUNet | 4.27 | 18.8% |
| RespFormer | 4.20 | 17.5% |
| Latent regression | 4.07 | 14.8% |
| Waveform-space diffusion | 4.00 | 13.3% |
| **ViTaL-Diff** | **3.47** | – |

MAE by 28.9%, 14.3%, and 14.6%. These results indicate that modeling respiratory dynamics as a conditional distribution is more effective than producing a single deterministic estimate from video.

The improvements are consistent across distinct sensing conditions. The in-house RGB dataset contains controlled but visually diverse recordings with multiple postures, viewpoints, and breathing patterns. AIR-125 contains heterogeneous pediatric recordings with varied resolution, frame rate, and visual quality. The Sleep dataset evaluates low-light NIR/IR recordings where respiratory motion can be subtle. The consistent gains across these settings suggest that ViTaL-Diff is not over-specialized to a single acquisition condition, but instead benefits from learning a structured respiratory latent space conditioned on video evidence.

### 5.2. Cross-Dataset Aggregate Performance

To summarize performance across datasets, we also compute the average MAE over the in-house RGB, AIR-125, and Sleep datasets. As shown in Table 2, ViTaL-Diff obtains the lowest average MAE of 3.47 bpm. This corresponds to a 35.4% reduction compared with Peak/FFT, an 18.8% reduction compared with E2RespUNet, a 17.5% reduction compared with RespFormer, a 14.8% reduction compared with latent regression, and a 13.3% reduction compared with waveform-space diffusion.

The aggregate comparison shows that the improvement is not driven by a single dataset. ViTaL-Diff consistently reduces average error across RGB, pediatric, and low-light sleep recordings. Its gains over latent regression and waveform-space diffusion show that both conditional diffusion and the compact respiratory latent space contribute to performance. These trends support combining respiratory

*Table 3.* Component ablation across datasets. Each removed module increases MAE, highlighting the contribution of the full ViTaL-Diff design.

| Variant | In-house RGB | AIR-125 | Sleep |
|---|---|---|---|
| **Full ViTaL-Diff** | **2.7** | **4.2** | **3.5** |
| w/o latent diffusion | 3.6 | 4.4 | 4.2 |
| w/o respiratory latent space | 3.2 | 4.7 | 4.1 |
| w/o evidence dropout | 3.0 | 4.5 | 3.9 |
| w/o uncertainty sampling | 3.1 | 4.6 | 4.0 |

tokenization with video-conditioned latent diffusion.

### 5.3. Benefit of Latent Diffusion

The controlled variants clarify the source of the improvement. Waveform-space diffusion benefits from probabilistic generation, but it denoises raw respiratory samples directly. ViTaL-Diff instead denoises compact respiratory tokens, reducing MAE by $15.6\%$, $10.6\%$, and $14.6\%$ over waveform-space diffusion. This supports moving diffusion from raw waveform samples to a learned respiratory token space. Latent regression uses the same tokenizer but removes diffusion sampling, reducing the model to a deterministic video-to-latent mapping. ViTaL-Diff improves over this variant by $25.0\%$, $4.5\%$, and $16.7\%$, showing that tokenization alone is insufficient and that conditional sampling helps under ambiguous visual evidence. The largest relative improvements occur over the Peak/FFT baseline, where ViTaL-Diff reduces MAE by $41.3\%$, $31.1\%$, and $35.2\%$. This is expected because frequency-domain estimates can be unstable when respiratory motion is weak, mixed with non-respiratory movement, or affected by harmonic ambiguity. In contrast, ViTaL-Diff learns respiratory evidence directly from video tokens and constrains generation through a respiratory latent prior.

### 5.4. Component Analysis

To examine the contribution of each module, Table 3 reports controlled ablations across the three datasets. Removing latent diffusion and replacing it with deterministic latent regression increases MAE from 2.7 to 3.6 bpm on the in-house RGB dataset and from 3.5 to 4.2 bpm on the Sleep dataset, showing that deterministic video-to-latent prediction is less effective under ambiguous visual evidence. Removing the learned respiratory latent space also degrades performance, especially on AIR-125, supporting diffusion over structured respiratory tokens rather than raw waveform samples. Evidence dropout improves robustness to incomplete or weak visual evidence, while removing uncertainty sampling worsens posterior averaging and removes the ability to identify unreliable clips. Overall, the ablations show that ViTaL-Diff benefits from combining respiratory tokenization, latent diffusion, robust conditioning, and posterior sampling.

*Figure 2.* Uncertainty–coverage analysis. ViTaL-Diff is compared with deterministic and generative baselines; MAE decreases as high-$u_{\mathrm{RR}}$ clips are rejected.

### 5.5. Uncertainty-Aware Reliability

Figure 2 evaluates whether the generative model provides a useful reliability signal. We compare ViTaL-Diff with RespFormer and waveform-space diffusion, representing the strongest deterministic and closest generative baselines. We use the sampled RR variance $u_{\mathrm{RR}}$ across diffusion samples as the uncertainty score. As clips with high sampled RR variance $u_{\mathrm{RR}}$ are rejected, MAE decreases monotonically, indicating that uncertain clips are also harder to estimate accurately. On the in-house RGB dataset, rejecting the top $20\%$ most uncertain clips lowers ViTaL-Diff MAE from 2.7 bpm to 2.0 bpm, a $25.9\%$ reduction; rejecting the top $30\%$ lowers MAE to 1.8 bpm, a $33.3\%$ reduction. Thus, sample variability is not merely a byproduct of diffusion sampling, but a meaningful indicator of weak-motion, occluded, or visually ambiguous clips. This property is useful for contactless health monitoring because the system can flag low-confidence predictions rather than always returning an unqualified estimate. In practice, uncertainty scores can reject unreliable windows or down-weight ambiguous segments. Overall, latent diffusion improves RR estimation while providing a practical reliability signal for contactless monitoring.

## 6. Conclusion

We presented ViTaL-Diff, a video-conditioned latent diffusion framework for contactless respiratory waveform reconstruction from upper-body video. By learning a compact respiratory latent space and generating respiratory tokens from video evidence, ViTaL-Diff improves respiratory-rate estimation across RGB, pediatric, and low-light sleep settings without relying on handcrafted motion features. The generative formulation also provides a practical uncertainty signal through sample variability, allowing visually ambiguous clips to be identified during inference. Future work will focus on broader clinical validation, real-time deployment, and improved robustness across cameras, lighting conditions, postures, and patient populations.

## Acknowledgements

This work was partially supported by NSF CAREER Award #1750936, NSF CNS Grant #2233879, NSF IIS Grant #2509680, NSF I-Corps Grant #2502886, U.S. Army Grants #W911NF2120076 and #W911NF2410367, and ONR Grant #N00014-23-1-2119.

## Impact Statement

This work aims to advance contactless respiratory monitoring by reconstructing respiratory waveforms from video without requiring contact sensors at inference. The potential benefits include more accessible respiratory assessment in remote monitoring, sleep monitoring, pediatric care, and other settings where contact sensors may be uncomfortable or difficult to deploy.

At the same time, video-based health monitoring raises important privacy, fairness, and deployment considerations. Respiratory cues can be affected by camera placement, lighting, clothing, skin tone, posture, occlusion, and non-respiratory motion, so performance should be validated across diverse populations and acquisition conditions before clinical use. The method is intended to support respiratory monitoring and reliability estimation, not to replace clinical judgment. Any real-world deployment should follow appropriate consent, data-protection, and institutional review procedures, especially because video data can contain identifiable information.

## References

Vernier go direct respiration belt. https://www.vernier.com/product/go-direct-respiration-belt/.

Chen, W. and McDuff, D. Deepphys: Video-based physiological measurement using convolutional attention networks. In *Proceedings of the european conference on computer vision (ECCV)*, pp. 349–365, 2018.

Cretikos, M. A., Bellomo, R., Hillman, K., Chen, J., Finfer, S., and Flabouris, A. Respiratory rate: the neglected vital sign. *Medical Journal of Australia*, 188(11):657–659, 2008.

Figueroa, I. R. A., Nuño, J. V. M., and Mendizabal-Ruiz, E. G. Remote optical estimation of respiratory rate based on a deep learning human pose detector. In *Latin American Conference on Biomedical Engineering*, pp. 234–241. Springer, 2019.

Ghosal, D., Majumder, N., Mehrish, A., and Poria, S. Text-to-audio generation using instruction guided latent diffusion model. In *Proceedings of the 31st ACM international conference on multimedia*, pp. 3590–3598, 2023.

Hasan, Z., Ahmed, M., Sakib, S., Chugh, S., Khan, M. A., Zaher MD Faridee, A., and Roy, N. Rrpips: Respiratory waveform reconstruction using persistent independent particles tracking from video. In *2025 IEEE/ACM Conference on Connected Health: Applications, Systems and Engineering Technologies (CHASE)*, pp. 13–24, 2025.

Ho, J., Jain, A., and Abbeel, P. Denoising diffusion probabilistic models. *Advances in neural information processing systems*, 33:6840–6851, 2020.

Hu, M., Zhai, G., Li, D., Fan, Y., Duan, H., Zhu, W., and Yang, X. Combination of near-infrared and thermal imaging techniques for the remote and simultaneous measurements of breathing and heart rates under sleep situation. *PloS one*, 13(1):e0190466, 2018.

Khanam, F.-T.-Z., Perera, A. G., Al-Naji, A., Gibson, K., and Chahl, J. Non-contact automatic vital signs monitoring of infants in a neonatal intensive care unit based on neural networks. *Journal of Imaging*, 7(8):122, 2021.

Langley, P. Crafting papers on machine learning. In Langley, P. (ed.), *Proceedings of the 17th International Conference on Machine Learning (ICML 2000)*, pp. 1207–1216, Stanford, CA, 2000. Morgan Kaufmann.

Liu, X., Fromm, J., Patel, S., and McDuff, D. Multi-task temporal shift attention networks for on-device contactless vitals measurement. *Advances in Neural Information Processing Systems*, 33:19400–19411, 2020.

Liu, X., Hill, B., Jiang, Z., Patel, S., and McDuff, D. Efficientphys: Enabling simple, fast and accurate camera-based cardiac measurement. In *Proceedings of the IEEE/CVF winter conference on applications of computer vision*, pp. 5008–5017, 2023.

Liu, Y.-T., Wang, K.-C., Chao, R., Siniscalchi, S. M., Yeh, P.-C., and Tsao, Y. Msemg: Surface electromyography denoising with a mamba-based efficient network. In *ICASSP 2025-2025 IEEE International Conference on Acoustics, Speech and Signal Processing (ICASSP)*, pp. 1–5. IEEE, 2025.

Majumder, S., Mondal, T., and Deen, M. J. Wearable sensors for remote health monitoring. *Sensors*, 17(1):130, 2017.

Manne, S. K. R., Zhu, S., Ostadabbas, S., and Wan, M. Automatic infant respiration estimation from video: A deep flow-based algorithm and a novel public benchmark. In *International Workshop on Preterm, Perinatal and Paediatric Image Analysis*, pp. 111–120. Springer, 2023.

Massaroni, C., Nicolò, A., Lo Presti, D., Sacchetti, M., Silvestri, S., and Schena, E. Contact-based methods for measuring respiratory rate. *Sensors*, 19(4):908, 2019a.

Massaroni, C., Schena, E., Silvestri, S., and Maji, S. Comparison of two methods for estimating respiratory waveforms from videos without contact. In *2019 IEEE International Symposium on Medical Measurements and Applications (MeMeA)*, pp. 1–6. IEEE, 2019b.

Massaroni, C., Nicolo, A., Sacchetti, M., and Schena, E. Contactless methods for measuring respiratory rate: A review. *IEEE Sensors Journal*, 21(11):12821–12839, 2020.

Miao, Y., Chen, Z., Li, C., and Mandic, D. P. Respdiff: An end-to-end multi-scale rnn diffusion model for respiratory waveform estimation from ppg signals. In *ICASSP 2025-2025 IEEE International Conference on Acoustics, Speech and Signal Processing (ICASSP)*, pp. 1–5. IEEE, 2025.

Queiroz, L., Oliveira, H., Yanushkevich, S., and Ferber, R. Video-based breathing rate monitoring in sleeping subjects. In *2020 IEEE International Conference on Systems, Man, and Cybernetics (SMC)*, pp. 2458–2464. IEEE, 2020.

Sakib, S., Shinde, G., Chugh, S., Anwar, M. S., and Roy, N. Respformer: A motion-guided temporal-frequency multimodal fusion transformer for contactless respiratory monitoring. In *2025 International Conference on Machine Learning and Applications (ICMLA)*, pp. 631–638. IEEE, 2025a.

Sakib, S., Shinde, G., Dev, E., and Roy, N. E2respunet: End-to-end respiratory signal reconstruction and rate prediction using a unified attention-enhanced u-net. In *2025 IEEE International Conference on Smart Computing (SMARTCOMP)*, pp. 154–161. IEEE, 2025b.

Siam, A. I., El-Bahnasawy, N. A., El Banby, G. M., Abou Elazm, A., and Abd El-Samie, F. E. Efficient video-based breathing pattern and respiration rate monitoring for remote health monitoring. *JOSA A*, 37(11):C118–C124, 2020.

Tan, K. S., Saatchi, R., Elphick, H., and Burke, D. Real-time vision based respiration monitoring system. *CSNDSP 2010*, pp. 770, 2010a. doi: 10.1109/CSNDSP.2010.

Tan, K. S., Saatchi, R., Elphick, H., and Burke, D. Real-time vision based respiration monitoring system. In *2010 7th International Symposium on Communication Systems, Networks & Digital Signal Processing (CSNDSP 2010)*, pp. 770–774. IEEE, 2010b.

Tashiro, Y., Song, J., Song, Y., and Ermon, S. Csdi: Conditional score-based diffusion models for probabilistic time series imputation. *Advances in neural information processing systems*, 34:24804–24816, 2021.

Yu, Z., Li, X., and Zhao, G. Remote photoplethysmograph signal measurement from facial videos using spatio-temporal networks. In *30th British Machine Visison Conference: BMVC 2019. 9th-12th September 2019, Cardiff, UK*. The British Machine Vision Conference (BMVC), 2019.

Yuan, X. and Qiao, Y. Diffusion-ts: Interpretable diffusion for general time series generation. *arXiv preprint arXiv:2403.01742*, 2024.

# Appendix

## A. Additional Experimental Details

### A.1. In-House Dataset Construction

We developed an in-house video-based respiratory dataset to reflect both clinical and remote monitoring conditions. The dataset currently contains 50 RGB videos collected from 10 adult volunteers, with each recording lasting approximately 2.5–3.5 minutes at $1920 \times 1080$ resolution and 30 fps. Recordings include spontaneous breathing and paced breathing patterns, including slow, regular, and fast breathing. Videos were collected under multiple postures and viewpoints, including front-facing, back-facing, oblique, overhead, and supine views. The dataset also includes variation in lighting conditions, background motion, and environmental noise.

Participants rested for five minutes between recording sessions to reduce fatigue and carry-over effects. The cohort includes variation in age, gender, and skin tone. Data collection was conducted under an institutionally approved IRB protocol, and all participants provided informed consent with privacy and data-protection provisions. We are continuing to expand the dataset with additional subjects, postures, viewpoints, and environmental conditions. After de-identification and institutional approval, we plan to release the dataset to support reproducible research in contactless respiratory monitoring. Ground-truth respiratory waveforms were collected using a Vernier Go Direct Respiration Belt, which measures breathing-induced pressure changes. The belt signal was synchronized with the video stream and used as the reference respiratory waveform. Reference respiratory rates were derived from the belt waveform using peak detection in 30-second sliding windows advanced every 10 seconds.

### A.2. Belt-Signal Processing and Reference RR Construction

Let $s(t)$ denote the raw Vernier belt signal. The signal is first temporally aligned with video timestamps. To reduce subject-specific amplitude variation, the belt waveform is normalized within each recording:

$$\tilde{s}(t) = \frac{s(t) - \mu_s}{\sigma_s + \epsilon}. \tag{15}$$

Here, $\mu_s$ and $\sigma_s$ are the mean and standard deviation of the belt signal, and $\epsilon$ prevents numerical instability.

To remove slow baseline drift and high-frequency noise, the normalized signal is filtered within the respiratory band:

$$s_f(t) = \text{Bandpass}(\tilde{s}(t); f_{\min}, f_{\max}). \tag{16}$$

Here, $f_{\min}$ and $f_{\max}$ define the plausible breathing-frequency range. Peaks are detected from $s_f(t)$. For a

*Figure 3.* Example chest-motion trajectory from the in-house RGB dataset. The vertical chest-center coordinate varies over time due to breathing-induced upper-body motion and occasional non-respiratory movement.

window of length $W$, respiratory rate is computed as

$$RR = \frac{60}{W} |\mathcal{P}_W|. \tag{17}$$

Here, $|\mathcal{P}_W|$ is the number of detected inhalation peaks in the window. In our processing, $W = 30$ seconds and the window advances every 10 seconds. Each 6-second video segment is assigned the nearest belt-derived RR value for segment-level evaluation. For waveform-supervised training, the aligned belt waveform segment is used directly:

$$y_i = s_f(t_i : t_i + T). \tag{18}$$

Here, $T$ is the duration of the waveform segment aligned with video clip $V_i$.

### A.3. Video Processing and Segment Construction

Each video is paired with its corresponding Vernier belt file using the recording identifier. Frames are decoded sequentially at 30 fps, invalid or unreadable frames are removed, and the remaining frames are resized and normalized. Each recording is segmented into 6-second windows with 3-second overlap, producing paired video–waveform examples:

$$(V_i, y_i). \tag{19}$$

Here, $V_i$ is an upper-body video time series and $y_i$ is the synchronized belt-derived respiratory waveform segment. This window length preserves local respiratory dynamics while keeping the video input short enough for efficient video-token modeling.

### A.4. Diagnostic Motion-Signal Construction

In addition to belt-based reference construction, we extract diagnostic video motion signals to inspect respiratory visibility and to implement classical baselines. For ROI-center tracking, the vertical chest displacement signal is represented as

$$m_y(t) = c_y(t). \tag{20}$$

Here, $c_y(t)$ is the vertical coordinate of the tracked chest-region center at time $t$. Since respiration induces cyclic

*Figure 4.* Diagnostic motion-signal processing for classical baselines. The raw motion magnitude is smoothed to reveal respiratory cycles, and peaks/troughs are detected to estimate respiratory rate.

chest and abdomen displacement, $m_y(t)$ often contains respiratory periodicity. Figure 3 shows an example chest-center trajectory over time, where periodic oscillations correspond to respiratory motion and abrupt deviations may indicate body movement or tracking noise.

For frame-difference or optical-flow-based motion, the motion magnitude at frame $t$ is computed over a chest ROI $\Omega$:

$$m(t) = \frac{1}{|\Omega|} \sum_{p \in \Omega} \sqrt{u_t(p)^2 + v_t(p)^2}. \qquad (21)$$

Here, $u_t(p)$ and $v_t(p)$ are the horizontal and vertical motion components at pixel $p$. The resulting one-dimensional motion signal is smoothed using a moving average or low-pass filter:

$$\bar{m}(t) = \frac{1}{K} \sum_{k=0}^{K-1} m(t-k). \qquad (22)$$

Respiratory peaks and troughs are then detected from $\bar{m}(t)$. The complete diagnostic signal-processing pipeline is shown in Figure 4, where the raw motion magnitude is smoothed to reveal respiratory cycles and detected peaks/troughs are used for peak-based respiratory-rate estimation.

## A.5. Frequency-Domain RR Estimation for Classical Baselines

For the Peak/FFT baseline, a one-dimensional motion signal $m(t)$ is filtered within the respiratory band. The dominant frequency is estimated from the Fourier magnitude spectrum:

$$f^\star = \arg \max_{f \in [f_{\min}, f_{\max}]} |\mathcal{F}\{m(t)\}(f)|. \qquad (23)$$

Here, $\mathcal{F}\{\cdot\}$ denotes the Fourier transform. The respiratory rate is computed as

$$RR_{\text{FFT}} = 60 f^\star. \qquad (24)$$

Peak-based RR is computed from the number of detected respiratory peaks:

$$RR_{\text{peak}} = \frac{60}{W} N_{\text{peaks}}. \qquad (25)$$

Here, $N_{\text{peaks}}$ is the number of peaks in a $W$-second window.

## A.6. Algorithm

Algorithm 1 summarizes the two-stage training and inference pipeline of ViTaL-Diff. The respiratory tokenizer is first trained to learn a compact latent representation of belt-derived waveforms. After freezing the tokenizer, the video-conditioned latent diffusion transformer is trained to generate respiratory latent tokens from video evidence. During inference, the model samples multiple respiratory waveforms from video alone and uses their variability to estimate uncertainty.

---

**Algorithm 1** VITAL-DIFF: Training and Inference

---

1: **Input:** Paired data $\mathcal{D} = \{(V_i, y_i)\}_{i=1}^N$, respiratory tokenizer $(E_R, D_R)$, video encoder $E_V$, denoising transformer $f_\theta$.
2: **Stage 1: Learn respiratory latent space**
3: **for** each minibatch of waveforms $y$ **do**
4:  Encode respiratory tokens: $z_0 = E_R(y)$.
5:  Reconstruct waveform: $\hat{y} = D_R(z_0)$.
6:  Update $E_R, D_R$ using $\mathcal{L}_{tok} = \|y - D_R(E_R(y))\|_1$.
7: **end for**
8: Freeze $E_R$ and $D_R$.
9: **Stage 2: Train video-conditioned latent diffusion**
10: **for** each minibatch $(V, y)$ **do**
11:  Compute clean respiratory tokens: $z_0 = E_R(y)$.
12:  Compute video evidence tokens: $C = E_V(V)$.
13:  Sample $t$ and $\epsilon \sim \mathcal{N}(0, I)$.
14:  Corrupt respiratory tokens: $z_t = \sqrt{\bar{\alpha}_t} z_0 + \sqrt{1 - \bar{\alpha}_t} \epsilon$.
15:  Apply evidence dropout to $C$, and optionally replace $C$ with $\varnothing$ for classifier-free guidance.
16:  Predict clean tokens: $\hat{z}_{0,\theta} = f_\theta(z_t, t, C)$.
17:  Update $E_V, f_\theta$ using $\mathcal{L}_{diff} = \mathbb{E}[\|z_0 - f_\theta(z_t, t, C)\|_2^2]$.
18: **end for**
19: **Inference: video only**
20: Encode video evidence $C = E_V(V)$.
21: **for** $k = 1, \ldots, K$ **do**
22:  Sample $z_T^{(k)} \sim \mathcal{N}(0, I)$ and denoise with DDIM to obtain $\hat{z}_0^{(k)}$.
23:  Decode $\hat{y}^{(k)} = D_R(\hat{z}_0^{(k)})$ and estimate $RR^{(k)}$.
24: **end for**
25: Compute $\bar{y} = K^{-1} \sum_{k=1}^K \hat{y}^{(k)}$ and $u_{\text{RR}} = \text{Var}(RR^{(1)}, \ldots, RR^{(K)})$.
26: **Output:** Posterior mean waveform $\bar{y}$, RR estimate, and uncertainty $u_{\text{RR}}$.

---

## B. Extended Experimental Evaluation

### B.1. Full Respiratory-Rate Results

The main paper reports MAE and component ablations. Table 4 supplements the main results with full respiratory-rate estimation metrics, including MAE, RMSE, and Pearson correlation $r$. MAE and RMSE are reported in breaths per minute. Lower MAE/RMSE and higher $r$ indicate better performance. These complementary metrics provide a more complete view of respiratory-rate estimation performance than MAE alone. RMSE captures sensitivity to large errors, while Pearson correlation reflects whether the predicted

*Table 4.* Full respiratory-rate estimation results across datasets. MAE and RMSE are reported in breaths per minute.

| Dataset | Method | MAE↓ | RMSE↓ | $r \uparrow$ |
|---------|--------|------|-------|-----|
| *In-house* | Peak/FFT | 4.6 | 7.0 | 0.56 |
| | E2RespUNet | 3.8 | 5.9 | 0.66 |
| | RespFormer | 3.3 | 5.3 | 0.72 |
| | Latent regression | 3.6 | 5.6 | 0.69 |
| | Waveform-space diffusion | 3.2 | 5.1 | 0.73 |
| | **ViTaL-Diff** | **2.7** | **4.4** | **0.79** |
| *AIR-125* | Peak/FFT | 6.1 | 8.2 | 0.49 |
| | E2RespUNet | 4.9 | 6.7 | 0.62 |
| | RespFormer | 4.8 | 6.4 | 0.65 |
| | Latent regression | 4.4 | 6.1 | 0.67 |
| | Waveform-space diffusion | 4.7 | 6.3 | 0.66 |
| | **ViTaL-Diff** | **4.2** | **5.7** | **0.73** |
| *Sleep NIR* | Peak/FFT | 5.4 | 7.4 | 0.53 |
| | E2RespUNet | 4.1 | 5.9 | 0.68 |
| | RespFormer | 4.5 | 6.1 | 0.66 |
| | Latent regression | 4.2 | 5.8 | 0.68 |
| | Waveform-space diffusion | 4.1 | 5.7 | 0.69 |
| | **ViTaL-Diff** | **3.5** | **5.0** | **0.76** |

respiratory trend follows the reference trend across clips. Therefore, the full table helps distinguish methods that only reduce average error from methods that also improve stability and agreement with the reference signal.

ViTaL-Diff achieves the best performance across all metrics and datasets. On the in-house RGB dataset, it reduces MAE from 3.3 to 2.7 bpm compared with RespFormer and improves correlation from 0.72 to 0.79. On AIR-125, ViTaL-Diff obtains the lowest MAE and RMSE despite the dataset containing heterogeneous pediatric recordings with varied resolutions, frame rates, and visual conditions. On the Sleep dataset, ViTaL-Diff reduces MAE from 4.5 to 3.5 bpm compared with RespFormer and improves RMSE from 6.1 to 5.0 bpm. These results indicate that the latent diffusion formulation reduces both average error and large-error failures while better preserving agreement with the reference respiratory trend.

Compared with waveform-space diffusion, ViTaL-Diff consistently improves MAE, RMSE, and correlation. This supports the benefit of performing diffusion in a learned respiratory latent space rather than directly denoising raw waveform samples. Compared with latent regression, ViTaL-Diff also improves performance across datasets, showing that the gain is not only from tokenizing the respiratory waveform, but also from modeling a conditional distribution over plausible respiratory trajectories.

### B.2. Evaluation Metrics

We evaluate respiratory-rate estimation using mean absolute error, root mean square error, and Pearson correlation. MAE is the average absolute difference between the estimated and reference respiratory rates. RMSE penalizes larger errors

more strongly by averaging squared errors before taking the square root. Pearson correlation measures agreement between the estimated and reference respiratory-rate trends. Lower MAE and RMSE indicate smaller respiratory-rate error, while higher Pearson correlation indicates better agreement with the reference trend.

### B.3. Reproducibility and Implementation Notes

All experiments use the same preprocessing, windowing, and respiratory-rate estimation protocol across methods whenever applicable. Video clips are segmented into 6-second windows with 3-second overlap, and respiratory references are temporally aligned to the corresponding video segment. The respiratory tokenizer is trained first and then frozen before training the video-conditioned latent diffusion model. During inference, only video is used; the Vernier belt signal is used only for training supervision and evaluation.

For diffusion sampling, ViTaL-Diff generates multiple respiratory waveform samples per clip and computes the final RR estimate from the posterior mean waveform. The same generated samples are also used to compute the uncertainty score $u_{\mathrm{RR}}$. This keeps the reliability estimate tied directly to the model posterior rather than requiring an additional confidence head. All reported results use the same evaluation metrics and post-processing rules across datasets to ensure a consistent comparison.

### B.4. Uncertainty-Aware Evaluation

Table 5 reports the uncertainty-aware evaluation on the in-house RGB dataset. For each clip, ViTaL-Diff generates $K = 10$ respiratory waveform samples. RR is computed from each generated waveform, and the variance across sampled RR estimates is used as the uncertainty score $u_{\mathrm{RR}}$. Clips with the highest uncertainty are rejected, and MAE is computed on the retained subset.

*Table 5.* Uncertainty-aware evaluation on the in-house RGB dataset.

| Coverage | MAE↓ | Relative reduction |
|----------|------|--------------------|
| 100% | 2.7 | – |
| 90% | 2.3 | 14.8% |
| 80% | 2.0 | 25.9% |
| 70% | 1.8 | 33.3% |

### B.5. Extended Evaluation Analysis

The monotonic decrease in MAE shows that $u_{\mathrm{RR}}$ is correlated with prediction difficulty. Rejecting the top 20% most uncertain clips reduces MAE from 2.7 bpm to 2.0 bpm, while rejecting the top 30% reduces MAE to 1.8 bpm. This indicates that sample variability is not merely a byproduct

*Figure 5.* Normalized multi-metric comparison. MAE, RMSE, correlation, and reliability are normalized so that higher values indicate better performance. ViTaL-Diff achieves the strongest overall profile.

*Figure 6.* Uncertainty–coverage trend across baselines. ViTaL-Diff maintains the lowest MAE across coverage levels, and its error decreases as high-uncertainty clips are rejected.

*Figure 7.* Qualitative waveform reconstruction. The Vernier belt reference is compared with the ViTaL-Diff posterior mean and uncertainty band computed from $K = 10$ generated samples.

of diffusion sampling, but a useful reliability signal. High-uncertainty clips typically correspond to weak respiratory motion, occlusion, non-respiratory movement, poor visual quality, or ambiguous periodic patterns.

In deployment, $u_{RR}$ can be used as a quality indicator for continuous monitoring. Predictions from low-uncertainty clips can be aggregated to form a stable RR estimate, while high-uncertainty clips can be down-weighted or flagged for review. This is especially useful when respiratory motion is partially occluded, weak, or mixed with non-respiratory body movement. Thus, ViTaL-Diff provides both a respiratory estimate and a mechanism for identifying when that estimate may be unreliable.

Figure 5 summarizes normalized MAE, RMSE, correlation, and reliability scores across all methods. The metrics are scaled so that higher values indicate better performance. ViTaL-Diff achieves the strongest overall profile, with the best normalized scores across all four criteria. This supports the main quantitative results and shows that the improvement is not limited to MAE alone. Figure 6 shows the uncertainty–coverage trend across all baselines. ViTaL-Diff maintains the lowest MAE across retained coverage levels. As high-uncertainty clips are rejected, its MAE decreases consistently, indicating that sampled RR variance is correlated with prediction difficulty. This supports the use of $u_{RR}$ as a practical reliability signal for contactless respiratory monitoring.

Figure 7 provides a qualitative waveform reconstruction example. The ViTaL-Diff posterior mean closely follows the Vernier belt reference, while the uncertainty band highlights regions where the generated samples vary more. This illustrates that the model can recover respiratory waveform structure while also providing uncertainty information from multiple generated samples.

### B.6. Failure Case Discussion

Failure cases mainly occur when the video contains limited visible respiratory evidence. Examples include weak chest or abdominal motion, partial occlusion, posture changes, loose clothing, abrupt body movement, and low illumination. These conditions can make respiratory motion difficult to distinguish from non-respiratory motion, leading to larger RR errors or wider posterior variability. The uncertainty estimates help identify many of these difficult cases. Clips with high $u_{RR}$ often correspond to ambiguous visual evidence, where multiple plausible respiratory waveforms can explain the observed motion. In practice, such clips can be flagged for review, down-weighted during temporal aggregation, or excluded when a more reliable estimate is required. This makes the uncertainty output useful not only as an analysis tool, but also as a practical mechanism for safer deployment in uncontrolled contactless monitoring settings.

From a deployment perspective, these failure cases are expected rather than unusual. Contactless respiratory monitoring depends on subtle motion cues, so the model should not be forced to treat every clip as equally reliable. The uncertainty output provides a simple way to separate confident predictions from visually ambiguous segments, improving the practical usability of ViTaL-Diff in unconstrained monitoring environments.